

Programmable AC Power Supply

IT7900 Series User Manual



Model: IT7900

Version: V1.1/2024.08



Notices

© Itech Electronic, Co., Ltd. 2024
No part of this manual may be
reproduced in any form or by any
means (including electronic storage
and retrieval or translation into a foreign
language) without prior permission and
written consent from Itech Electronic,
Co., Ltd. as governed by international
copyright laws.

Manual Part Number

IT7900

Revision

1st Edition: Aug.01, 2024. Itech Electronic, Co., Ltd.

Trademarks

Pentium is U.S. registered trademarks of Intel Corporation.

Microsoft, Visual Studio, Windows and MS Windows are registered trademarks of Microsoft Corporation in the United States and/or other countries and regions.

Warranty

The materials contained in this document are provided "as is", and is subject to change, without prior notice, in future editions. Further, to the maximum extent permitted by applicable laws, ITECH disclaims all warrants, either express or implied, with regard to this manual and any information contained including but not limited to the implied warranties of merchantability and fitness for a particular purpose. ITECH shall not be held liable for errors or for incidental or indirect damages in connection with the furnishing, use or application of this document or of any information contained herein. Should ITECH and the user enter into a separate written agreement with warranty terms covering the materials in this document that conflict with these terms, the warranty terms in the separate agreement shall prevail.

Technology Licenses

The hardware and/or software described herein are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

Restricted permissions of the U.S. government. Permissions for software and technical data which are authorized to the U.S. Government only include those for custom provision to end users. ITECH provides this customary commercial license in software and technical data pursuant to FAR 12.211 (Technical Data) and 12.212 (Computer Software) and DFARS 252.227-7015 (Technical Data – Commercial Items) and DFARS 227.7202-3 (Rights in Commercial Computer Software or Computer Software Documentation).

Safety Notices

CAUTION

A CAUTION sign denotes a hazard. It calls attention to an operating procedure or practice that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION sign until the indicated conditions are fully understood and met.

WARNING

A WARNING sign denotes a hazard. It calls attention to an operating procedure or practice that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING sign until the indicated conditions are fully understood and met.



NOTE

A NOTE sign denotes important hint. It calls attention to tips or supplementary information that is essential for users to refer to.



Quality Certification and Assurance

We certify that IT7900 series power supply meets all the published specifications at time of shipment from the factory.

Warranty

ITECH warrants that the product will be free from defects in material and workmanship under normal use for a period of one (1) year from the date of delivery (except those described in the Limitation of Warranty below).



Visit https://www.itechate.com/en/support/register.html to complete product registration by filling out the necessary information to extend the warranty to two (2) years.

For warranty service or repair, the product must be returned to a service center designated by ITECH.

- The product returned to ITECH for warranty service must be shipped PREPAID. And ITECH will pay for return of the product to customer.
- If the product is returned to ITECH for warranty service from overseas, all the freights, duties and other taxes shall be on the account of customer.

Limitation of Warranty

This Warranty will be rendered invalid if the product is:

- Damaged resulting from customer-wired circuits or customer-supplied parts or accessories;
- Modified or repaired by customer without authorization;
- Damaged resulting from customer-wired circuits or use in an environment not designated by us;
- The product model or serial number is altered, deleted, removed or made illegible by customer;
- Damaged as a result of accidents, including but not limited to lightning, moisture, fire, improper use or negligence.

Safety Symbols

===	Direct current		ON (power)
~	Alternating current	0	OFF (power)
\sim	Both direct and alternating current	ф	Power-on state
	Chassis (earth ground) symbol.	Д	Power-off state
4	Earth (ground) terminal	土	Reference terminal



4	Caution (possibility of electric shock)	+	Positive terminal
	Warning (refer to this manual for specific Warning or Caution information)		Negative terminal
777	A chassis terminal	-	-

Safety Precautions

The following safety precautions must be observed during all phases of operation of this instrument. Failure to comply with these precautions or specific warnings elsewhere in this manual will constitute a default under safety standards of design, manufacture and intended use of the instrument. ITECH assumes no liability for the customer's failure to comply with these precautions.

WARNING

- Do not use the instrument if it is damaged. Before operation, check the casing to see whether it cracks. Do not operate the instrument in the presence of inflammable gasses, vapors or dusts.
- The instrument is provided with a power cord during delivery and should be connected to a socket with a protective earth terminal, a junction box or a three-phase distribution box. Before operation, be sure that the instrument is well grounded.
- Check all marks on the instrument before connecting the instrument to power supply.
- Use electric wires of appropriate load. All loading wires should be capable of bearing maximum short-circuit of electronic load without overheating. If there are multiple loads, each pair of the load power cord must be carry out the full rated short-circuit output current of the power securely.
- Ensure the voltage fluctuation of mains supply is less than 10% of the working voltage range in order to reduce risks of fire and electric shock.
- Do not install alternative parts on the instrument or perform any unauthorized modification.
- Do not use the instrument if the detachable cover is removed or loosen.
- To prevent the possibility of accidental injuries, be sure to use the power adapter supplied by the manufacturer only.
- We do not accept responsibility for any direct or indirect financial damage or loss of profit that might occur when using the instrument.
- This instrument is used for industrial purposes, do not apply this product to IT power supply system.
- Never use the instrument with a life-support system or any other equipment subject to safety requirements.



WARNING

- SHOCK HAZARD Ground the Instrument. This product is provided with a protective earth terminal. To minimize shock hazard, the instrument must be connected to the AC mains through a grounded power cable, with the ground wire firmly connected to an electrical ground (safety ground) at the power outlet or distribution box. Any interruption of the protective (grounding) conductor or disconnection of the protective earth terminal will cause a potential shock hazard that could result in injury or death.
- Before applying power, verify that all safety precautions are taken. All
 connections must be made with the instrument turned off, and must be
 performed by qualified personnel who are aware of the hazards involved.
 Improper actions can cause fatal injury as well as equipment damage.
- SHOCK HAZARD, LETHAL VOLTAGES This product can input the dangerous voltage that can cause personal injury, and the operator must always be protected from electric shock. Ensure that the input electrodes are either insulated or covered using the safety covers provided, so that no accidental contact with lethal voltages can occur.
- Never touch cables or connections immediately after turning off the instrument. Verify that there is no dangerous voltage on the electrodes or sense terminals before touching them.

CAUTION

- Failure to use the instrument as directed by the manufacturer may render its protective features void.
- Always clean the casing with a dry cloth. Do not clean the internals.
- Make sure the vent hole is always unblocked.

Environmental Conditions


The instrument is designed for indoor use and an area with low condensation. The table below shows the general environmental requirements for the instrument.

Environmental Conditions	Requirements
Operating temperature	0°C∼40°C
Operating humidity	20% \sim 80% (non-condensation)
Storage temperature	-10°C∼70 °C
Altitude	Operating up to 2,000 meters
Installation category	II '
Pollution degree	Pollution degree 2



To make accurate measurements, allow the instrument to warm up for 30 min.

Regulatory Markings





The instrument complies with the WEEE Directive (2002/96/EC) marking requirement. This affix product label indicates that you must not discard the electrical/electronic product in domestic household waste.
This symbol indicates the time period during which no hazardous or toxic substances are expected to leak or deteriorate during normal use. The expected useful life of the product is 10 years. The product can be used safely during the 10-year Environment Friendly Use Period (EFUP). Upon expiration of the EFUP, the product must be immediately recycled.

Waste Electrical and Electronic Equipment (WEEE) Directive



2002/96/EC Waste Electrical and Electronic Equipment (WEEE) Directive

This product complies with the WEEE Directive (2002/96/EC) marking requirement. This affix product label indicates that you must not discard the electrical/electronic product in domestic household waste.

Product Category

With reference to the equipment classifications described in the Annex 1 of the WEEE Directive, this instrument is classified as a "Monitoring and Control Instrument".

To return this unwanted instrument, contact your nearest ITECH office.



Compliance Information

Complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- Electromagnetic Compatibility (EMC) Directive 2014/30/EU
- Low-Voltage Directive (Safety) 2014/35/EU

Conforms with the following product standards:

EMC Standard

IEC 61326-1:2012/ EN 61326-1:2013 123

Reference Standards

CISPR 11:2009+A1:2010/ EN 55011:2009+A1:2010 (Group 1, Class A)

IEC 61000-4-2:2008/ EN 61000-4-2:2009

IEC 61000-4-3:2006+A1:2007+A2:2010/ EN 61000-4-

3:2006+A1:2008+A2:2010

IEC 61000-4-4:2004+A1:2010/ EN 61000-4-4:2004+A1:2010

IEC 61000-4-5:2005/ EN 61000-4-5:2006

IEC 61000-4-6:2008/ EN 61000-4-6:2009

IEC 61000-4-11:2004/ EN 61000-4-11:2004

- 1. The product is intended for use in non-residential/non-domestic environments. Use of the product in residential/domestic environments may cause electromagnetic interference.
- 2. Connection of the instrument to a test object may produce radiations beyond the specified limit.
- 3. Use high-performance shielded interface cable to ensure conformity with the EMC standards listed above.

Safety Standard

IEC 61010-1:2010/ EN 61010-1:2010



Content

Warranty		i
vvarianty.		i
Limitation	of Warranty	i
Safety Sym	nbols	i
•	cautions	
	ental Conditions	
	y Markings	
	ctrical and Electronic Equipment (WEEE) Directive	
Compliand	ce Information	v
Chapter1	Overview	9
1.1 Brief Ir	ntroduction	g
1.2 Model	ls and Options	10
1.3 Option	nal accessories	11
1.4 Instrur	ment Size Introduction	12
Chapter2	Instrument Introduction	19
2.1 Unpac	king and Transportation	19
2.2 Verifyi	ing the Shipment	20
2.3 Front F	Panel	21
2.4 Keybo	ard	21
2.5 Push-c	on Knob	23
2.6 Rear P	anel	24
Chapter3	Installation	27
3.1 Conne	ecting the Power Cord	27
	ecting Test Lines (Optional)	
3.3 Installi	ing Fiber Cables (Only for IT7930-350-180)	35
Chapter4	Getting Started	37
4.1 Power	-on the Instrument	37
4.2.11	-Screen Overview	
4.2 Home-		38
	Screen Introduction	
4.3 Touch		42
4.3 Touch 4.4 Set ou	Screen Introduction	42 42
4.3 Touch 4.4 Set ou	Screen Introductiontput parameters	42 42 43
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5	Screen Introductiontput parametersttput On/Off Control	42 42 43
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-0	Screen Introduction	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-0 5.2 Select	Screen Introduction	
4.3 Touch 4.4 Set ou 4.5 Outpur Chapter5 5.1 full 4-0 5.2 Select 5.3 Select 5.3.1 AG	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character the Power Supply Mode the Output Mode C Output Mode	
4.3 Touch 4.4 Set ou 4.5 Outpur Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode C Output Mode C Output Mode	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC	Screen Introduction	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode COutput Mode COutput Mode CHDC Mode CHDC Mode CHAC Mode	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-0 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode COutput Mode COutput Mode CHDC Mode CHDC Mode CHAC Mode Ithius Mode and Power Limit Mode	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode IC Output Mode	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes: 5.6 Wavefo	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode IC Output Mode IC Output Mode IC COUTPUT Mode IC CHAC Mode INTERIOR IN	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes: 5.6 Wavef 5.7 Full 4-C	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode COutput Mode COutput Mode CHDC Mode CHAC Mode Int Limit Mode and Power Limit Mode Issional Anti-islanding Test Mode Quadrant Power Amplifier	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-0 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes: 5.6 Wavef 5.7 Full 4-0 5.8 Auto-r	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode COutput Mode COutput Mode CHDC Mode Ithinit Mode and Power Limit Mode Issional Anti-islanding Test Mode Issional Power Amplifier Itange Function	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-0 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes: 5.6 Wavefi 5.7 Full 4-0 5.8 Auto-r 5.9 Progra	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode COutput Mode COutput Mode CHDC Mode CHAC Mode Int Limit Mode and Power Limit Mode Issional Anti-islanding Test Mode Ouddrant Power Amplifier Itange Function International Impedance	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes 5.6 Wavef 5.7 Full 4-C 5.8 Auto-r 5.9 Progra 5.10 Swee	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode IC Output Mode IC Output Mode IC HOC	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes: 5.6 Wavefe 5.7 Full 4-C 5.8 Auto-r 5.9 Progra 5.10 Swee 5.11 Polyp	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode IC Output Mode IC Output Mode IC HOC Mode INT Limit Mode and Power Limit Mode Issional Anti-islanding Test Mode Icorm Selection IC Quadrant Power Amplifier Icorm Selection ICORD Impedance Icorm Selection ICORD Impedance ICORD Impedance Icorm Selection ICORD Impedance ICORD IMPEDATE IMPEDAT	
4.3 Touch 4.4 Set ou 4.5 Outpur Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes: 5.6 Wavefi 5.7 Full 4-C 5.8 Auto-r 5.9 Progra 5.10 Swee 5.11 Polyp 5.11.1 C	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode COutput Mode COutput Mode CHDC Mode It Limit Mode and Power Limit Mode Isional Anti-islanding Test Mode Isoma Selection Quadrant Power Amplifier Irange Function Immable Output Impedance Imped	
4.3 Touch 4.4 Set ou 4.5 Outpu Chapter5 5.1 full 4-C 5.2 Select 5.3 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profes: 5.6 Wavef 5.7 Full 4-C 5.8 Auto-r 5.9 Progra 5.10 Swee 5.11 Polyp 5.11.1 C 5.11.2 C	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character The Power Supply Mode The Output Mode COutput Mode COutput Mode CHDC Mode CHAC Mode The Limit Mode and Power Limit Mode Torm Selection Quadrant Power Amplifier Trange Function The Sep Function The Sep Function The System Bus Method Torn Connecting the System Bus Method	
4.3 Touch 4.4 Set ou 4.5 Outpur Chapter5 5.1 full 4-0 5.2 Select 5.3.1 AC 5.3.2 DC 5.3.3 AC 5.3.4 DC 5.4 Curren 5.5 Profess 5.6 Wavef 5.7 Full 4-0 5.8 Auto-r 5.9 Progra 5.10 Swee 5.11 Polyp 5.11.1 C 5.11.2 C 5.11.3 C	Screen Introduction Itput parameters It On/Off Control Operation and Application Quadrant output character Ithe Power Supply Mode Ithe Output Mode COutput Mode COutput Mode CHDC Mode It Limit Mode and Power Limit Mode Isional Anti-islanding Test Mode Isoma Selection Quadrant Power Amplifier Irange Function Immable Output Impedance Imped	



Chapter6	System-Related Functions	62
6.1 Syster	m Menu Reference	62
6.1.1 N	Лепи function	69
6.1.2 Se	et the communication interface	70
6.1.3 V	iew the System Information	71
_	guration Menu Reference	
	ock Function	
6.4 Switch	hing Local/Remote Mode	74
6.5 Save a	and Recall Operations	75
6.6 Proteo	ction Function	76
	Current RMS Protection	
6.6.2 Se	et the Current Peak protection	77
	et Voltage/Current/Power Limit Range	
	Over-temperature protection (OTP)	
	Over-power protection (OPP)	
	n Capture Function	
	er Function	
-	y the System Log	
	ry the Energy	
	Parallel Operation Mode	
	ote Measurement Function	
_	tal I/O Function	
6.14 Anal	logue Function (Ext-Program) (Optional)	88
Chapter7	Current Source Mode Operation	92
7.1 Select	t the Power Supply	92
7.2 Select	t the Output Mode	92
7.2.1 A	NC Output Mode	92
7.2.2 A	NC+DC Output Mode	93
Chapter8	Measurement Functions	94
8.1 Meter	r Mode	94
8.2 Oscillo	oscope Mode	95
8.3 Harmo	onic measurement	97
8.3.1 H	larmonic measurement	97
8.3.2 H	Harmonic Analysis	99
8.4 Record	der Function	101
Chapter9	Configuration Arbitrary Waveform	103
9.1 List Fu	unction	103
9.1.1	Create a new List file	103
9.1.2 S	elect/Run List File	105
9.1.3 In	mport/Export List file	106
9.2 Settin	ng of Surge/Sag Configuration	106
9.3 Self-de	efined Waveform Function	108
9.3.1 7	THD	108
9.3.2 L	User-defined waveform	110
	ge Sampling	
9.5 Voltag	ge Signals Simulation	114
9.6 Interh	narmnics function	115
9.7 Aircra	ıfts Power Interruption Test	117
Chapter10	Technical Specifications	118
10.1 Supp	olemental characteristics	118
	n technical parameters	
	?-350-10U	
	!-350-20U	
IT7905	i-350-30U	122



(112011		Overview
IT7906-3.	350-30	
IT7906-3.	350-90	127
IT7909-3.	350-90	130
IT7912-3	350-90	132
IT7915-3	350-90	135
IT7930-3	350-180	138
IT7945-3.	350-270	141
IT7960-3.	350-360	144
IT7975-3.	350-450	146
IT7990-3.	350-540	149
IT79105-	-350-630	152
IT79120-	-350-720	155
IT79135	-350-810	158
IT79150-	-350-900	160
IT79165-	-350-990	163
Chapter11	Remote Control	167
11.1 USB In	iterface	167
11.2 LAN In	nterface	167
11.3 CAN In	nterface	170
11.4 GPIB Ir	nterface (Optional)	171
11.5 RS-232	2 Interface (Optional)	172
11.6 Comm	nonly Used Commands Overview	173
11.7 Demo	Software Introduction	174
Appendix		175
Specificatio	ons of Red and Black Test Lines	



Chapter1 Overview

This chapter introduces the front panel, the rear panel, key functions and LCD display function of the IT7900 series power supply, make sure that you can quickly know the appearance, instruction and the key function before you operate the power supply. Help you make better use of this series of power supply.

1.1 Brief Introduction

ITECH IT7900 series represents a new generation of programmable, full power source and sink 4-Quadrant grid simulators and also can perform as a 4-Quadrant power amplifier, providing the flexibility to test widest range of grid-tied devices, including PCS, energy storage systems, micro grids, BOBC (V2X) and power-related hardware in the loop emulation (PHiL).

The IT7900 series is equipped with a professional anti-islanding test mode, allowing the user to set R, L, C and active and reactive power parameters to simulate non-linear loads on the grid and achieve anti- islanding testing requirement. The IT7900 series supports energy recovery function that provides 100% current absorption and feeds back into local grid through the device, saving electricity bills and heat dissipation costs.

ITECH IT7900 Regenerative Grid Simulator adopts high power density design of 15kVA in 3U enclosure and provides wide voltage ranges up to 350V L-N and 500VL-N. Easily expandable by Master-slave parallel configuration for higher power up to 960kVA. Comprehensive options of operating modes can meet users' requirements for single-phase, three-phase, reversed-phase and multichannel testing requirements. Extended AC voltage range to 700V L-N enabled by reversed mode. The powerful arbitrary waveform editing function can simulate a variety of grid disturbance waveforms, which is an ideal solution for testing and R&D laboratories.



- High power density/minimum rack space, 3U up to 15kVA,16Hz~150Hz
- Regenerative grid simulator & full 4-Quadrant AC&DC power sources
- Power Amplifier function for PHiL applications
- Professional anti-islanding test mode, can set and simulate the RLC (resistive-inductive-capacitive), active and reactive power circuit for antiislanding detection.
- CC/CV/CP modes
- wide voltage ranges: 350V L-N and 500V L-N*1



- Master-slave parallel with current sharing technology, up to 960kVA,
- Intuitive touch screen; Built-in single/3-phase AC power meter; Scope function
- Fast response time and high accuracy 0.1%+0.2%FS
- waveforms Library: Sine wave, Square wave, Triangle wave, Clipped sine waves, Staircase wave, self-defined waves
- Harmonics and Interharmonics waveform synthesizer
- Power line disturbance simulation testing by List programming/Sweep/Surge&Sag functions
- The harmonic measurement function can measure 50th order harmonics of voltage and current.
- Output 0-360 ° start/stop phase angle can be set
- AC, DC, AC+DC or DC+AC output capability
- Comprehensive working modes selectable: single-phase ,three-phase, reversed phase and multi-channel
- Programmable Output Impedance, allows simulation of Real-World Utility Grid Impedance.
- compliance tests incl LVRT /Phase Jump/Frequency variation /Harmonic Injection
- Supported regulatory testing include IEC61000-4-11/4-13/4-14/4-17/4-28/4-29 etc.
- Front panel USB port for data and waveform import and export
- Provides various trigger input/output signals. When amplitude/frequency changes, trigger signals can be generated to synchronously capture the current waveform of DUTs
- Relay CTRL function, to cut off the connection between instrument and DUT
- Frequency lock and phase lock function, tracking the external signal frequency and phase, to achieve 6 phase& 12 phase power output
- Built-in USB/CAN/LAN/Digital IO interface, optional GPIB /Analog&RS232

1.2 Models and Options

Model	Power	Voltage	Current	Phase	Height
IT7902-350-10U	2kVA	350V	10A	1Ф	2U
IT7902-350-10U-ATE	2kVA	350V	10A	1Ф	1U
IT7904-350-20U	4kVA	350V	20A	1Ф	2U
IT7904-350-20U-ATE	4kVA	350V	20A	1Ф	1U
IT7905-350-30U	5kVA	350V	30A	1Ф	3U
IT7906-350-30	6kVA	350V	30A	1Ф or 3Ф	2U
IT7906-350-30-ATE	6kVA	350V	30A	1Ф or 3Ф	1U
IT7906-350-90	6kVA	350V	90A	1Ф or 3Ф	3U
IT7909-350-90	9kVA	350V	90A	1Ф or 3Ф	3U
IT7912-350-90	12kVA	350V	90A	1Ф or 3Ф	3U
IT7915-350-90	15kVA	350V	90A	1Ф or 3Ф	3U



Model	Power	Voltage	Current	Phase	Height
IT7930-350-180	30kVA	350V	180A	1Ф or 3Ф	6U
IT7945-350-270	45kVA	350V	270A	1Ф or 3Ф	15U
IT7960-350-360	60kVA	350V	360A	1Ф or 3Ф	27U
IT7975-350-450	75kVA	350V	450A	1Ф or 3Ф	27U
IT7990-350-540	90kVA	350V	540A	1Ф or 3Ф	27U
IT79105-350-630	105kVA	350V	630A	1Ф or 3Ф	27U
IT79120-350-720	120kVA	350V	720A	1Ф or 3Ф	37U
IT79135-350-810	135kVA	350V	810A	1Ф or 3Ф	37U
IT79150-350-900	150kVA	350V	900A	1Ф or 3Ф	37U
IT79165-350-990	165kVA	350V	990A	1Ф or 3Ф	37U

Note

Naming rules for this series of model are as follows: IT79XXX-YYY-ZZZ, wherein, XXX means rated power, YYY means rated voltage, and ZZZ means rated current. Take IT7915-350-90 for an example, the rated power is 15kVA, rated voltage is 350V and rated current is 90A. In addition, ATE indicates that this model is a system model without an operator panel.

1.3 Optional accessories

The IT7900 series supports the following optional accessories (sold separately), the details are as below:

The interface expansion slot provided on the rear panel of the IT7900 series

instrument allows users to flexibly expand according to their needs. Different interface cards can be selected to achieve different functions.

The following optional accessories from ITECH are sold separately. Users need to purchase separately.


Device Name	Model	Description
GPIB communication interface	IT-E176	When the user needs to use GPIB interface to enable remote operation, this option is the right choice.
RS232+Analog	IT-E177	Interface card that includes RS-232 communication interface and external analog. When the user needs to use RS-232 or external analog interface to enable remote operation, this option is the right choice.





Device Name M		Model	Description
Fiber modules cables	optic and	IT-E168	Used for parallel connection between the units in a cabinet, including one fiber module and two fiber cables, with lengths of 1.5 meters and 0.3 meters.
			The fiber optic module and cable are the necessary accessories for the parallel connection. Different numbers of fiber optic modules and cables are used in different numbers of parallels.
Fiber modules cables	optic and	IT-E169	Used for parallel connection between cabinets, including one fiber module and one 2.5m fiber cable.
			The fiber optic module and cable are the necessary accessories for the parallel connection. Different numbers of fiber optic modules and cables are used in different numbers of parallels.

1.4 Instrument Size Introduction

The instrument should be installed at well-ventilated and rational-sized space. Please select appropriate space for installation based on the power supply size.












IT7902-350-10U/IT7904-350-20U/IT7906-350-30 机型

































IT7945-350-270 Model












IT7960-350-360/IT7975-350-450/IT7990-350-540/IT79105-350-630 Model








IT79120-350-720/IT79135-350-810/IT79150-350-900/IT79165-350-990 Model







Chapter2 Instrument Introduction

2.1 Unpacking and Transportation

Unpacking

For cabinet products, they are packaged in wooden boxes at the factory. After you receive them, please refer to the unpacking instructions provided with the box for disassembly; for products packaged in cartons, please use appropriate tools for unpacking.

It is recommended to keep the complete transport packaging for the lifetime of the device for relocation or return to the manufacture for repair.

Transportation

If you need to transport non-cabinet products, you must pay attention to the following to ensure the safety of equipment and personnel.

CAUTION

- Before moving, make sure that the cabinet or stand where the equipment will be placed has been fixed and meets the load-bear- ing conditions to avoid tilting and collapsing, causing personnel to be injured, and equipment broken.
- Due to the weight of the product, transport by hand should be avoided where possible. If unavoidable, carry it with two people and holding the product shell and not external parts (such as han-dles, electrodes, knobs, etc.).
- When carrying, be prepared to bear the weight to avoid sprains or being crushed by heavy objects.
- Use suitable safety clothing, especially safety shoes, when carrying the equipment, as due to its weight a fall can have serious consequences.

After unpacking the cabinet product, if you need to move it to other places, you must pay attention to the following matters to ensure the safety of equipment and personnel.

CAUTION

- The cabinet product is very heavy. Before moving to another loca- tion, confirm whether the ground load is in compliance.
- During the process of moving the cabinet, it is recommended that two
 or more people cooperate and push it slowly and at a constant speed.
 If you encounter a pit, you need to pay special attention. It is forbidden
 to push it quickly, otherwise it will easily cause excessive inertia and
 cause the casters at the bottom of the cabinet to jam and the cabinet
 to fall.
- It is not advisable to push down the slope to prevent the cabinet from falling down due to the shift of the center of gravity. It is rec- ommended



to use a forklift or crane to move the cabinet.

- ITECH 27U and 37U cabinets are equipped with hoisting rings as standard on the top. It is recommended to use a crane equipped with a four-leg hoisting belt structure for horizontal hoisting and moving, and ensure that the four hoisting belts are the same length to avoid cabinet skew during movement. As shown below.
- After moving to the destination, please lock the four casters to se-cure the cabinet.
- The cabinet should be placed on a level ground. It is forbidden to place the cabinet on a slopedground.



2.2 Verifying the Shipment

Open the package and check the articles within package box before operation. In case of any non-conformity, missing or appearance wearing, please contact ITECH immediately.

The package box should comprise:

Device name	Quantity	Model	Remarks
Programmable AC Power Supply	x1	IT7900 series	For the specific models included in this series, refer to 1.2 Models and Options.
Power Cord	X1	-	Number of the power cords vary depending on the model, See the Section 3.1 Connectiong the Power Cord for power cord connection.
USB cable	x1	-	Communication with PC



LAN cable	x1	-	Communication with PC
Calibration Certificate	x1	-	Test report before delivery




After confirming that package contents are consistent and correct, please appropriately keep package box and related contents. The package requirements should be met when the instrument is returned to factory for repair.

The IT7930-350-180 model comes with a yellow optical fiber cable. The fiber cable need to install by yourself. For details, see 3.3 chapter.


2.3 Front Panel

The front panel of IT7900 is as shown below.

2U model front panel:







- 1 Power Switch
- 3 LCD touch screen
- 5 Number key
- 7 Rotary knob
- 2 USB interface /Print/Trig/Menu
- 4 Function key
- 6 Up, down, left and right key and enter key
- 8 Vent hole

2.4 Keyboard

The keyboard introduction of IT7900 series Power Supply is shown as follows.





Keys	Description		
Print	Used for saving screen images		
Trig	Used for manual trigger		
Power	Power Switch		
Menu	Used for going back to menu page		
[Set]	Set the output voltage value		
[F-set]	AC mode: set the output frequency		
	DC mode: [F-set] is invalid		
Config	Enter to Configuration menu		
3	Basic metering key used for basic metering.		
~	Waveform Display key When this key is pressed, the waveform corresponding to current measurement data will be displayed.		
Lim	Harmonic Measurement key When this key is pressed, the harmonic measurement results and the menu of harmonic measurement parameter configuration will be displayed.		
[On/Off]	Turn the power supply output on or off. When lit, indicates that the output is enabled or on.		
Shift	Composite key, combined with other keys to realize functions marked above keys.		
Esc	Press this key to exit the current operation interface.		
[0]-[9]	Number key. Enter the number directly		
+/_	Positive and negative signs		
	Decimal point		
Left / Right	The left and right navigation keys are used to adjust the		
Navigation keys	cursor to the specified position or scrolls pages to view menu items.		
Up / Down	The up and down navigation keys are used to scroll page		
Navigation keys	up and down to view menu items.		
Enter	Operation confirmation key		

Composite key **[Shift]**, combined with other keys to realize functions marked above keys. Firstly, press **[shift]** and the shift key will be lighted, and then press the function key, the detailed functions are listed as follows.



Keys	Description
[Shift]+[Set](List)	Enter the List function menu.
[Shift]+[F-set] (Sweep)	Enter the Sweep function menu.
[Shift]+[Config](Protect)	Enter the protection setting menu.
[Shift]+ (System)	Enter the System setting menu. Used to set the system parameters.
[Shift]+ (Surge&Sag)	Enter the Surge and Sag menu, used to Set the Surge/trapped Configuration.
[Shift]+ (Standard)	Enter the standard wave selection interface.
[Shift]+[1](Log)	Enter the data logging function menu.
[Shift]+ [2] (Lock)	Turn the keyboard lock on or off.
[Shift]+[3] (Local)	Switch remote control mode to local control mode.
[Shift]+[4] (Save)	Save the common parameter settings.
[Shift]+ [5] (Recall)	Enter the Protect menu of the power supply.
[Shift]+ [6] (Hold)	When you need to keep the present meter status, you can press the keys. Then the present meter status display and will be kept no matter whether output is running.
[Shift]+ [7] (Help)	Obtain the help information.

2.5 Push-on Knob

The IT7900 series Power Supply provides a knob on the front panel as shown in the next figure.



The functions of the posh-on knob is described as follows.

- Adjust the value setting
- Select menu item
- Confirm the set value or the selected menu item

Adjust the Value Setting

In the value setting interface, rotate the knob clockwise to increase the set value and anticlockwise to decrease the set value.

Select Menu Item

The knob can also be used to view menu items. In the menu item display



interface, turning the knob clockwise indicates that the next menu item is selected, and turning the knob anticlockwise indicates that the previous menu item is selected.


Confirm settings

After completing the value setting or selecting a menu item, pushing the knob acts like pressing **[Enter]** key to confirm the operation.


2.6 Rear Panel

The rear panel of the IT7900 series 3U model is shown below.

2U model rear panel illustration.



3U model rear panel illustration.




No.	Name	Description
1	ground terminal	Ground screw for making chassis ground connections.
2	Output terminals(M6)	AC or DC output terminals and chassis ground terminals, used to connect DUT.
3	Remote sense Terminals Output lock terminals and	SL+ and SN- are remote sense terminals, used for maximizing measurement accuracy.




4	Optional expansion slot	Optional interfaces: (Plastic plugs are inserted by default when the user does not purchase the interface.) The optional interface as follows: GPIB RS-232/ Analog interface	
5 I/O terminals/CAN interface		Digital Port	
		CAN communication interface CAN-H and CAN-L	
6	LAN interface	LAN communication interface	
7	External control interface CTRL	This interface is used for the parallel connection between the master (with operation panel) and the slaves (without operation panel). Connect the interface on the rear panel of each unit to be connected in parallel, and the master can offer synchronous control over the power-on/off of the slaves.	
8	USB interface	USB communication interface.	
9	system bus	Used for communication between instruments in parallel operation feature.	
10	AC power input socket(M4)	Used to connect AC power to start instrument.	

The rear panel of IT7900 series 15U and above series cabinets are the same. Take IT7945-350-270 as an example, the rear panel is introduced as shown below.





- 1: Output socket, which can be used when users need single-phase lower than 220V/10A power supply
- 2: Output terminals and protective cover, remove the protective cover to show ABCN terminals and sense terminals
- 3: Input terminal, connected to the instrument's power supply
- 4: The communication interface for cabinet, the detailed diagram as follows, and the descriptions are the same as 3U model.





Chapter3 Installation

3.1 Connecting the Power Cord

Connect power cord of standard accessories and ensure that the power supply is under normal power supply.

Before connecting the power cord

To prevent electric shock and damage to the instrument, observe the following precautions.

WARNING


- Before connecting power cord, be sure to confirm that the power voltage matches with the rated input voltage of the instrument.
- Before connecting power cord, be sure to switch off the instrument. Verify that there is no dangerous voltage on the connection terminals.
- To avoid fire or electric shock, Make sure to use the power cord supplied by ITECH.
- Be sure to connect the power cord to the AC distribution box with protective grounding. Do not use terminal board without protective grounding.
- Do not use an extended power cord without protective grounding, otherwise the protection function will fail.
- Ensure that the power cord connection terminals are either insulated or covered by the supplied protective cover so that no accidental contact with lethal voltage can occur.

CAUTION


Safety agency requirements dictate that there must be a way to physically disconnect the AC mains cable from the unit. A disconnect device, either a switch or circuit breaker must be provided in the final installation. The disconnect device must be close to the equipment, be easily accessible, and be marked as the disconnect device for this equipment.

Connecting the power cord

The standard power cord specifications for this series instruments are divided into the following types according to different regions:







The yellow-green wire is grounding wire, which is connected to the PE terminal of power input on the rear panel; the others are live wires, which are correspondingly connected to the L1, L2 and L3 terminals of power input on the rear panel of the instrument.



IT7930-350-180 is standard equipped with two power cords.

AC Power Input Level

Input Voltage Level

The default AC input of this series of instruments is three-phase AC voltage (three-phase four-wire system), some models support single-phase input, please refer to the following table for specific model description.

AC input	requirements	and output	nower	derating.
AC IIIDUL	recontenients	and outbut	DOWEL	ueraunu.

AC Input	1-Phase (V _{LN})	3-phase 4-wire (V _{LL})	3-phase 4-wire (V _{LL})
Input Voltage	(200-220) ±10%	(200-220) ±10%	(380-480) ±10%
2 kVA	1.5 kVA	FS.	FS.
4 kVA	1.5 kVA	3 kVA	FS.
5 kVA	3 kVA	3 kVA	FS.
6 kVA(2U)	1.5 kVA	3 kVA	FS.
6 kVA(3U)	Not support	FS.	FS.
9 kVA	Not support	FS.	FS.
12 kVA	Not support	60%FS.	FS.
≥15 kVA	Not support	60%FS.	FS.

Input Current Level

Users are requested to use the maximum input current specification provided in the specifications when conditions are met. This current value provides a current margin while satisfying the maximum power output. If the user needs to calculate the exact current value under different input voltages, please use the following formula, and it is recommended that you also add current margin to this value. (The apparent power in the formula is



the value at full power output. Different efficiencies at different output powers result in different apparent powers. Please refer to the maximum power and maximum efficiency when calculating.)

AC Input Current (Per Phase) = (Apparent Power/3) / (V_{LL}/1.732)


Taking the IT7915-350-90 as an example, calculate the input current value for the input voltage of **380-10%**. At full power output, Apparent Power is **17.4kVA** in the specification.

AC Input Current (Per Phase) = (17400/3)/((380-380*0.1)/1.732)=29.37A


Connecting the Power Cord

When connecting single-phase AC input

Some models of this series support single-phase input. For single-phase input, please refer to the following figure for the power cord connection method. Single phase input connecting:




Multiple power supplies connect to three phase input:



When connecting three-phase AC input




All models of this product default three-phase power supply, please refer to the following diagram to connect the power cord



Connecting the power cord

- 1. Confirm that the switch of the AC power distribution box is off.
- 2. Confirm that the power switch is in the OFF position and verify that there is no dangerous voltage on the connection terminals.
- 3. Connect one end of the power cable's round terminal to the AC power input terminal on the instrument's rear panel.
 - a) You only need to connect the red/green/yellow (or brown/black/gray) live wires to the terminals on the rear panel, which are not required to correspond to L1, L2 and L3 terminals one by one.
 - The yellow-green wire is grounding wire, which is connected to the protective grounding terminal (PE).



4. Refer to the suggestion connection diagram, connect the other end of the power cable to the required AC distribution box.

Connect the two power cables of the IT7930-350-180 to the power distribution box.

3.2 Connecting Test Lines (Optional)

Test lines are not standard accessories of the instrument. Please select optional



red and black test lines for individual sales based on the maximum current value. For specifications of test lines and maximum current values, refer to "Specifications of Red and Black Test Lines" in "Appendix".

WARNING

- Before connecting test lines, be sure to switch off the instrument.
 Power switch is in Off position. Otherwise, contact with output terminals in rear panel may cause electrical shock.
- To avoid electrical shock, before testing, please make sure the rating values of the testing lines, and do not measure the current that higher than the rating value. All test lines shall be capable of withstanding the maximum short circuit output current of the power supply without causing overheat.
- If several loads are provided, each pair of load wires shall safely withstand the rated short circuit output current of the power supply under full load.
- Always use test lines provided by ITECH to connect the equipment.
 If test lines from other factories are used, please check that the test line can withstand maximum current.

Specification for Test Cables

Test cables are not standard accessories for the instrument. Please select optional red and black test cables for individual sales based on the maximum current value. For specifications of test cables and maximum current values, refer to A.1 Specifications of Red and Black Test Cables for more information.

Connecting the DUT (Local Measurement)

The instrument supports two kinds of wiring methods with the DUT: local measurement and remote measurement (SENSE). The default test mode is local measurement.




The T7930-350-180 consists of two 15kVA in parallel mode. The output terminals and Remote sense terminals use the master interface. The wiring method is the same as other models.

IT7900 series 2U model rear backplane terminal appearance is different from 3U model, but the wiring principle is the same, the following content and diagram to 3U model as an example. 2U model corresponding to the output terminals can be.

Please confirm that the Remote Sense function in the menu is set to Off, otherwise the instrument will report an error in the present connection mode.

The connection diagram of single phase is shown as follow:






NOTE


When the output voltage has DC voltage, the output terminal L is positive, and N is negative.

• The connection diagram of three phase is shown as follow:



The above diagram shows Y-type wiring as an example, if the user needs \triangle -type wiring, please directly suspend the N terminal can be.

• The connection diagram of reverse phase is shown as follow:




NOTE

When the output voltage has DC voltage, the output terminal A is positive, and B is negative.

The connection diagram of multi-channel is shown as follow:







NOTE

When the output voltage has DC voltage, the output terminal L is positive, and N is negative.

- 1. Confirm that the power switch is in the OFF position and verify that there is no dangerous voltage on the connection terminals.
- 2. Remove the output terminals cover of the power system.
- 3. Loosen the screws of the output terminals and connect the red and black test cables to the output terminals. Re-tighten the screws.

When maximum current that one test cable can withstand fails to meet the rated current, use multiple pieces of red and black test cables. For example, the maximum current is 1,200A, then 4 pieces of 360A red and black cables are required.

- 4. Thread the red and black test cables through the output terminals cover of the power system and install the cover.
- 5. (Optional) According to the actual situation of DUT, connect the grounding terminal on the rear panel of the instrument to the DUT to ensure the safe grounding.

For the location information, see 1.5 Rear Panel Introduction.

6. Connect the other end of the red and black cables to the DUT. The positive and negative poles must be properly connected and fastened when wiring.

Connecting the DUT (Remote Sensing)

Remote measurement is available for the following scenarios:

When the DUT consumes large current or the wires are too long, there is a voltage drop on the wires between DUT and output terminals of the power system.

To maximize measurement accuracy, the power system provides the remote measurement terminals VS+ and VS- on the rear panel, which can be used to measure the terminal voltage of the DUT.

When the power system is used for battery testing in actual applications, the voltage drop of the wire will lead to voltage inconsistency of both ends and




inconsistency of the cutoff voltage of power system and the actual voltage of battery, resulting in inaccurate measurement.

The connection diagram and steps of remote measurement are as follows:



The T7930-350-180 consists of two 15kVA in parallel mode. Connect the Remote sense terminals of master module. The wiring method is the same as other models.


The connection diagram of single phase is shown as follow:



₩ _{NOTE}


When the output voltage has DC voltage, the output terminal L is positive, and N is negative.

• The connection diagram of three phase is shown as follow:





The connection diagram of reverse phase is shown as follow:



O NOTE

When the output voltage has DC voltage, the output terminal A is positive, and N is negative.

- 1. Confirm that the power switch is in the OFF position and verify that there is no dangerous voltage on the connection terminals.
- 2. Remove the output terminals cover of the power system.
- 3. Refer to the wiring diagram and connect the Vs+ and Vs- with armored twisted-pair cables. Loosen the screws of the output terminals and connect the red and black test cables to the output terminals. Re-tighten the screws.

When maximum current that one test cable can withstand fails to meet the rated current, use multiple pieces of red and black test cables. For example, the maximum current is 1,200A, then 4 pieces of 360A red and black cables are required.

- 4. Thread the red and black test cables through the output terminals cover of the power system and install the cover.
- 5. (Optional) According to the actual situation of DUT, connect the grounding terminal on the rear panel of the instrument to the DUT to ensure the safe grounding.

For the location information, see 1.5 Rear Panel Introduction.

- 6. Connect the other end of the remote sense cables to the DUT.
- 7. Connect the other end of the red and black cables to the DUT. The positive and negative poles must be properly connected and fastened when wiring.
- 8. Power on the instrument and turn on the Sense function of the instrument.

3.3 Installing Fiber Cables (Only for IT7930-350-180)


IT7930-350-180 model is composed of two 15kVA parallel machines. In order to avoid damage to the fiber cords connected in parallel with the master and slave during transportation, the fiber cords are not installed before delivery. After receiving the instrument, users need to install the yellow fiber cable by themselves.



CAUTION

- Before connecting the cables, ensure that the instrument power switch is off and the main switch of the AC power input (distribution box) is off.
- Fiber optic cables cannot be flexed or folded. When the cable is too long and needs to be arranged, gently wrap the cable in a circle and gently tie it.

As shown in the following figure, connect the System Bus between the master and slave by fiber cords.





Chapter4 Getting Started

4.1 Power-on the Instrument

A successful selftest indicates that the purchased power product meets delivery standards and is available for normal usage.

Before operation, please confirm that you have fully understood the safety instructions.

Precautions

To prevent electric shock and damage to the instrument, please observe the following precautions.

WARNING

- Before connecting power cord, be sure to confirm that the power voltage matches with the supply voltage.
- Before connecting power cord, be sure to switch off the instrument.
 Verify that there is no dangerous voltage on the terminals before touching them.
- To avoid fire or electric shock, make sure to use the power cord supplied by ITECH.
- Be sure to connect the main power socket to the power outlet with protective grounding. Do not use terminal board without protective grounding.
- Do not use an extended power cord without protective grounding, otherwise the protection function will fail.
- Ensure that the input electrodes are either insulated or covered using the safety covers provided, so that no accidental contact with lethal voltages can occur.
- If you notice strange sounds, unusual odors, fire, or smoke around or from inside the instrument, flip the POWER switch to the (O) side to turn the instrument off, or remove the power cord plug from the outlet. The detachable power cord may be used as an emergency disconnecting device. Removing the power cord will disconnect AC input power to the unit.

CAUTION

Safety agency requirements dictate that there must be a way to physically disconnect the AC mains cable from the unit. A disconnect device, either a switch or circuit breaker must be provided in the final installation. The disconnect device must be close to the equipment, be easily accessible, and be marked as the disconnect device for this equipment.



Power Switch Introduction

User can adjust the power switch directly to turn on or turn off the instrument. The status of Power switch is as follows.

The switching knob of the IT7900 series power supply allows the user to turn the power on by 90° clockwise or to turn the power off by 90° counterclockwise.



Turning the POWER Switch On

Check that the power cord is connected properly.

Flip the POWER switch to the (ON) side to turn the instrument on. The front panel display will light up after a few seconds. It may take about 30 seconds or so for the power supply to initialize before it is ready for use.

If a self-test error occurs, an error message will be displayed in the front panel. Press the **[Esc]** button to try to clear the current fault status. The user can also restart the instrument to try to clear the fault status. Wait until the power is turned off and then start again. If the problem still cannot be solved after restarting, please contact the ITECH engineer.

Turning the POWER Switch Off

Flip the POWER switch to the (OFF) side to turn the instrument off. When it is turned off, the instrument interface will prompt "Power Down", and the instrument will store the setting information before shutdown in the group 1 nonvolatile memory.

After you turn the POWER switch off, wait at least 10 seconds after the fan stops before you turn the POWER switch back on. Turning the instrument on too soon after you turn it off can cause damage to the inrush current limiter circuit, as well as reduce the life of components such as the POWER switch and the internal input fuses.

4.2 Home-Screen Overview


IT7900 series power supply adopts touch screen design, the users can easily operation by touch screen.

The power supply can work in either single-phase mode or three-phase mode. The working mode can be set in the system menu. The display interface of the instrument is different under different modes.



Single Phase Mode

The meter interface of IT7900 series power supply is shown as follow.




Display area of measuring data

Three Phase Mode

The meter interface of IT7900 three phase mode is shown as follow.

The parameters are the same as the single phase.



Under three-phase mode, display total parameters corresponding to Phase A, B and C, the user can switch screen and view the total parameters of each phase.



The meter interface of phase A is shown as follow, the display parameters are the same as single mode. You can click the A, B or C to switch the phase interface



Introduction to Interface Symbols

The interface of IT7900 power supply will display the following symbols. All the symbols and description are listed in the table below.

Char	Function description	Char	Function description
Shift	Compound key	OFF	Output is off
U) ON	Output is on	LOCK	Key operation is locked
1Ø	Single phase AC mode	3Ø	Three phase AC mode
10	Single phase DC mode	3Ø	Three phase AC+DC mode
AC	Split phase AC mode	\mathbb{R}^{2}	Split phase AC mode
AC+DC	Split phase AC+DC mode	DC+AC	Split phase DC+AC mode
1Ø	Single phase DC+AC mode	1Ø	Single phase AC+DC mode
AC	Multi-channel AC mode	AC+DC	Multi-channel AC+DC mode
DC	Multi-channel DC mode	DC+AC	Multi-channel DC+AC mode
D D REM	The AC source is in remote mode	SAG	Surge&Sag function indicator
-WW-	External Simulation Test Function		Record log




Char	Function description	Char	Function description
LIST	LIST is running	LIST	LIST is finished
LIST	LIST function is waiting for trigger	SWEEP	Sweep function is waiting for trigger
₹ SWEEP	Sweep running indicator	USB	Found USB disk
SENSE	Sense indicator	SENSE	Sense Error
OV	Over voltage protection	OTP	Over temperature protection
PEAK	AC/AC+DC mode Peak OCP	A	AC/AC+DC mode RMS OCP
POWER	Over power protection	FIBERS	Fiber is error
SYNC	SYNC unlock	SYNC	SYNC lock
v(4)	Current Negative	V SAMPLE	Voltage sample mode
 	Unbalanced current	INHIBIT	Inhibit output
osc	Input oscillation protection	PEAK	Over voltage peak protection
POWER	Negative power limit protection	O V	Under voltage protection
WDG	Communication watch dog protection	VE	Voltage error
«	Click to collapse the present screen	ERROR	Command error
>>	Tip image is collapsed, click to expand all icons	10.101 SIMU.	Voltage signals simulation

Menu Introduction

Press the **[Menu]** key on the front panel and enter to the menu interface. Menu interface will display all of function icon, user can rotate the knob or direction key to select, or click the screen to enter the function interface.






4.3 Touch Screen Introduction

This series of power display is a touch screen LCD interface, users can select and set parameters by hand touch. The touch function can be set in the system menu.

Touch screen drag and corresponding functions are described as below. Convenient user quickly understand how to use this power supply.



4.4 Set output parameters

The voltage value and frequency value can be programmed, which can be set to different parameters within the specification range based on customer requirements. This can meet various test requirements of the customer.

After the user presses the V-set or F-set keys on the front panel, the instrument interface displays the parameters to be set and the cursor flashes for prompt. The user can use the following methods to set the values.

- Directly use the number keys to set the value.
- Rotate the knob to set the data in the cursor position. Rotate the knob lockwise to increase the set value and anticlockwise to decrease the set value. Once the data in the cursor position increases to ten, the value will add one to the front position automatically. and once the data in the cursor



position decreases to zero, the value will minus one from the front position automatically. This provides convenience for the user to set. The knob can works with the left or right keys. Use the left or right keys to move the cursor position.

■ NOTE

After entering the menu interface, the knob can also be used to scroll pages to view menu items.

4.5 Output On/Off Control

WARNING

- The [On/Off] key is used to turn the output on or off under normal circumstances. Even if the instrument is in control by PC or the keyboard is locked, the [On/Off] is still valid.
- The [On/Off] key light is off and turning the output off does not place the instrument in a safe state. Hazardous voltages may be present on all output and guard terminals. Putting the equipment into an output-off state does not guarantee that the outputs are powered off if a hardware or software fault occurs. See the cautions about connecting the test lines before connecting test lines.

Controlled by the [On/Off] key

You can press the **[On/Off]** key on the front panel to control the output status of the power supply. If the **[On/Off]** key light is on, indicates that the output is turned on. The VFD displays the meter value such as voltage, current, power and so on. If the **[On/Off]** key light is off, indicates that the output is turned off. The VFD displays that the power supply state is OFF.



It is recommended that you turn on the [On/Off] after the power supply is connected to the DUT. If the power supply has no output after the output is turned on, check the voltage and current setting value, set the voltage and current to a non-zero value, and then turn on the output.

Controlled by remote command

When communicating remotely, the [On/Off] switch of the power supply or load can be controlled by using the corresponding SCPI command, as detailed in the programming guide. For example, use OUTP 1 to turn on the power output and OUTP 0 to turn off the output.

Controlled by Digital IO pins


The Digital IO-1 pin, which comes standard with this series of instruments, supports external level/pulse signal control output, and in combination with external circuitry, enables emergency stop control of the output. Refer to the following use case.

Digital IO-1 with the function of inhibit output, can disable the output by inputting low level (0V) to IO-1 in the default setting. at which time the [**On/Off**] key light is on and there is no actual output. Restoring a high level (5V) on IO-1 restores the output on.

1. Go to the System→I/O menu.



- 2. Select the Digital IO-1:Remote Inhibit Input items, and setup the configurations.
- Reverse: set to Off it means the valid signal is low level (0V).
- Function: set to Inhibit-living it means when input an inhibit signal and the instrument output is inhibit. When the input signal undoes, the output returns to normal.
- 3. Connect Pin 1 (positive) and Pin PE (negative) of Digital IO to the external signal as follows:



At this time, 5V is output between Pin1 and Pin PE.

- 4. After connecting the DUT, turn on [On/Off] button.
- 5. Input 0V to Pin1 (positive) and Pin PE (negative), or directly short Pin1 and Pin PE. At this point, output is disabled.
- 6. Input 5V to Pin1 (positive) and Pin PE (negative), or disconnect the short wire between Pin1 and Pin PE. At this point, output is restored.




Chapter5 Operation and Application


This chapter describes operations of the keys on the front panel of the AC power supply. The AC power supply must be in the local mode when controlled by the front panel. The default mode is the local mode after the AC power supply is turned on. In the local mode, the user can enable all functions of the power supply through the front panel.

5.1 full 4-Quadrant output character

The IT7900 series represent the newest generation of programmable, full four-quadrant grid simulators with full 100% of current rating in both source and sink mode, and provides energy recovery capability. The power generated by the DUT during the test can be easily retuned to the AC grid, rather than being dissipated as heat. which protects the environment and lowers the cost of operation. It can be applied to the applications in testing grid-tied products that inject energy into the grid, such as the frequency changes, voltage transients and DC injection of grid-tied/off-grid inverters testing.



The relationship between output voltage, current, power and frequency is shown in the figure below.





5.2 Select the Power Supply Mode

The IT7900 series provides multiple modes such as single-phase, three-phase, reverse phase and multi-channel, which can be selected by the user through the panel menu.

Under the reverse mode, the rated voltage will be extended to 200%.

The multi-channel function allows users to test 3 independent DUT at the same time without adding additional hardware configuration.

The operation steps are as follows.





(System) enter to system menu.

2. Under the Source setting interface, touch the screen or rotate the knob to select the **Phase** and set the power supply phase mode.

Single Phase

Under the single phase, the power supply works as a single phase power supply. the output mode can be set to AC/DC/ACDC/DCAC.

Three Phase

Under the three phase mode, the power supply works as a three phase power supply. the output mode can be set to AC/ACDC.

Reverse Phase

Under the reverse phase, the power supply works as a single phase power supply, and the rated voltage will be extended to 200%. For example, AC power supply rated voltage is 350V, under the reverse phase mode, the reted output voltage is 700V. And the output mode can be set to AC/DC/ACDC/DCAC.

Multi-channel

Under the Multi-channel phase, the power supply works as a single phase power supply and have three channels.

The multi-channel function of the IT7900 series allows users to test 3 independent DUTs at the same time without adding additional hardware configuration. In the traditional solution, the user needs to configure 3 AC power supplies; and one IT7900 device can meet multi-channel testing requirements. Under multi-channel mode, parameters of each channel can be set independently. The output mode can be set to AC/DC/ACDC/DCAC.

5.3 Select the Output Mode

The IT7900 series has four output modes: AC, DC, AC+DC, DC+AC. It not only provides pure AC/DC output, but also can use AC+DC and DC+AC output modes to realize "AC output plus DC bias" And "DC output waveform with ripple", which cover a wider range of applications.

The output mode can be select in the system menu.





(System) enter to system menu.

2. Under the Source setting interface, touch the screen or rotate the knob to select the **[Output couple mode]** and set the output mode.



5.3.1 AC Output Mode

If the output mode select to AC Mode, the instrument will simulate AC power supply. The default set of IT7900 series power supply is AC Mode.

Set the output parameters of the power supply in the main interface, including the output voltage, output frequency.

- Press up/down keys to select setting value and then press Enter to confirm.
- Rotate the knob to select setting value and then press **Enter** to confirm.
- Touch screen and then press **Enter** to confirm.



5.3.2 DC Output Mode

If the output mode select to DC Mode, the instrument will simulate DC power supply. Set the output voltage and current of the DC power supply in the main interface, as shown in the figure below.

- Press up/down keys to select setting value and then press Enter to confirm.
- Rotate the knob to select setting value and then press Enter to confirm.
- Touch screen and then press **Enter** to confirm.



□ NOTE

When using DC mode for testing, please know the DC ripple parameters in DC mode in advance, if there are strict noise requirements, you need to add extra DC noise filters to get low noise and good stable DC voltage for testing.



5.3.3 AC+DC Mode

If the output mode select to AC+DC Mode, the instrument will simulate AC and DC power supply, which can add DC component to AC voltage. Set the output voltage in the main interface, as shown in the figure below.



Under AC+DC mode, Set the output voltage and frequency of the AC power supply in the main interface, and set the DC voltage in the configure menu.

Vac: you can set the Vac under the main interface or under the config interface.

DC: set the DC component under the config interface.




The AC + DC mode not only expands the application range of pure AC voltage, but also expands the application range of DC component in laboratory testing. When using AC+DC for testing, please first understand the ripple parameters of the instrument when it is used as a DC power supply. If there are strict noise requirements, additional DC noise filters are needed to obtain low noise and good stable DC voltage for testing.

5.3.4 DC+AC Mode

If the output mode select to DC+AC Mode, the instrument will simulate DC and AC power supply, which can add AC component to DC voltage. Set the output voltage in the main interface, as shown in the figure below.





Under DC+AC mode, Set the Vdc in the main interface, and set the Vac in the configure menu.

Vdc: you can set the Vdc under the main interface or under the config interface.

Ripple control: set the wave, Vac component and frequency under the config interface.



When using DC mode for testing, please know the DC ripple parameters in DC mode in advance, if there are strict noise requirements, you need to add extra DC noise filters to get low noise and good stable DC voltage for testing.

5.4 Current Limit Mode and Power Limit Mode

The IT7900 series power supply defaults to constant voltage CV output mode. The output voltage can be set in main interface.

When the actual current value is higher than the setting current limit value, power supply works in current limit mode, and output voltage will be reduce.

When the actual power value is higher than the setting power limit value, power supply works in power limit mode, and output voltage and current will be adjust.

The current limit value can be set in the Protection menu. Refer to 6.6 Protection Function for detailed setting methods.

MOTE

Current limited or power limited operation mode is only valid when pulling load with constant resistance, please contact ITECH engineer for other test situation.

5.5 Professional Anti-islanding Test Mode

The IT7900 series provide a professional anit-islanding test mode. Users can adjust RLC parameters or configure the parameters of active power and reactive power to simulate islanding status between inverter and load when the main network is off, and verify the anti-islanding protection response time of grid-tied DUTs under different equivalent impedances, three-phase load balancing and non-balancing conditions.

The solution helps engineers to simplify the test circuit and cost savings of




additional equipment such as oscilloscopes and power meters.

1. Click the RLC function in Menu interface and enter to islanding simulator function.



2. Select the mode.



- PQ mode: configure the parameters of active power(P), reactive power Q_L and Q_C .
- RLC mode: configure the parameters of R, L, C.
- 3. Set the Phase, it means the output stop phase angle when close the K1 switch.
- Users can click **More** to enter the menu setting interface to set the test method.





The IT7900 Grid Simulator provides two end-of-island state detection methods:

- The Inhibit control enable parameter is set to Enable, the inverter outputs a signal after it is turned off, and the IO-1 pin of the IT7900 grid simulator receives the signal and ends the islanding simulation state, the K1 and K2 switches are disconnected, and the On/off switch is turned off. This stopping method requires that the inverter has the function of outputting a signal after disconnection. The inverter output signal is connected to the IO-1 pin and set to the default function.
- The other is to set the detected parameter information in A/B/C Phase.
 IT7900E Grid simulator detects that the input voltage/current/power is less than the set value, then it ends the islanding simulation state, the K1 and K2 switches are disconnected, and the On/Off switch is closed.
- 5. Close the K2 switch, connecting the load and UUT, and then close the K1 switch, power on the UUT.
- 6. Open the K1 switch, enable the islanding status. Test the anti-islanding protection response time, The interface will display the running time value in the island state.

After the K1 switch is disconnected, the machine can output a voltage signal externally as a reference point for the island start time, this signal needs to be connected to the IO-6 pin to get it, and the pin function is defined as Trigger1-out.

5.6 Waveform Selection

The user can set the output waveform in the config menu of IT7900 series power supply. Eight output waveforms below are available, user can select the waveform in **Config->Waveform** menu.





- Sine
- Square
- Sawtooth
- Triangle
- Trapezoid
- Clipped-sine
- Rectifier
- THD wave
- User-defined

When Trapezoid, Clipped-sine, Rectifier, Rectifier, THD Wave and User-define are selected, the user should configure the parameter of waveform.

5.7 Full 4-Quadrant Power Amplifier


The IT7900 series regenerative grid simulator can be used as a power amplifier to complete power hardware in the loop (PHIL) applications for microgrids, energy storage and new energy vehicles. The digital I/O or a standard suite of analog signal can be input via an external analog interface (optional) and then amplified without distortion to a real power waveshape.

The power amplification function can only be realized by selecting external analog interface. Please refer to the Analogue function in 6.14 for the specific use method.

5.8 Auto-range Function

IT7900 series power supply can achieve the combined output of multiple voltage and current at a fixed power. Single power supply can meet different DUT tests with high voltage low current or high current low voltage, at the same time, because the output of voltage and current is controlled by the limit power, it will show the switching of voltage and current auto ranging. To avoid Intermittent output shutdown when switching the high range and low range.





I-V Curve Graph

5.9 Programmable Output Impedance

The function of programmable output impedance allows user to edit the output R and L so as to simulate the impedance of the AC grid.

5.10 Sweep Function

The Sweep function is used to test efficiency of switching power supply and capture the voltage and frequency at the maximum power point. The voltage and frequency of power may be altered in the form of step ladder by setting the initial voltage, final voltage, step voltage, initial frequency, final frequency, step frequency and one-step time. As the test closes, voltage and frequency at the maximum power point may be displayed. Sweep function is not applicable under DC mode and DCAC mode.

The Sweep function is not supported in multi-channel mode.

Operating steps

1. Press [Shift] + [F-set](Sweep)on the front panel to enter the sweep interface, as shown in the figure below.



2. Set the voltage and frequency parameters in the sweep interface.



Parameters in the sweep interface are described as follows:

Parameter	Explain
Start voltage	Set the starting voltage.
Stop Voltage	Set the ending voltage.
Step Voltage	Set the step voltage.
Start freq	Set the starting frequency.
Stop freq	Set the ending frequency.
Step freq	Set the step frequency.
Mode	Set the sweep mode. Time: Time Sweep Mode
	Trig: Trigger mode
	Time-back-forth: Step switching according
	to time and scanning back and forth.
	Trigger-back-forth: Step switching
	according to trigger and scanning back and
	forth.
Step time	Set the step time.
Repeat Count	Set the sweep repeat count.
Priority	Priority setup
	Volt-Priority
	Freq-priority
	Volt&Freq
Waveform	Select sweep waveform
	● Sine
	Square
	• Saw
	Triangle
	Trapezoid
	Clipped-sine
	Rectifier
	● THD
	User-defined
Finish	Set the running state after the sweep
	execution is finished:
	off: Directly off the output after the
	execution is finished;
	Last: Keep the last waveform unchanged
	after the execution is finished.
	Normal: return to normal mode after the
T.::	sweep execution is finished.
Trig source	Select the Trigger source:
	Manual BUS
	Trigger1
	Trigger2
	11199012

- 3. Press [On/Off] on the front panel and enable the output.
- 4. Press [Run] on the Sweep interface, Then the running status will be displayed on LCD. Output parameters and measurement parameters are



displayed on the left side of the Sweep interface. You can also press the Meter to observe the output parameters in the main interface.

5. After sweeping, **[On/Off]** on the front panel will be off, and status will be displayed on LCD. You can press **[Stop]** on the Sweep interface to stop the Sweep function.

5.11 Polyphase Function

The IT7900 series power supplies implement the frequency-locked and phase-locked function between power supplies through the digital IO interface or System bus interface, which achieve 6 phase 12 phase power output.


This power supply provides two ways to achieve multi-phase functionality:

- One is to connect the System bus interface between instruments directly through fiber optic cable, no other connection is needed, the wiring is simple and convenient, this wiring method requires customers to purchase fiber optic module and fiber optic cable separately.
- The other is via a digital IO interface, which requires no additional accessories and the instrument is connected by default.

Take the simulated 6-phase power supply as an example, and introduce the multi-phase function application method.

5.11.1 Connecting the DUT

When two IT7900 power supplies are combined in multiple phases, the outputs ABC are connected separately and the N wires of the two IT7900 power supplies need to be connected together as six-phase N wire terminals. The wiring schematic is shown below.



5.11.2 Connecting the System Bus Method


The two IT7900 power supplies are connected via the outer ring interfaces TX and RX in the System bus. Fiber optic connection method customers need to




purchase fiber optic modules and fiber cables.

Connect the System Bus as shown in the diagram below for fiber optic communication between instruments.

1. Insert the fiber optic module into the hole corresponding to TX RX.



2. Insert the plug of the fiber optic cable into the fiber optic module and hear a click sound to indicate that it is inserted in place. The fiber optic cable connection schematic is as follows.



Power Supply Configuration

- 1. Press [Shift] + (System) and enter to the system menu interface.
- 2. Set the role of fiber control in the **Source** settings menu, you can choose one set as the Fiber-Phase-Master, the other set as the Fiber-Phase-Slave, and the slave can choose the way to synchronize.

External Lock- frequency control	Set the external frequency lock function.	
	Status	Enables or disables the external frequency lock function. Off/On Lock-Freq Lock-Phase Fiber-Phase-Master Fiber-Phase-Slave



Phase I	Delay Set the phase deviation between the output phase and the external I/O input signal: 0-360°.
Freq lim	Set the frequency difference upper limit between the output frequency and the external I/O input signal.
Freq lim	Set the frequency difference upper limit between the output frequency and the external I/O input signal.
Exception	Set the output mode when the frequency lock fails: Output Disable: stop output Limit: Output according to the set frequency.
Sync se	Synchronization mode selection (settable only in fiber slave mode). Local: The slave parameters take effect in real time. Fiber: The slave is controlled by the master fiber.

- First instrument selection: Source -> External Lock-frequency control -> Status -> Fiber-Phase-Master
- Second instrument selection: Source -> External Lock-frequency control
 -> Status -> Fiber-Phase-Slave。

And then select **Sync select** to Local or Fiber.

When **local** is selected, the slave's amplitude setting and modification will take effect in real time, without waiting for the host's command.


When **Fiber** is selected, the slave's amplitude setting does not take effect immediately and needs to wait for the host's command.

When the host changes the parameters or changes the on/off status, the value set by the slave will take effect.

5.11.3 Connecting Digital I/O Method

Connect the IO-4 pins of two IT7900 power supplies. One IT7900 outputs synchronization signals to the other IT7900. The second power supply runs synchronously with the frequency and phase of the first power supply to realize the six-phase output function. The wiring diagram is as follows:

First Second



Power Supply Configuration

Defining the digital IO-4 Pin

User can define the IO function in IO configure menu.



- 3. Press [Shift] +
 - (System) and enter to the system menu interface.
- 4. Select I/O -> Digital IO-4: SYNC.

Set one IT7900 power supply as a synchronization signal output to **Syncout**, and the others set to **Sync-out**.

【Sync-in】: Synchronous input function, which is used to output frequency lock or phase lock with the external signal. At this time, the machine synchronizes the frequency or phase information input from the IO-4 pin.

【 Sync-out 】: Synchronous output function, the IT7900 produces synchronous signal to the outside, which is AC zero crossing pulse signal sent from the IO port.

- 5. Set the IO-5 function of the first power supply. Select I/O -> Digital IO-5: On-off Status, set Reverse to Off, and set Function to OnOff-status.
- 6. Set the IO-1 function of the second power supply. Select I/O -> Digital IO-1:Remote inhibit input, set Reverse to On, and set Function to Inhibit-Living.

□ NOTE

The IO-1 function of the second power supply must set according as above. Otherwise, the polyphase will be error.

 Frequency and phase lock configuration set frequency and phase synchronization Settings for the second power supply.

On the **Source** Settings menu, select **External Lock-Frequency Control** -> **Status**-> **Lock-Phase**.

->Status->Lock-Pilase.			
External Lock- frequency control	Set the external frequency lock function.		
	Status	Enables or disables the external frequency lock function. Off/On Lock-Freq Lock-Phase Fiber-Phase-Master Fiber-Phase-Slave	
	Phase Delay	Set the phase deviation between the output phase and the external I/O input signal: 0-360°.	
	Freq limit+	Set the frequency difference upper limit between the output frequency and the external I/O input signal.	
	Freq limit-	Set the frequency difference upper limit between the output frequency and the external I/O input signal.	
	Exception	Set the output mode when the frequency lock fails: Output Disable: stop output Limit: Output according to the set frequency.	
	Sync select	Synchronization mode selection (settable only in fiber slave mode). Local: The slave parameters take effect in real time.	



	Fiber: The slave is controlled by the
	master fiber.

If in balance mode, the Phase Delay between the first instrument and the second intrument is set to 60° to achieve six-phase balanced output of 60° between six phases.

5.12 Standard IEC Regulations

IT7900 series AC/DC power supplies provide standard test curves in accordance with IEC 61000-4-11/4-13/4-14/4-17/4-28/4-29 regulations. It can be invoked directly by the user when testing IEC compliance tests.

This function provides both the test curve that meets the standard requirements of regulations and the curve customization function. Users can customize the curve according to the project requirements of regulations and perform the extended test of the test object.

The IEC 61000-4-11/4-13/4-14/4-28 regulation waveforms can be executed in AC voltage source mode, the IEC 61000-4-17 regulation waveforms can be executed in DC+AC voltage source mode, and the IEC 61000-4-29 regulation curve can be executed in DC or DC+AC voltage source mode.

Click Standard on the Menu screen directly press Shift+ to enter the regulation test screen.






For parameters of corresponding curves and related regulations, refer to IEC standards.

Interface Instruction

Take IEC 61000-4-11 as an example to introduce the interface display and how to use the regulation test. The usage method of other regulatory interfaces is the same, but the displayed parameters are different.





Standards Selection Area

In the Standards selection area, users can select the regulation to be tested. Each option is described as follows:

Standards	Select the regulations to be tested.		
	the following regulations are supported:		
	AC voltage source mode:		
	● IEC 61000-4-11		
	• IEC 61000-4-13		
	• IEC 61000-4-14		
	• IEC 61000-4-28		
	DC+AC voltage source mode:		
	• IEC 61000-4-17		
	• IEC 61000-4-29		
Category	Test items defined by standard regulations. The third class projects are divided into several sub-projects according to the requirements defined by regulations, and customers select them according to the level of needs.		
	Voltage dips		
	short interruptions		
	Voltage variations		
Voltage	Voltage value of AC output. User can setting the voltage level according to the DUT requirements.		
Frequency	Frequency value of AC output. Different regulatins have different definitions. This parameter can be set only for User defined categories.		
Phase	Select the phase to execute the test. If not selected, the test as defined by the regulation will not occur.		

Standards Edit Area

If you select different regulations, the execution items and parameters in the regulation editing area are different. You can edit parameters such as the start Angle of test execution. The parameters displayed in green on the page can be modified.

This section uses IEC 61000-4-11 as an example to describe the regulatory parameters. For other regulatory parameters, see the regulatory documents.

Level %	Level of test items defined by the regulation.
Phase	The starting phase Angle position of the A-phase waveform output.
	For three-phase mode, the Angle difference between PHASE B and A is 120°, PHASE C and A is 240°
Cycle	The number of continuous periods for maintaining the drop voltage, According to regulations, different test items have different cycles.



Interval	The interval time of test, from the beginning of the drop to the recovery of normal operation. Unit: second (s)
Repeat counts	Repeat number.
Delay	Time delay, the time interval between test items. Unit: second (s)
Selected	Test item selection: Select to Yes, indicates the test item will be executed. Select to No, indicates the test item is not executed.

Configure

Click Configure to enter the configuration interface of other parameters, and set the Rise time and Fall time.

Run/Stop

running or stopping the regulation tests. You must enable the output before running. Otherwise, a message is displayed indicating that the output is not enabled.



Chapter6 System-Related Functions

6.1 System Menu Reference

Press [Shift] + (System) to enter the menu function. At this time, LCD displays optional menus. Select and edit the menu items by pressing the Up, Down, Left and Right keys. Specific menu items are shown below.

Voltage Source Menu:

System				
.,	Device operation mode	Select the operat	tion mode to Voltage Source.	
		Volt Source: Voltage source mode Current Source: current source mode		
	Phase mode	Set the power supply mode		
		1-Phase	Single mode	
		3-Phase	Three phase mode	
		Multi-Channel	Multi-channels mode	
		Reverse	Reverse phase mode	
	Output couple mode	Set the output mo	·	
		AC	AC mode	
		DC	DC mode	
		ACDC	AC+DC mode	
		DCAC	DC+AC mode	
	Output impedance	Set the output im	<u> </u>	
		Status	Enable or disable this function	
		R	Set the output resistance	
Source		L	Set the output inductance	
	Off mode	Set the output turn off mode		
		Open-Z	Open circuit mode	
		High-Z	High impedance mode	
		Short	Short circuit mode	
	Loop speed	Output loop speed control		
		High	High speed	
		Low	Low speed	
	External Lock- frequency control	Set the external Lock-frequency control mode		
		Status	Set the Lock-frequency ON/OFF state Off/On Lock-Freq Lock-Phase Fiber-Phase-Master Fiber-Phase-Slave	
		Phase Delay	Set the phase deviation between the output phase and the external I/O input signal: 0-360°	



IIII			System-Related Functions
		Freq limit+	Set the upper limit of output frequency.
		Freq limit-	Set the lower limit of output frequency.
		Exception	Set the output mode when the frequency lock fails: Output Disable: stop output Limit: Output according to the set frequency.
		Sync select	Synchronization mode selection (settable only in fiber slave mode). Local: The slave parameters take effect in real time. Fiber: The slave is controlled by the master fiber.
	External programme	External analog fu	nction:
		Status	Set the ON/OFF state
		Mode	AM: Adjust the amplitude Amplifier: Real-time output and power Amplifier.
		Monitor phase	Select the phase to monitor. This cannot be set in single phase mode.
		U ratio	The external program ratio of voltage.
		I ratio	The external program ration of current.
	Remote sense	Set the sense fund	tion state.
		Mode	On: Enable the remote sense Measurement Off: disable the sense function
	On/Off phase	On-mode	 Output on phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope
		Off-mode	 Output off phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope
	Measurement	Set the measurem	ent speed
		Lower (1000ms)	Lower mode, measuring every 1000ms
		Slow (500ms)	Slow mode, measuring every 500ms
		Medium (300ms)	Medium mode, measuring every 300ms
		Fast (150ms)	Fast mode, measuring every 150ms
		Filter	Whether to turn on the filter during measurement.
	Power Unit Setting	Unit of power, can	be set to KW/kVA/KVar or W/VA/Var.



Current Source Menu:

System	T		
	Device operation mode	Select the operation mode to Current Source.	
		Current Source	
	Phase mode	Set the power supply mode	
		1-Phase Single mode	
		Reverse	Spilt phase mode
	Output couple mode	Set the output mode	
		AC AC mode	
		DC	DC mode
		ACDC	AC+DC mode
		DCAC	DC+AC mode
	Output impedance	Set the output imp	
		Status Enable or disable this function	
		R	Set the output resistance
		L	Set the output inductance
	Off mode	Set the output turr	
	On mode	Open-Z	Open circuit mode
		High-Z	High impedance mode
		Short	Short circuit mode
	Loop speed		
	Loop speed	Output loop speed control High High speed	
		Low	Low speed
	External Lock-	LOW	Low speed
Current Source	frequency control	Set the external Lock-frequency control mode	
		Status	Set the Lock-frequency ON/OFF state Off/On Lock-Freq Lock-Phase Fiber-Phase-Master Fiber-Phase-Slave
		Phase Delay	Set the phase deviation between the output phase and the external I/O input signal: 0-360°.
		Freq limit+	Set the frequency difference upper limit between the output frequency and the external I/O input signal.
		Freq limit-	Set the frequency difference lower limit between the output frequency and the external I/O input signal.
		Exception	Set the output mode when the frequency lock fails: Output Disable: stop output Limit: Output according to the set frequency.
		Sync select	Synchronization mode selection (settable only in fiber slave mode). Local: The slave parameters take effect in real time.



External programme External analog function: Status Set the ON/OFF state AM: Adjust the amplitude Amplifier: Real-time output and power Amplifier. Select the phase to monitor. This cannot be set in single phase mode. U ratio The external program ratio of voltage. I ratio The external program ration of current. Remote sense Set the sense function state. On: Enable the remote sense Measurement Off: disable the sense function On/Off phase On-mode On-mode On-mode On-mode On-mode Measurement Off-mode Measurement Off-mode Measurement Set the measurement speed Lowest (1Hz) Lowest mode, measuring every 2s
Status Set the ON/OFF state
AM: Adjust the amplitude Amplifier: Real-time output and power Amplifier. Select the phase to monitor. This cannot be set in single phase mode. U ratio The external program ratio of voltage. I ratio The external program ration of current. Remote sense Set the sense function state. On: Enable the remote sense Measurement Off: disable the sense function On/Off phase On-mode On-mode On-mode Off-mode Off-mode Off-mode Measurement Off-mode Measurement Off-mode Measurement Off-mode Measurement Off-mode Output of phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Output off phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Output off phase Off-mode
Mode Amplifier: Real-time output and power Amplifier. Select the phase to monitor. This cannot be set in single phase mode. U ratio The external program ratio of voltage. I ratio The external program ration of current. Remote sense Set the sense function state. On: Enable the remote sense Measurement Off: disable the sense function On/Off phase On-mode Imm: immediately Onf-mode On-mode Ontput off phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Off-mode Imm: immediately Imm: immediately Slope: Controlled according to the set slope Off-mode Set the measurement speed
Monitor phase U ratio The external program ratio of voltage.
Pratio Voltage. I ratio The external program ration of current. Remote sense Set the sense function state. On: Enable the remote sense Measurement Off: disable the sense function Output on phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Off-mode Imm: immediately Off-mode Imm: immediately Off-mode Imm: immediately Slope: Controlled according to the set slope Measurement Set the measurement speed
Remote sense Set the sense function state. On: Enable the remote sense Measurement Off: disable the sense function Output on phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Off-mode Off-mode Measurement Off-mode Measurement Set the measurement speed
On: Enable the remote sense Measurement Off: disable the sense function Output on phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Off-mode Off-mode Off-mode Measurement Off: disable the remote sense Measurement Off: disable the sense function Output on phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Measurement Set the measurement speed
Mode Measurement Off: disable the sense function Output on phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Off-mode Off-mode Measurement Off: disable the sense function Output on phase control Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Measurement Set the measurement speed
On/Off phase On-mode Output off phase control On-mode On-mode Output off phase control Output off phase
Off-mode Off-mode Off-mode Phase: setting the phase Imm: immediately Slope: Controlled according to the set slope Measurement Set the measurement speed
Lowest (1Hz) Lowest mode, measuring every 2s
Lower (1000ms) Lower mode, measuring every 1000ms
Slow (500ms) Slow mode, measuring every 500ms
Medium (300ms) Medium mode, measuring every 300ms
Fast (150ms) Fast mode, measuring every 150ms
Filter Whether to turn on the filter during measurement.
Power Unit Setting Unit of power, can be set to KW/kVA/KVar or W/VA/Var.

General Menu:

	Buzzer	Set the keyboard sound	
General		Key	Buzzer on/off for key
		Protect	Buzzer on/off for protect
	Brightness	Set the screen brightness.	
		1-10	Set the screen brightness level
	Factory-default- settings	Select whether to reset the factory default settings or not.	
		Enter	Confirm to reset operation



TILON I			System-Related Functions	
	Power-on setup	Set the power-on s	state.	
			When the instrument is powered	
		Reset	on, the instrument will initialize	
			some settings and [On/Off] state.	
			When the instrument is powered	
			on, the instrument will remain	
		Last	the same settings and [On/Off]	
			state as last time	
			you turned off the instrument.	
			When the instrument is powered	
			on, the instrument will remain	
		Last-OFF	the same settings as last time you	
		Lasi-OFF	turned off the instrument,	
			•	
	Dorollol mode	Cat the instrument	but the [On/Off] is OFF state.	
	Parallel mode	Set the instrument	s to parallel operation mode.	
			Master: Set the instrument to	
			master mode.	
		Б.	Single: Set the instrument to	
		Role	single mode, i.e., disable the	
			parallel operation mode.	
			Slave: Set the instrument to	
			slave mode.	
		Numbers	Number of parallel instruments.	
	Touch function	Lock the touch scr		
		Status	Set the ON/OFF state	
			ke effect immediately. If set to ON,	
	Knob immediately			
	effective	OFF, press ENTER to confirm the effect after the		
		setting is completed.		
	Language	Set the language of	of display	
		English	English	
		Chinese	Chinese	
	Soft keyboard	Set the soft keyboa		
			Turn on the soft keyboard.	
		Off	Turn off the soft keyboard.	
	LISB type		Tam on the soit keyboard.	
	USB type	Set the USB type.		
		DEVICE: the USB device is used to communication		
		with PC.		
	LIOD 1 : 1		evice is used to storage disk.	
	USB device class	USB communication		
		VCP	Virtual serial port	
		TMC usbtmc	USB-TMC protocol	
	LAN config	LAN communication		
			DHCP: automatically configure	
Communication			the address of the instrument.	
		Mode	 Manual: manually configure 	
		MIDGE	the address of the instrument	
			by entering values in the	
			following five fields.	
		IP	Set the IP address.	
		Mask	Set the subnet mask.	
		Gateway	Set the gateway address.	
		Port	Set the port number.	
	CAN config	CAN communication		
	o, ar coming	C/ (14 CCITITICATIOATIC	JII III.OIIUOO.	



ITEUN			System-Related Functions	
		Baud rate	Select the baud rate	
		Addr	Set the instrument address to a	
			number	
		Select RS-232 con	nmunication interface. When insert	
	RS232 config	IT-E177 communication board into expansion sl		
	1	the menu displays this information.		
		Baud rate Baud rate		
		Databits	Data bit: 5/6/7/8	
		Stopbits	Stop bit: 1/2	
		Even-odd check	Parity bit: N (No parity) / E (Even	
			parity) / O	
		A 1 1	(Odd parity)	
		Addr	Address	
		Select GPIB communication interface. When ins		
	GPIB config		on board into expansion slot, the	
		menu displays this		
		Addr	Set the communication address	
	Digital IO-1:Remote	Function setting of	nin 1	
	Inhibit Input	i unclion selling of	piii i	
			On/Off, Select Invert or not under the	
		Reverse	IO Settings. If setting to ON, it means	
			the valid signal is reversed.	
			Inhibit-living:	
			Inhibit-latch	
		Function	Input	
			Output	
	Digital IO-2: PS		Ουτραί	
	Clear	Function setting of pin 2		
		Reverse	On/Off, Select Invert or not under the	
			IO Settings. If setting to ON, it means	
			the valid signal is reversed.	
			PS Clear	
			• Input	
		T dilotion	Output	
	Digital IO-3: PS	Function setting of		
10	Digital 10-3. F3	i unouon seung or	On/Off, Select Invert or not under the	
IO		Dovorce		
		Reverse	IO Settings. If setting to ON, it means	
			the valid signal is reversed.	
			• PS	
		Function	• Input	
			Output	
	Digital IO-4: SYNC	Function setting of		
			On/Off, Select Invert or not under the	
		Reverse	IO Settings. If setting to ON, it means	
			the valid signal is reversed.	
			Sync-in	
			Sync-out	
			• Input	
			Output	
	Digital IO-5:			
	ON/OFF Status	Function setting of pin 5		
			On/Off, Select Invert or not under the	
		Reverse	IO Settings. If setting to ON, it means	
			the valid signal is reversed.	
	1		and tand signal is reversed.	



IIEGII			System-Related Functions
		Function	ON/OFF StatusInputOutput
	Digital IO-6: TRIG1	Function setting of	
	2.g.u. re o. rrue r	Reverse	On/Off, Select Invert or not under the IO Settings. If setting to ON, it means the valid signal is reversed.
		Function	Trigger1-outTrigger1-inInputOutput
		AC	On/Off: When On is selected, a trigger signal is output when AC amplitude changes. The accuracy of voltage change is 100mV and is not restricted by phase. (This configuration is displayed only when the IO pin is set to Trigger1-out)
		DC	On/Off: When On is selected, a trigger signal is output when the DC amplitude changes, and the voltage change precision is 100mV. (This configuration is displayed only when the IO pin is set to Trigger1-out)
		Freq	On/Off: When On is selected, a trigger signal is output when the frequency changes. The accuracy of frequency change is 0.1 Hz (This configuration is displayed only when the IO pin is set to Trigger 1-out)
		List	On/Off: When On is selected, the List generates a trigger signal and outputs a trigger signal. (This configuration is displayed only when the IO pin is set to Trigger1-out)
	Digital IO-7: TRIG2	Function setting of	pin 7
		Reverse	On/Off
		Fun	Trigger1-outTrigger1-inInputOutput
		AC	On/Off: When On is selected, a trigger signal is output when AC amplitude changes. The accuracy of voltage change is 100mV and is not restricted by phase. (This configuration is displayed only when the IO pin is set to Trigger2-out)
		DC	On/Off: When On is selected, a trigger signal is output when the DC amplitude changes, and the voltage change precision is 100mV. (This configuration is displayed only when the IO pin is set to Trigger2-out)
		Freq	On/Off: When On is selected, a trigger signal is output when the frequency



			·	
			changes.The accuracy of frequency change is 0.1Hz	
			(This configuration is displayed only	
			when the IO pin is set to Trigger2-out)	
			On/Off: When On is selected, the List	
			generates a trigger signal and outputs a	
		List	trigger signal.	
			(This configuration is displayed only	
			when the IO pin is set to Trigger2-out)	
	Product model	Display the instrument model.		
Information	Serial number	Display the serial number.		
	Software version	Display the control board version.		
	MAC address	MAC address		
	Rbf Version	Rbf version		
	Ctrl1 version	Ctrl1 version		
	Ctrl2 version	Ctrl2 version		
	Hardware version	Hardware version		
	Inner numbers	The inner numbers setting item is required when using		
		F-TX and F-RX fiber to achieve master-slave model		
		parallel.		

6.1.1 Menu function

Set OFF Mode Status

This parameter is used to set the status after the power output is off.

- Select to High-Z, After the power output is Off, the dc impedance between the terminals of the power supply is high, and the resistance value varies with different models.
- Select to Short, When the power output is Off, the terminals of the power supply are short-circuited, the voltage is 0.
- Select to Open-Z, After the power output is Off, the power supply is in open mode, and the internal circuit and external load of the power supply are disconnected by relay.

Set the keyboard sound

This item can set the key sound state. If in ON mode, then when you press a button, the power supply will beep. If in OFF mode, the beeper will not make a sound. The default set is in ON mode.

Set the screen brightness

This item can set the screen brightness. Set the screen brightness within the range 1 to 10 by pressing number keys on the front panel. The larger the number is, the higher the screen brightness is. You can also set the screen brightness by rotating the knob on the front panel.

Restored to Factory Setting

This menu item is used to restore some parameter settings to factory setting values.

The procedures to set the menu item are as follows.

Select the General under system menu.



2. Press **Enter** in **Factory_default_settings**. After the parameter settings are complete and return to main interface.

Set the Power-on State

This parameter determines the state of the AC source after power up.

The procedures to set the menu item are as follows.

- 1. Press the **General** under the system menu.
- 2. Press the Up/Down key or turn the knob to select the **Power-on setup** and press **[Enter]**.
- Reset: Default value, indicates when the instrument is powered on, the instrument will initialize some parameter settings or state, such as output voltage, output frequency, start phase and stop phase.
- Last: Indicates when powered on, the instrument will remain the same parameter settings and output status as last time you powered off the instrument.
- Last+Off: Indicates when powered on, the instrument will remain the same settings as last time you powered off the instrument, but the output status is Off.

Lock the Touch Screen

This parameter determines the state of the touch screen.

- 1. Press the **General** under the system menu.
- 2. Press the Up/Down key or turn the knob to select the **Touch screen lock** and press **[Enter]**.
- On: enable the touch screen
- Off: disable the touch screen

Set the Loop Speed

This item can control stability of the loop. When the connected load is capacitive load or inductive load, select Slow; when the connected load is resistance, select High.

Set the Knob Function

Set the knob setting function. If set to ON, the Knob setting will take effect immediately. If set to OFF, press Enter to confirm the effect after the Knob setting is completed.

Select Language

Users can select the instrument language type from the menu.

Set the Soft Keyboard

The user can open the soft keyboard in the menu. When the parameter is set to ON, the soft keyboard is enabled. And when setting parameters on the screen, the soft keyboard appears. Convenient users directly touch screen to select the number.

6.1.2 Set the communication interface

This menu item is used to set the communication information between



instrument and PC. The standard communication interfaces for IT7900 series power supply are USB, LAN and CAN. You can also select the non-standard interface GPIB or RS-232 based on personal requirement.

The user does not need to select the communication interface. The instrument will automatically select the present communication mode according to the communication interface accessed by the rear panel. The user only needs to set the communication parameters and keep them consistent with the PC Settings.

☐ NOTE

- When select USB interface, the USB type need to be DEVICE.
- The instrument supports the non-standard RS-232 and GPIB interfaces, and the configuration items in the menu are dynamically displayed according to the interface user selected

6.1.3 View the System Information

System Information menu is used to view the system information of the instrument. Include product model, SN, software version and MAC address.

6.2 Configuration Menu Reference

Press [Config] key and enter to configuration menu interface. At this interface, user can setup the power supply output parameters, detailed parameters are shown as follows:

DC			
Config	Configuration setting for DC mode		
_	Voltage DC	Vdc: Output voltage of DC, range from 0- full scale Slew Rate: Slope , range from 0-5000V/ms	
DC+AC Config	Configuration settir	ng for DC+AC mode	
	Voltage DC	Output voltage of DC, range from 0-full scale Slew Rate: Slope , range from 0-5000V/ms	
	Ripple control	Wave: Select the output wave type Vac: output voltage of AC, range from 0- 10% of full scale Freq: frequency value	
AC Config	Configuration setting for AC mode		
	Balance control	Only displayed in 3-phase mode. Balance control in 3-phase mode. If select On, the output of ABC phase is synchronous. If select Off, the output is not balanced.	
	Voltage AC	Output voltage Vac: output voltage of AC, range from 0-full scale Slew Rate: Slope , range from 0.0001- 5000V/ms	
	Frequency	Freq: output frequency, range from 16- 150Hz Slew Rate: slope, range from 0.00001- 5000Hz/ms	



		•
	Phase control	Phase difference between AB and AC, only valid in balance control is Off state. AB: Phase difference between A and B. AC: Phase difference between A and C.
	Waveform	Selece and edit the output wave type. For detailed introduction please refer to 5.6 waveform selection.
	Dimming	Status: Dimming function switch. Edge: select front edge or back adge Phase: phase control, range from 0-180°
AC+DC Config	Configuration settin	ng for AC+DC mode
	Balance control	Only displayed in 3-phase mode. Balance control in 3-phase mode. If select On, the output of ABC phase is synchronous. If select Off, the output is not balanced.
	Voltage AC	Output voltage Vac: output voltage of AC, range from 0-full scale Slew Rate: Slope , range from 0.0001- 5000V/ms
	Voltage DC	Output voltage of DC, range from 0-10% of full scale Slew Rate: Slope , range from 0-5000V/ms
	Frequency	Freq: output frequency, range from 16- 150Hz Slew Rate: slope, range from 0.00001- 5000Hz/ms
	Waveform	Selece and edit the output wave type. For detailed introduction please refer to 5.6 waveform selection.
	Dimming	Status: Dimming function switch. Edge: select front edge or back adge Phase: phase control, range from 0-180°

Current Source Mode:

AC Config	Configuration setting for output.	
	Current AC	Output current. lac: output current of AC, range from 0-full scale Slew Rate: Slope, range from 0.0001- 5000A/ms
	Frequency	Freq: frequency of output, rang from 16- 2400Hz Slew Rate: slew rate of frequency, 0.00001-5000Hz/ms
	Waveform	Selece and edit the output wave type. For detailed introduction please refer to 4.3 waveform selection.
	Dimming	Status: Dimming function switch. Edge: select front edge or back adge Phase: phase control, range from 0-180°



AC+DC Config	Configuration setting for output.	
	Current AC	Output current. lac: output current of AC, range from 0-full scale Slew Rate: Slope , range from 0.0001- 5000A/ms
	Current DC	Idc: DC current value, range from 0-full scale Slew Rate: slew rate, 0-5000A/ms It is not displayed in three-phase balanced mode, if you need to set the three-phase DC component, balanced mode needs to be turned off.
	Frequency	Freq: frequency of output, rang from 16- 2400Hz Slew Rate: slew rate of frequency, 0.00001-5000Hz/ms
	Waveform	Selece and edit the output wave type. For detailed introduction please refer to 4.3 waveform selection.
	Dimming	Status: Dimming function switch. Edge: select front edge or back adge Phase: phase control, range from 0-180°

Balance Control Setting

In three-phase mode, the instrument is in three-phase balance mode by default, and the three-phase output voltage will be set synchronously, and the phase Angle between phases is fixed at 120°, which cannot be set.


You can set the Balance Control to Off and disable the Balance control function in the Config menu. And then, the voltage of each phase can be set independently, and the config menu displays the phase Angle difference setting parameter between AB and AC.

Dimming Function


The front and back edge of the waveform can by concealed and the phase angle set with Dimmer function to regulate the active power, thus adjusting the lighting intensity.

Select Front edge, and phase set to 90° in Configuration menu.





Select Back edge, and phase set to 90° in Configuration menu.



6.3 Key Lock Function

Press [Shift] + [2] (Lock) button to set the key lock state. If keyboard has been locked, the indicator light "Lock" will display on the LCD. In addition, when keyboard are locked, all buttons can't be used except Local key Press [Shift] + [2] (Lock) once again will relieve key lock function.

6.4 Switching Local/Remote Mode

You can press the **[Shift] +[3]** (Local) button to change the AC source from remote to local operation.

After you power on the AC source, it defaults in local mode, all buttons are enabled. While in remote mode, most buttons are disabled except **[Shift] +[3]** (Local) keys. You can switch Local/Remote mode via PC. In addition, the mode modification will not affect the output parameters.



6.5 Save and Recall Operations

The power system can save up to 10 common parameters in nonvolatile memory (No. 1 to No. 10) for user to recall conveniently.

The saved parameters include:

- Power supply mode
- Present output mode
- Config menu settings

You can do the save and recall operations by the following two methods.

- Press the composite keys [Shift]+[4] (Save) to save the parameters. Pressthe composite keys [Shift]+[5] (Recall) to recall the parameters.
- SCPI commands: *SAV and *RCL

Save Operation

The save operation procedures are as follows:

- 1. Press the composite keys [Shift]+[4] (Save) to enter the parameter save interface.
- Select the storage location. up to 10 position can be select.



3. Press [Enter] to save the parameters.

Finished, the saved parameters will be display at the bottom of the interface.

Recall Operation

You can recall the parameters you saved in the specified memory location as the setting values.

- 1. Press the composite keys [Shift]+[5] (Recall) to enter the parameter recallinterface.
- Set the storage location.

Press the direction keys to set the storage location, and then, the saved parameters will be display at the bottom of the interface.

Press [Enter] to recall the parameters.



6.6 Protection Function

IT7900 series source includes the following protection functions: overcurrent protection (Current RMS protection, Current peak protection), voltage limit protection and over-temperature protection (OTP).

AC/ACDC Mode

Press [Shift]+[Config] (Protect) and enter to Protect configure menu, where you can set the following protection.

Current RMS protection	RMS current for OCP protection	
·	Rms	Protection point
	Time	delay time of protection
	Туре	Protection type.
		Limit
		Output Disable
Current peak protection	Current Pea	ak OCP
	Peak	Protection point
	Time	delay time of protection
Voltage peak range	Voltage peak range setting	
	V limit	Maximum voltage setting
Power limit range	Power limit	setting
	P limit	Power limit point
	Time	delay time of protection
	Туре	Protection type.
		Limit
		Output Disable

Current Source mode:

Current Limit range	Current limit range	
	Mode	Can not be set.
	Time	Delay time, range from 0s-10s.
	I-High	Current high value
	I-Low	Current low value
Voltage Limit range	Voltage limit	t range
	Mode	Can not be set
	Time	Delay time, range from 0s-10s.
	V-High	Voltage high value
	V-Low	Voltage low value
Power Limit range	Power limit	range
	Mode	Can not be set
	Time	Delay time, range from 0s-10s.
	P-High	Power high value
	P-Low	Power low value

DC/DCAC Mode

Press [Shift]+[Config] (Protect) and enter to Protect configure menu, where you can set the following protection.

Current limit range	Current limit setting	
	I+ limit	Maximum current setting, the current setting will be limited to
		this range.



	I- limit	Minimum current setting, the current setting will be limited to this range.
Voltage limit range	Voltage limit setting	
	V+ limit	Maximum voltage setting, the voltage setting will be limited to this range.
	V- limit	Minimum voltage setting, the voltage setting will be limited to this range.
Power limit range	Power limit setting	
	P+ limit	Maximum power setting, the power setting will be limited to this range.
	P- limit	Minimum power setting, the power setting will be limited to this range.

6.6.1 Current RMS Protection

The user can set the over-current protection point, delay time and protection type for the Current RMS protection function. The function is mainly used to protect the DUT connected during test to prevent it from damage due to over-current. The OCP (rms) function is always enabled.

RMS OCP Type:

- Limit: If the output current reaches the RMS current protection point, the protection will be activated. The power supply output current is controlled within the current limiting value. The power supply will output in current limit mode.
- Output Disable: When the output current reaches the over-current protection point and the period is greater than the set protection delay time, the over-current protection (rms) will turn off the output.

How to Set

- 1. Press [Shift]+[Config] (Protect) keys and enter to Protection menu.
- 2. Press the up/down key or rotate the knob to select Current RMS protection and press [Enter].
- 3. Set the protection RMS Level, the delay time and protection type in sequence, and press [Enter] to confirm.

Clear RMS OCP Protection

When RMS OCP protection occurs, the instrument responds as follows:

- Instrument output is off:
- The buzzer sounds;
- The interface displays "OCPrms" and flashes

To clear the RMS OCP and return to normal operation, firstly remove the conditions that caused the protection fault. Press [Shift] +[Esc] key (or send the command PROTection:CLEar) to clear the protection status. The message displayed in front panel is cleared and the instrument exits protection status.

6.6.2 Set the Current Peak protection

The user can set the current peak protection point and delay time for the Current peak protection function. The function is mainly used to protect the DUT



connected during test to prevent it from damage due to over load. The OCP peak function is always enabled.

How to Set

- 1. Press [Shift]+[Config] (Protect) keys and enter to Protection menu.
- 2. Press the up/down key or rotate the knob to select Current peak protection and press [Enter].
- 3. Set the protection peak Level and the delay time in sequence, and press [Enter] to confirm.

Clear Peak OCP Protection

When peak OCP protection occurs, the instrument responds as follows:

- Instrument output is off;
- The buzzer sounds;
- The interface displays "OCPpeak" and flashes

To clear the peak OCP and return to normal operation, firstly remove the conditions that caused the protection fault. Press [Shift]+[Esc] key (or send the command PROTection:CLEar) to clear the protection status. The message displayed in front panel is cleared and the instrument exits protection status.

6.6.3 Set Voltage/Current/Power Limit Range

The maximum voltage, current and power of the power supply ranges from minimum to full-rated output range. You can setting maximum and minimum values of power supply voltage/current/power in the protection menu. When limit setting is finished, the voltage setting value, current setting value and power setting value can only be set within the maximum and minimum limits.

Limit factory setting is the rated output voltage/current/power of corresponding model of the power supply. Minimum is 0.

Take the voltage limit setting for an example, the operating as follows:

- 1. Press [Shift]+[Config] (Protect) keys and enter to Protection menu.
- 2. Press the up/down key or rotate the knob to select Voltage limit range and press [**Enter**].
- 3. Set the V+ limit and the V- limit in sequence, and press [Enter] to confirm.

6.6.4 Over-temperature protection (OTP)

When the temperature of the power component in the power supply exceeds 95°C, the temperature protection will be enabled. In this case, the power supply

will be automatically OFF, and the LCD will display . At the same time, the OT position in the status register will be set and kept until power supply is reset.

Clearing over-temperature protection:

When the power supply temperature decreases to the protection temperature, press [Shift]+[Esc] key on the front panel (or send the command

"PROTection:CLEar"). Then on the power supply screen will disappear, and the power supply will exit the OTP status.



6.6.5 Over-power protection (OPP)

When the output power exceeds the set power, OPP will be enabled, and will appear on the LCD screen.



Clearing over-power protection:

In the case of over-power protection, disconnect the tested object at first. Press [Shift]+[Esc] on the front panel (or send the command "PROTection:CLEar") to clear on the front panel and exit the OPP mode.

6.7 Screen Capture Function

IT7900 series power supply has the screen capture function. Insert the USB equipment into the USB interface of the front panel, and press [Print] on the front panel to capture and save the current screen into the USB disk.

When you need the screen capture function, the USB type under the system menu needs to be set to Host.

6.8 Trigger Function

IT7900 series has four trigger source to choose: trigger by keys (Key), Bus trigger (Bus) and External signal trigger (TRIG1/TRIG2).

- Manual: if [Trig] on the front panel is pressed in the valid manual trigger mode, the power supply will be triggered once.
- Bus: if the bus trigger command is received by the power supply in the valid bus trigger mode, the power supply will be triggered once.
- Trigger1: if one trigger signal is connected to the Digital IO-6 of digital I/O Interface terminal on the rear panel, the power supply will be triggered once.
- Trigger2: if one trigger signal is connected to the Digital IO-7 of digital I/O Interface terminal on the rear panel, the power supply will be triggered once.

6.9 Query the System Log

The IT7900 series power supply provides the system operation Log query function. On the Menu interface of the front panel of the instrument, click Log or directly press Shift +1[Loq] to enter the Loq query interface. You can view historical system operation records on this screen.

6.10 Query the Energy

IT7900 series power supply provides the energy statistics function. Click WHours button in the Menu interface of the front panel of the instrument to enter the quantity query interface. You can view power statistics on this screen.

6.11 Set Parallel Operation Mode

The IT7900 series power supply supports multiple instruments to work in parallel mode to provide more power and current output capability. Under the parallel mode, All features are set up from the master unit.

This chapter takes three instruments (with operation panel) as an example to describe how to parallelize the single units and how to return from parallel mode to single mode.



Connect the instruments (3U model)

CAUTION

- Before connecting the system bus, you must ensure that each instrument is in single mode (Single).
- Fiber optic cables cannot be flexed or folded. When the cable is too long and needs to be arranged, gently wrap the cable in a circle and gently tie it.


WARNING

- Before connecting the cables, ensure that the instrument power switch is off and the main switch of the AC power input (distribution box) is off.
- Before connecting 3 single instruments to the AC distribution box, ensure that the distribution box capacity is sufficient. Refer to the corresponding specifications for the AC input parameters of a single instrument.


For 3U models, the same model can be connected in parallel. Taking three 3U single instruments (with operation panel) as an example, the steps for parallel operation are as follows.


- 1. Ensure that the power switches of the three units and the main switch of the AC power distribution box are off.
- 2. Refer wiring connection diagram to connect three units.

For long distance testing requirements:











- 3. Connect the AC input terminals of the three units separately, and connect them to the AC distribution box.
- 4. Connect the output terminals of the three units in parallel and connect them to the DUT.
- 5. Refer to the blue wiring legend in the figure, connect the System Bus (i.e., the fiber outer ring interfaces TX and RX) for fiber-optic communication between the master and slaves.
- a) Insert the fiber optic module into the hole corresponding to TX RX.



b) Insert the plug of the fiber optic cable into the fiber optic module and hear a click sound to indicate that it is inserted in place. The fiber optic cable connection schematic is as follows.





Set the Parallel Mode

- 1. Turn on the main switch of the AC distribution box and power on each of the three units.
- 2. Set three units in parallel mode with one master and two slaves.
- 3. Press the composite keys [Shift]+ (System) on the front panel to enter the system menu.
- 4. Select General menu.
- 5. Set the **Parallel Mode**, set them to one master unit and two slave units. In each group, one instrument must be the master unit and all other instruments connected in parallel are slave units. All features are set up from the master unit.
- Single: Default value, indicates that the instrument is in single mode.
- Master: Indicates that the single unit is set to master in parallel mode. Numbers: total number of units in the parallel relationship, when the instrument set to master, you need to set the Numbers. For example, Numbers set to 3.
- Slave: Indicates that the instrument is set to the slave in parallel mode.
- 6. After the parallel menu of the three units are set, restart the instrument separately.

After the instrument is restarted, the screen shows that the instrument is working in parallel mode.

Revert to Single Mode

1. Set each of the three instruments to single mode.

a) Press the composite keys [Shift]+



(System) on the front panel to



enter the system menu.

- b) Select General menu.
- c) Set the Parallel Mode, set them to single.
- 2. Power off the three instruments and turn off the main switch of the AC distribution box.
- 3. Remove the cables connection of the System Bus and output terminals between three units.
- 4. Power on the three instruments separately.

After the instrument is restarted, the screen shows that the instrument is working in single mode.

Connect the instruments (Cabinet model)

CAUTION

- Before connecting the system bus, you must ensure that each instrument is in single mode (Single).
- Fiber optic cables cannot be flexed or folded. When the cable is too long and needs to be arranged, gently wrap the cable in a circle and gently tie it.

WARNING

- Before connecting the cables, ensure that the instrument power switch is off and the main switch of the AC power input (distribution box) is off.
- Before connecting 3 single instruments to the AC distribution box, ensure that the distribution box capacity is sufficient. Refer to the corresponding specifications for the AC input parameters of a single instrument.


For cabinet models, the same model can be connected in parallel. Taking two 15U instruments (with operation panel) as an example, the steps for parallel operation are as follows.

1. Ensure that the power switches of the instrument and the main switch of the AC power distribution box are off.

Refer wiring connection diagram to connect two units.

The configuration method is the same as 3U model instrument.





Notice:


- The connection cables from the cabinet terminals to the mid-node are as short as possible and bundled up after the connection is completed. The cable length from the two cabinets must be the same.
- Connect TX, RX fiber cable of two cabinets for communication, connected in the same way as a 3U model instrument.
- Each cabinet has a red role selection button that allows the user to switch the red button to select the present instrument as a master or as a slave.



- When paralleling as a master role, several cabinets in parallel need to be powered on or powered down separately.
- When parallel as a slave role, the present slave cabinet up and



down power to follow the host operation, at this time the slave cabinet and the master cabinet need to connect the CTRL interface. The wiring method is shown in the figure below.



6.12 Remote Measurement Function

The IT7900 series power supply supports two connection methods: Local measurement and Remote sensing. The remote sensing is used for maximizing measurement accuracy. (Refer to 2.4 Connecting Test Lines).

The procedures to set the menu item are as follows.


- 1. Press the composite keys [Shift] + (System) on the front panel to enter the system menu.
- 2. Press the up/down key or rotate the knob to select **Sense compensation** and press [**Enter**].
- Local: Default value, indicates turn the sense function off.
- Remote: Indicates turn the sense function on.
- 3. After the parameter settings are complete, press [Enter].

6.13 Digital I/O Function

This series power supply supports digital I/O function. The user can realize logic control over high and low level input or output by related configurations in the system menu.

Pins Introduction

Different I/O implements different functions. The detailed functions description are shown in the figure below:





Pin	Туре	Description	Properties
Digital IO-1	Input/Output	Turn off the output under emergency status	Level signal
Digital IO-2	Input/Output	Clear the protection state	Pulse signal
Digital IO-3	Input/Output	Protection state indicator	Level signal
Digital IO-4	Input/Output	synchronous control	Pulse signal
Digital IO-5	Input/Output	OnOff-status indicator	Level signal
Digital IO-6	Input/Output	Trigger1	Pulse signal
Digital IO-7	Input/Output	Trigger2	Pulse signal
<u></u>	Ground terminal, that is, the negative terminal corresponding to each of the above 7 pins.		

General Digital I/O Function

Signal definition

Digital I/O functions involve input and output levels and pulse signals. The input signal is the control signal provided externally to IT7900, the output signal is the level signal provided externally by IT7900, and the pulse signal is the edge signal switched between high and low levels.

		Typical: 5V
	High level signal	Range: 1.6V-15V
Input signal		Current: ≤100mA
pat oigna.	Low level signal	Typical: 0V
		Range: -5V-0.8V
		Current: ≤100mA
	High level signal	Voltage level: 5V
Output signal		Current: ≤1mA
Output signal	Low level signal	Voltage level: 0V
		Current: 0.5mA
	Level rise slope	10us
Pulse	Level fall slope	2us
	Width	30us

Input/Output Function

The IO-1 \sim IO-7 pins are featured default function, the user can setting the function of pin according to requirement. The Input and Output are the general digital I/O function, and the parameter settings and functions of the



seven pins are the same.

The IO-1~IO-7 pins provide default functions. Users can realize control according to the functions defined. Users can also reset the input or output properties of the present pin and customize the function use of the pin according to their needs.

When pins 1 to 7 are configured to Output function, when send the command (IO:STATe 1/0) to instrument, the IO pin can output high level (False) or low level (True).

When pins 1 to 7 are configured to Input function, an external signal can be Input to this pin, and the instrument can detect the state of the external signal.

Signal Revert

Select Invert or not under the IO Settings menu. If setting to OFF, it means the default level will be valid. If setting to ON, it means the valid signal is reversed. For example, the IO-1 pin is inhibit output by default and the high level is valid, when select revert ON, the low level is valid and the instrument output is disabled.

Digital IO-1

IO-1 pin can be set to 【Inhibit-latch】,【Inhibit-living】, 【Input】, 【Output】

The default function is inhibit output. When the IO pin is configured for a Inhibit function and the level signal is low, the output of the machine is forbidden. At this point, Pin 1 has a bi-directional I/O function, which can both receive the level signal input from the external instrument and output the level signal outward.default level is high, and low is valid when entering. Outgoing output also generates low level signals.

Inhibit function has two mode: Latch and Living.

- Living: When input an inhibit signal and the instrument output is turned OFF. The status bar of the LCD screen displays INH warning icon and the output is marked as OFF. If power supply output is ON state before, the ON/OFF button will be lit. When the input signal undoes, the output returns to normal. This function can be used to control the output of the power supply.
- Latch: When input an inhibit signal and the instrument output is turned OFF.
 The ON/OFF button will be lighted off, the status bar of the LCD screen
 displays INH warning icon. In this case, user need to remove the input signal
 and press [Shift]+[Esc] to cleare protection, then manually turn on [On/Off]
 again.

Digital IO-2

IO-1 pin can be set to 【PS-clear】, 【Input】, 【Output】

The default function is to clear the protected state. When the protection occurs, the protection state can be cleared through this pin, so that the instrument can continue to output normally.

IO-2 is bi-directional, that is, when the power supply is in a protected state, the instrument can receive a pulse signal from an external input through IO-2 for clean protection operation, or when the power supply is in a protected state, the clean protection can generate a pulse signal from IO-2.

Digital IO-3

IO-3 pin can be set to 【PS】, 【Input】, 【Output】



The default function is protection state indicator. IO-3 pin will output high or low level based on whether the instrument is under protection or not. Under normal conditions (Not under protection), and when pin3 is under default setting (Not Invert), pin 3 outputs high level; when the instrument is under protection, pin 3 outputs low level. When pin3 is set to Invert, the output level is completely opposite.

Digital IO-4

IO-4 pin can be set to [Sync-in], [Sync-out], [Input], [Output]

This function can be used to simulate the six-phase output mode, in which one IT7900 outputs a synchronous signal to another IT7900, and the second power supply runs synchronously with the frequency and phase of the first power supply to realize the six-phase output function.

[Sync-in]: Synchronous input function, which is used to output frequency lock or phase lock with the external signal. At this time, the machine synchronizes the frequency or phase information input from the IO-4 pin.

[Sync-out]: Synchronous output function, the IT7900 produces synchronous signal to the outside, which is AC zero crossing pulse signal sent from the IO port.

Digital IO-5

IO-5 can be set to 【OnOff-status】, 【Input】, 【Output】

The default function is to indicate the output state of the power supply, in case of output is ON, output 5V, otherwise, output 0V.

When pin5 is set to Invert, the output level is completely opposite.

Digital IO-6

IO-6 can be set to 【Trigger1-in】, 【Trigger1-out】, 【Input】, 【Output】

【Trigger1-in】: The input trigger signal, the pulse signal sent to the IO-6 pin can be used as the trigger source. Users can select as the trigger source of the corresponding function in the menu.

【Trigger1-out】: The output trigger signal, when the instrument generates a trigger signal, the pin6 generates an pulse signal.

Digital IO-7

IO-7 can be set to [Trigger2-in], [Trigger2-out], [Input], [Output]

【Trigger2-in】: The input trigger signal, the pulse signal sent to the IO-6 pin can be used as the trigger source. Users can select as the trigger source of the corresponding function in the menu.

【Trigger2-out】: The output trigger signal, when the instrument generates a trigger signal, the pin7 generates an pulse signal.

6.14 Analogue Function (Ext-Program) (Optional)


The interface expansion slot provided on the rear panel of the IT7900 series. This function is not standard with the instrument and is optional for users.

When the interface card selected by the user is RS232+Analog interface (IT-E177), the analog interface can realize the external analog function.



- Remotely control voltage
- Power amplifier function
- Remotely monitor voltage and current values

The pins description is as below.



Pins	Name	Description
4 pin	GND	Grounding of analog signals,
5 pin	V_Monitor	Monitor voltage. When output is DC, -10V ~ 10V voltage value to monitor the output or absorption voltage between negative full range and positive full scale. When output is AC, 0-10V voltage value to monitor the 0 to full scale.
6 pin	I_Monitor	Monitot Current. When output is DC, -10V ~ 10V voltage value to monitor the output or absorption voltage between negative full range and positive full scale. When output is AC, 0-10V voltage value to monitor the 0 to full scale.
8 pin	A input	Set the voltage of A phase. When output is DC, -10V ~ 10V voltage value to set the voltage between negative full range and positive full scale. When output is AC, 0-10V voltage value to set the 0 to full scale.
9 pin	B input	Set the voltage of B phase. When output is DC, -10V ~ 10V voltage value to set the voltage between negative full range and positive full scale. When output is AC, 0-10V voltage value to set the 0 to full scale.
10 pin	C input	Set the voltage of C phase. When output is DC, -10V ~ 10V voltage value to set the voltage between negative full range and positive full scale. When output is AC, 0-10V voltage value to set the 0 to full scale.

Enable/disable analog control

The user needs to select the corresponding function settings in the System menu. The detailed parameter description is as below.

External program	External analog function		
	Status	Set the ON/OFF state	
		 AM: Adjust the amplitude 	
	Mode	 Amplifier: Monitor the real-time output 	
		values	



	Monitor phase	When under 3-phase mode, select the phase to be monitor. Invalid under single phase mode.
	U ratio	Set the ratio between the external signal and the output voltage. Can be set to 50V/1 or 100V/1
	I ratio	Set the ratio between the external signal and the output voltage. Can be set to 5A/1 or 10A/1

- 1. Press the composite keys [Shift] + (System) on the front panel to enter the system menu.
- 2. Press the up/down key or rotate the knob to select **Source->External programme** and press [Enter].

Set the External control state and control mode, press [Enter].


Remote Control

Through the analog input interface, the analog signal can be input to remotely set the output voltage value or power amplification function. For the detailed function definition of the pin, please refer to the definition of the analog pins description. The following is an example of amplitude modulation to introduce how to connect and how to use.

When the voltage setting is controlled through the analog interface, the external voltage $(-10V \sim 10V)$ is connected to program the voltage value between 0 and full scale(AC mode).

For example, analog control AC range of 0~350V voltage, when the analog signal voltage is set to 5V, the voltage ratio set to 50V/1 under source menu, the instrument output voltage is set to 5*50=250V.

When parallel machine operation, can be controlled through the host analog interface.






Power Amplification


Analog signals can be input through the analog input interface and power amplification function can be realized. The operation method is as follows:

Connecting the analog interface, different interface function are different, please refer to pins description as above.



Voltage and current monitoring

Through the analog interface, the existing output voltage/current can be monitored. Connect a digital voltmeter or oscilloscope between pin 54 (V_Monitor), pin 6 (I_Monitor) and ground wire 4 (GND) of the analog interface. The -10V \sim 10V voltage reading corresponds to the power voltage and current output between negative full range and positive full scale (For AC, 0 to 10V corresponds to 0 to full scale). The wiring diagram is shown in the figure below.





Chapter7 Current Source Mode Operation

When the user sets the device operation mode to Current Source mode in the system menu, the instrument is a constant current source. This chapter describes the functions and features of the instrument in Current Source mode.

Only Normal function and List function are supported in constant current source mode, and the function usage is similar to voltage source.

7.1 Select the Power Supply

The IT7900 series instrument can be used as an AC constant current power supply. In the current source mode the power supply supports only single phase and reverse phase modes.

The operation steps are as follows.



- 1. Press [Shift] +
- (System) enter to system menu.
- 2. Under the **Source** setting interface, touch the screen or rotate the knob to select the **Phase** and set the power supply phase mode.

Single Phase

Under the single phase, the power supply works as a single phase power supply. the output mode can be set to AC/ AC+DC.

Reverse Phase

Under the reverse phase, the power supply works as a single phase power supply, and the rated voltage will be extended to 200%. For example, AC power supply rated voltage is 350V, under the reverse phase mode, the reted output voltage is 700V. And the output mode can be set to AC/AC+DC.

7.2 Select the Output Mode

The IT7900 series has four output modes: AC, DC, AC+DC, DC+AC. It not only provides pure AC/DC output, but also can use AC+DC and DC+AC output modes to realize "AC output plus DC bias" And "DC output waveform with ripple", which cover a wider range of applications.

The output mode can be select in the system menu.





(System) enter to system menu.

2. Under the Source setting interface, touch the screen or rotate the knob to select the **[Output couple mode]** and set the output mode.

7.2.1 AC Output Mode

If the output mode select to AC Mode, the instrument will simulate AC power supply. The default set of IT7900P series power supply is AC Mode.

Set the output parameters of the power supply in the main interface, including the output voltage, output frequency.

- Press up/down keys to select setting value and then press Enter to confirm.
- Rotate the knob to select setting value and then press Enter to confirm.



Touch screen and then press Enter to confirm.



7.2.2 AC+DC Output Mode

If the output mode select to AC+DC Mode, the instrument will simulate AC and DC power supply, which can add DC component to AC voltage. Set the output voltage in the main interface, as shown in the figure below.



Under AC+DC mode, Set the output voltage and frequency of the AC power supply in the main interface, and set the DC voltage in the configure menu.

 Vac: you can set the Vac under the main interface or under the config interface.

DC: set the DC component under the config interface.



Chapter8 Measurement Functions

This chapter describes the characteristics and operations of the basic metering function of IT7900 series source.

IT7900 series source has rich functions of basic metering of electric energy and can accurately measure the parameters such as Vrms, Irms, Ipeak, Idc, CF, PF.

8.1 Meter Mode

Press on the front panel to enter the metering interface. Different modes display different interfaces, but the displayed measurement parameters are the same. Take AC single-phase mode as an example, the measurement interface is shown in the figure below.



Description of the metering parameters:

Parameter	Description
Α	Single mode
Voltage Vrms	Setting voltage
	AC/ACDC mode: Vac setting value
	DC/DCAC mode: Vdc setting value
Frequency Hz	Setting frequency
	Only valid in AC/ACDC/DCAC mode.
Р	Active power [W]
PF	Power factor
CF	Crest factor
Ithd	Current THD
Uthd	Voltage THD
Ipeak	Current peak value [A]
lp+	Positive current peak value [A]
lp-	Negative current peak value [A]
S	Apparent power




Parameter	Description
Q	Reactive power
Udc	DC voltage
Idc	DC current
Sa	Total three-phase apparent power
ΡΣ	Total three phase Active power
PFa	PFa=P_total/S_total

8.2 Oscilloscope Mode

IT7900 series source has the function of displaying the waveform based on sampling data. The user can select to display or hide the voltage and current waveform of the input unit. Only the necessary waveform is displayed, which can facilitate observation. The waveform display interface includes the vertical axis and horizontal axis.

Press on the front panel and the following waveform display interface will appear. Different modes display different interfaces. the Oscilloscope interface is shown in the figure below.



Description of keys on the waveform display interface:

Voltage/Current/Time: Adjust voltage/current/time base range

Single: Single measurement key: when single measurement is enabled in the Stop status, the stop status is enabled again after one measurement based on the current data updating rate. When single measurement is enabled in the Ready status, the instrument immediately restarts one measurement and then enters the Stop status.



Run/Stop: press the corresponding soft key to run or stop the waveform status. AutoSet: Automatically adjusts the scale of the appropriate vertical axis.

- : Advanced configuration options for oscilloscope function.
- Trigger source: Select trigger source, Voltage/Current and rise edge or fall edge can be select.
- Trigger mode: Auto and Normal can be select.
- Print data: save the data
- Line selection: Select the displayed curve, which is used to select whether to display the voltage/current waveform of the corresponding phase. Up to 6 oscillographic data curves can be displayed.

U↑: Voltage trigger setting

Vertical calibration

The voltage range and current range are subject to vertical calibration (voltage/scale and current/scale). Press the soft key **[Function]** and select Volt-Range or Curr-Range. Rotate the knob to set the voltage or current range of each interval.

Horizontal calibration

If the soft key **[Function]** is pressed and "Trig-TimeBase" is selected, you can rotate the knob to adjust the horizontal scale (scanning speed). When the horizontal (time/scale) setting is changed by rotating the knob, you can observe the change of time/scale on the screen. During data collection, the sampling speed can be changed by adjusting the horizontal calibration knob. After collection is stopped, the collected data can be amplified by adjusting the horizontal calibration knob.

Trigger waveform

When the specified trigger conditions are satisfied, the trigger waveform will be displayed. The triggering time is the trigger point, generally on the right of the screen. When the trigger point is reached, the screen will display the waveform from left to right over time. The user should set the following parameters before using the trigger function.

Trigger mode

The trigger mode refers to the condition to update the contents on the screen. It is divided into the Auto mode and Normal mode. In the Auto mode, the displayed waveform will be updated when triggering occurs in the suspension time; otherwise, the displayed waveform will be updated automatically.

In the Normal mode, the displayed waveform will be updated in the case of triggering and not updated in the case of no triggering.

Trigger source

The trigger source is used for generating trigger conditions. The user can select the trigger source in the input signal of the input unit.

Trigger slope



The slope refers to the change of the signal from low level to high level (rising edge) or from high level to low level (falling edge). The slope used as a trigger condition is referred to as the trigger slope.

Trigger level

The trigger level refers to the level which the trigger slope passes through. If the signal of the trigger source passes through the set trigger level according to the specified trigger slope, triggering occurs. When the soft key **[Function]** is pressed and "Trig-Level" is selected, you can rotate the knob to adjust the trigger level. In this case, the trigger level can be changed by rotating the knob and you can observe trigger level changes on the screen.

Print data

In the advanced menu, users can select the Print Data item and select the data logging mode. Data of oscilloscope interface will be recorded to U disk.

Print data mode:

- Off: turn off the print data function.
- Post: The recorded data is consistent with the data displayed on the oscilloscope interface
- Raw: The recorded data is original data, The default data sampling interval is 10us
- Both: Post and Raw, record two data file.


8.3 Harmonic measurement

IT7900 series source can display harmonic parameters in the list or bar chart form to make the analysis of test result clear.

8.3.1 Harmonic measurement

Press key on the front panel, and the following initial interface of harmonic measurement will appear.

Harmonic bar chart screen





Description of keys on the harmonic display interface:

: Select the voltage/current harmonic

": Distortion factor calculation formula.

%r: displaying harmonics in the form of percentage to the overall voltage amplitude of all harmonics.

%f: displaying harmonics in the form of percentage to the fundamental voltage.

THD: Under single phase mode, display total harmonic distortion of the single phase. Under three-phase mode, display total harmonic distortion (THD) factors corresponding to Phase A, B and C.

Single harmonic parameters: Under single phase mode, display single harmonic distortion factor and phase of the single phase. Under three-phase mode, display the single harmonic distortion factor and corresponding phase. Rotate the knob to select the number.

Harmonic list screen

When LIST mode is selected in the harmonic measurement mode, the list of harmonic measurement results will be displayed, when "U" is selected, the list will display the voltage and phase values under various harmonics. When "I" is selected, the current and phase values will be displayed. Under single-phase mode, the list only displays measurement results for Phase A, and displays 0 for the other two phases, as shown in the figure below; under three-phase mode, display effective values for the three phases.




Harmonic number list: the signal data of 0-50th harmonic(s) will be shown in the LIST. You can press the Up and Down key to display the hidden rows, i.e. hidden data of single harmonic data.

Introduction to vector interface

When vector mode is selected in the harmonic measurement mode, to enter the vector measurement interface, as shown in the figure below.





8.3.2 Harmonic Analysis

This series of instruments can set the analysis conditions during harmonic measurement, filter the harmonic measurement data according to the set conditions, and mark the corresponding non-conformities in red. It is convenient for users to do harmonic test analysis. The set conditions can be THD value of voltage and current, THC, POHC, PWHC and IEC61000-3-2/IEC61000-3-12 regulation items.




Item	Description
Total THD observe setting	THD interface observation switch, this setting is globally
	effective and takes precedence over each regulation
	module setting item.
	Enable: Harmonic analysis according to conditions
	Disable: Does not perform harmonic analysis
U-THD	Total harmonic distortion of voltage
I-THD	Total harmonic distortion of current



Item	Description	
THC	total harmonic current (The total rms value of the	
	harmonic current components of order 2 to 40)	
POHC	Partial odd harmonic current	
PWHC	Partial weighted harmonic current	
Catagory	Category of Regulations	
Subset (IEC61000-3-2)	Class A:	
	Harmonic order h	
	Max harmonic current Ih/Iref	
	Class B	
	Harmonic order h	
	Max harmonic current Ih/Iref	
	Class C	
	Harmonic order h	
	Max Permissible Ithd%	
	Class D	
	Harmonic order h	
	Max Permissible Ithd per watt mA/W	
	Max harmonic current lh/Iref	
	Note: Parameter items are displayed only and cannot be	
	edited	
Subset (valid for Self-	I THD	
defined)	U THD	
	U/I THD	
	1. Form parameter items can be edited, 51 parameters can	
	be edited	
	2. I THD or U THD title suffix %f/%r dynamically changes	
	according to the harmonic interface configuration items	
	3. I THD/ U THD, The parameter value is set to -1 to delete	
	the current edit line parameter	
	4. After editing the parameters, sort them dynamically	
Tuno (IECC1000 2.1)	from smallest to largest by harmonic order h settings	
Type (IEC61000-3-1)	Phase type Non-balance 3-Phase	
	Balance 3-Phase	
	Balance 3-Phase Balance 3-Phase(a,b,c)	
	Balance 3-Phase(d,e,f)	
	Note: Table parameter entries are displayed only and are	
	not editable.Iref: reference current, Ih: harmonic current	
	component	
Rise (IEC61000-3-12)	Minimum short-circuit ratio limit value	
11130 (10001000 3 12)	iviliani short circuit ratio mint value	

After setting, press Esc to return to the harmonic measurement interface, and the instrument measurement interface will display the results after harmonic analysis.





8.4 Recorder Function

The data recording function allows you to observe and record output status data for a long time. On the recorder interface, you can select a maximum of six data curves to be displayed. See the figure below.



Stop: Stop refresh data.

Clean: Clean all of data curve.

More: Enter to the advanced menu to set more parameters.

- Viewing control: Select the data curve, up to 6 curves can be select.
- File format: select the file format, include Tdms and CSV.
- Export to udisk: exporting all of present data to U disk.
- Meter sample: Set the sampling interval.

Hold-On/Hold-Off: Pause screen data refresh (for data observation)/ Start



dynamically observing the data

Auto: Automatically adjusts the scale of the appropriate vertical axis.

Time: The time value of each of the horizontal coordinates, unit is s/Div

Vernier: Position information of the vernier caliper.



Chapter9 Configuration Arbitrary Waveform

This chapter will introduce configuration of arbitraryny waveform of IT7900. Besides the five standard waveforms (see 5.6 Waveform Selection) built in the main interface, the user can enable the outputting of AC wave sequences of different amplitudes through List and self-define function.

9.1 List Function


At maximum of 200 steps can be edited for each List file. The user can choose any waveform from existing 11 waveforms. Or, the user can choose a waveform file stored in the device through self-define function (see 8.3 "Self-defined Waveform Function" for details) and combine the waveform files into a List file based on required order. Meanwhile, the user can set the starting steps of the List file, run count and end state of the waveform, and edit parameters such as frequency, amplitude, running time and rising slope of each waveform.

9.1.1 Create a new List file

Create a new single-phase List file

Under single-phase mode, the user can output AC waveform sequences with different amplitudes by creating a new List file. Detailed operation steps are as below:

1. Press [Shift]+[V-set](**list**) on the front panel to enter the List function configuration interface, as shown in the figure below.



888.csv: the list file name to execute. Trig source: select the trigger source Run/Stop: Run/stop the list function. Open: Select the List file to execute.


New: Create a new List file.



Edit: Edit present list file

Delete: delete the present List file.

2. Press [New] and enter to the List file edit interface.



List edit description:

Description: Description of List, display list file name.

Repeat: Edit the cycles of the List file.

End: Set the final waveform, with the following options available:

- Off: directly off the output after operation.
- Normal: return to normal after operation.
- Last: keep the last waveform output unchanged after operation.

No.: step number of list. Click the numer, you can operate such as copy/paste/cut/insert/delete.

ACrms V: Voltage RMS value

Freq Hz: Frequency

Time S: width time, range from 0.0001s to 42949s

More: other settings, click ... and setup the slew rate, waveform, phase and so

Export: Export the internal List file of the instrument to an external USB flash drive.

Save: Save the list file.

Config: configure the list file to make it effective.

Clear all: delete all of step information

Trig source: select trigger source

3. Click (More)... enter to advanced menu of list file.





List parameters description:

Voltage AC	Voltage value and slope.
Start Phase	Start phase setting, displayed in AC mode. By default, the program automatically calculates this value (which is guaranteed to be continuous with the previous step). If the user modifies this parameter, the user set value is used without guarantee of continuity.
Phase Difference	Phase difference between ABC, only displays in AC 3-phase mode.
Frequency	Frequency and slope, displayed in AC mode.
Waveform	Waveform type, every basic waveform can be selected, displayed in AC mode.
Step jump	Method of step jumps to next step. Time: when the time is out, jumps to next step Trig: receive a trigger signal, jumps to next step. Phase: jumps to next step at this phase.
Trig out	Whether outputs a signal when this step is ending.

- Press [Esc] to return Edit interface, Press [Save].
 At list Edit interface, click the step number, the [Insert]/[Paste]/[Cut]/[Copy] /[Delete] will display, click the key to edit.
- 5. Press [Esc] to return.

9.1.2 Select/Run List File

If several List files are edited, press Recall to recall the List file to be tested. Detailed operation steps are as below:

- 1. Press [Shift]+[V-set](list) on the front panel to enter the List function configuration interface.
- Press [Open], select the saved List01 csv file, and press [Enter] to enter the file.
- 3. Press [On/Off] on the front panel, turn on the output.
- 4. Press [Run] in the list function interface.



Running indicator will appear in interface.Press [Scope] key to view the output waveform.

9.1.3 Import/Export List file

Import List file


IT7900 series support import list file function, The user can finish the editing of List file in Excel and import it into the software. This function simplifies the List file edit and facilitates user operation.

To help user define an Excel file format, please export a CSV template from the List interface.

Detailed operation steps are as below:

- Create a new Excel document on local PC and name it List02.
- 2. Open the Excel document and save it as in "other formats" i.e. "(*.csv)".
- 3. Open the List02.csv document and edit the List. Set every step of the List and corresponding parameters and save the document in the USB disk.

List import file formats under single-phase mode:



- 4. Insert the USB disk into the USB interface of the front panel. Press [Shift]+[V-set](list) on the front panel to enter the List function configuration interface.
- Press [Open]. Select the List02.csv file and open it. The List file will be imported.

Export List file

After editing the List file, the user can directly save it into the device or export and save it into the peripheral memory disc. The exported List is saved in the format of. (*.csv). Detailed operation steps are as below:

- 1. Insert the U disk into the USB interface of the front panel.
- Press [Shift]+[V-set](list) on the front panel to enter the List function configuration interface.
- Select [Edit], enter to list file edit interface.
- 4. Press [Export]. This file will be exported into the USB disk.

9.2 Setting of Surge/Sag Configuration

IT7900 provides surge/sag simulation. The user can add surge/sag to simulate abnormal voltage fluctuation on the basis of outputting wave, and test usage of the DUT under this circumstance.



Surge/sag can be added to any waveform, the basic waveform is selected from the Config menu.

The Surge/Sag function is not supported in multi-channel mode.



When using the dimming function, this function will automatically turn off, and when using this function, dimming function is turned off.

Operating steps

1. Press [Shift]+ (Surge&Sag)on the front panel to enter the List function configuration interface.



2. Set the voltage and frequency parameters in the sweep interface.

Parameters in the sweep interface are described as follows:

Parameter	Description
Mode	surge/sag executing mode.
	Trig: Trigger mode. Executing the surge/sag after receiving the trigger signal.
	Period: Period mode. Execute the surge/sag based on the period.
Action	This setting is valid when the mode select to Trigger.
	Imme: Executing the surge/sag immediately.
	Angle: Executing the surge/sag at specific angle.
Trig source	Select the trigger source when the mode select to Trigger.
Start angle	Set the start phase angle of the surge/trap when the mode select to trigger.
Angle width	Set the period of the surge/trap. For example, start angle=30 degree, Angle width=30 degree, then, the waveform will execute surge/sag at 30 to 60 degree.



Parameter	Description		
Symmetry	Whether to produce symmetrical surge/sag waves.		
	If Start angle + Angle width >180°,this setting is Off state.		
Repeat count	repeat count of surge/sag waves.		
Period count	Number of cycles to generate surge/sag waves. This parameter is meaningful only in Peroid mode.		
	This setting is used in combination with Repeat count, for example, if Repeat count is set to 5 and Period count is set to 10, five surge/sag waves occur in every 10 cycles.		
Enable	Phase A/Phase B/Phase A&B/Phase B&C/Phase A&C/PhaseA&B&C:		
	Select the phase information where the surge/notch occurs.(displays under 3-phase mode)		
Enable	Synchronize: Each of the three phases executes surge/trap at the same time.		
	Specify Phase: Each of the three phases executes surge/trap at the specified phase.		
	(displays under 3-phase mode)		
Value select	Value select:		
	Percent: Set the percentage of the surge/trap amplitude to AC signal amplitude (RMS).		
	Setting: Set the value of the surge/trap amplitude.		

- 3. Press [On/Off] on the front panel and enable the output.
- 4. Press [Run] on the interface, Then the running status will be displayed on LCD. Output parameters and measurement parameters are displayed on the left side of the interface. You can also press the Meter to observe the output parameters in the main interface.
- 5. After execute, [On/Off] on the front panel will be off, and running status will be displayed on LCD. You can press [Stop] on the interface to stop the surge/sag function.

9.3 Self-defined Waveform Function


In AC mode or AC+DC mode, you can customize waveform curves and save them locally as output waveform options. This self-defined waveform can be used as a normal output waveform or as a wave option for scanning waveform, Surge/Sag waveform, and LIST waveform.

9.3.1 THD

The distorted waveform can simulate voltage harmonic wave in the circuit. The user can set the extent to which the output voltage waveform deviates from the Sin wave voltage and test the usage of the DUT under this circumstance.

THD includes built-in 30 waveforms and user - defined waveforms. The interface is shown below.





DST wave: Select waveforms which built into the instrument by default. When the waveform name is selected, the parameters and waveform diagram are displayed on the right side of the interface.

User THD wave: THD wave of user defined.

- "+" Create a new THD wave.
- "-" Delete the seleted THD wave.
- "I" Edit the THD wave.

Press the "+" or "I" enter to the edit interface.



Thd profile: THD file name

Thd formula: Distortion factor calculation formula.

%r: displaying harmonics in the form of percentage to the overall voltage amplitude of all harmonics.

%f: displaying harmonics in the form of percentage to the fundamental voltage.



Delete: select a row and click Delete.

Open: import Thd wave data.

Save: Save the THD wave.

Back: return back upper menu.

THD=: Total distortion rate calculated based on the user Thd configuration.

Import THD file

Α	В	С	D
Model	ITxxxx	Device_operation	0
Usage	Wave		
Name	Untitled-01.csv		
Туре	8		
Editable	1		
Formula	0		
Unit_number	39		
Fund_phase	0		
Order	Thd	Phase	
2	5	0	
3	9	180	
A	2	0	

Users can edit a custom harmonic waveform file in .csv format and import it into the instrument. User defined harmonic waveform template files can be obtained by contacting ITECH, or an empty file can be exported directly from the instrument.

The harmonic waveform template description as follows:

Model: instrument model, Keep default.

Device_operation: operation mode of instrument, 0: voltage source, 1: load, 2: current source, Users are cautious to modify, the waveform files are not common in different modes.

Usage: Usage, keep default

Name: keep default

Type: Waveform type, harmonic custom waveform is 8, which must be 8 in this

ile.

Editable: keep default

Formula: Harmonic calculation formula, 0 represents f%, 1 represents r%

Unit_number: The number of data points edited needs to be consistent with the

number of data rows.

Fund phase: phase of fundamental waves, range from 0-360°

Order: harmonic order, range from 2-50

9.3.2 User-defined waveform

"+" create a new wave.


"-" delete the wave file.

"I" edit the wave file.





Edit interface:



Profile: user-defined file name

Origin Symmetry: To select the waveform data type, you can select 512 origin

symm/512 origin asymm /1024 points

Open: import waveform data.

Save: Save the user-define wave.

Delete: select a row and click Delete.

Clean: delete all of data

Back: return back upper menu.



Import user-define waveform file

4	Α	В	С	D
	Model	IT7800	Device_operation	0
	Usage	Wave		
	Name	Userdefine.csv		
	Туре	9		
	Editable	1		
	Point_number	1024		
	Origin_symm	2		
	index	fval		
	0	0		
)	1	-0.0184		
L	2	-0.0368		
2	3	-0.0552		
3	4	-0.0736		
1	5	-0.0919		
5	6	-0.1102		

Users can edit a user-defined waveform file in .csv format and import it into the instrument. User defined harmonic waveform template files can be obtained by contacting ITECH, or an empty file can be exported directly from the instrument.

The user-defined waveform template description as follows:

Model: keep default

operation mode of instrument, 0: voltage source, 1: load, 2: current source, Users are cautious to modify, the waveform files are not common in different modes.

Usage: Usage, keep default

Name: keep default

Type: Waveform type, user-defined waveform is 9, which must be 9 in this file.

Editable: keep default

Point_number: The number of data points to be edited, which needs to be consistent with the data row in the table. Associated with the Origin_symm parameter.

Origin_symm: symmetry option, 0 represents 512 non-origin symmetry, 1 represents 512 origin symmetry, 2 represents 1024 points.

Index/fval: Data points and detailed parameter values.


9.4 Voltage Sampling

The IT7900 series instruments can be used as data sampling devices to collect voltage variation signals from other instrument. With this function, users can realize the acquisition of the actual grid voltage change state of the DUT, and then use the import function to reproduce the grid state.

Connecting method


When the voltage signal is sampling, the wiring method is the same as the test line connection method when outputting. For single-phase, connect the two terminals A and N, and for three-phase, connect the four terminals ABCN. For details, refer to section 2.5 Connecting the Test Lines, and take the sampling DUT supply voltage signal as an example, the wiring diagram is shown below.





Operation method

Click Menu > Volt sample and enter to Voltage signals sampling interface, the display is shown below.



- File: Input the name of the sampling file, the default display is null, you need to input the file name before sampling.
- 4: Represents the number of sampling. If you click Run once, the number of sampling will be increased once and a test-1 file will be formed.
- Tdms: the file format, can be select to Tdms or CSV.
- Export: export the sampling file.
- Run: start to sampling

Setting the configuration

Click Configure in the voltage sampling interface to enter the parameter setting interface.





- Sampling rate: Sampling rate, indicating the sampling interval of data points.
 Setting range 1-50, unit kHz
- Start Condition: setting the condition of sampling

Imme: Sampling is executed immediately after clicking Run.

Trigger1: After clicking Run, wait for the Trigger1 signal to be received to start executing sampling.

Trigger2: After clicking Run, wait for the Trigger2 signal to be received to start executing sampling.

Capture-Vac: The voltage signal sampling after compared. When this mode is selected, additional parameters need to be set.

Trigger delay: Set the sample data delay time. Set a positive value to record data after the capture event occurs. A negative value means that data before the capture event occurs is recorded.

Compare: Select the object of comparison, either Vac/Rms or Vac/Peak

Up Limit: AC voltage upper limit value, grid voltage greater than the upper limit value will start sampling.

Down Limit: AC voltage lower limit value, grid voltage less than the lower limit value will start sampling.


9.5 Voltage Signals Simulation

In AC mode or AC+DC mode, users can import TDMS waveform curve files for simulation, and set parameters such as repetition times and waveform offset for imported waveform.

By default, the open waveform is only for real-time simulation. If you remove the USB disk, the waveform disappears. Users can also permanently save the open waveform to the power supply memory.

Directly click Simulation in the Menu to enter the voltage signal Simulation interface, as shown below:





Start mode: can be selected as immediate execution or trigger execution.

Interval time: The interval between waveform repeats. This parameter is displayed only when start mode is selected for Imme.

Action: Displayed when Trigger1/trigger2 is selected for start mode, Imme or Phase can be selected.

∠: Displayed when Phase is selected for Action, sets the starting angle value.

Repeat: Number of repeated execution of waveform

Coefficient: Waveform amplification Coefficient

DC: DC offset

Open: Open the waveform file.

Delete: Delete the waveform file.

Import: Import and save the waveform file to the power supply memory.


Export: Export the waveform file to the U disk.

Run: Start running waveform. Power output should be turned on before operation.

9.6 Interharmnics function

Users can click the Interharmonics icon in the Menu interface to enter the Interharmonics function interface and edit the parameters such as interharmonic components to simulate the interharmonic output waveform.





Parameter descriptions of the interharmonic interface

Catalogue: Function category selection, Normal general mode, Sweep scanning mode.

Startup: start mode, select to Trig or Imme.

Trig source: Trigger source, when the Startup is Trig need to set.

Parameter introduction when Normal mode is selected to the function category

Selection: Calculation method selection, %f is equivalent to base wave percentage mode, Voltage direct voltage setpoint mode.

Thd levels: The interharmonic distortion rate setting, depending on Selection, can be a percentage and a voltage setting.

Frequency: Interharmonic frequency setting value.

Phase: Interharmonic phase setting value.

Timer type: Select the timing reference, Time method or Cycle method.

Hold time: Interharmonic operation time. If it is Cycle mode, it is the number of interharmonic operation cycles (Hold cycles).

Normal time: Normal waveform running time. If it is Cycle mode, it is the number of normal cycles.

Repeat counts: The number of times the interharmonics and normal waveforms are run alternately.

Parameter introduction when Sweep mode is selected for the function category

Selection: Calculation method selection, %f is equivalent to base wave percentage mode, Voltage direct voltage setpoint mode.

Thd levels: The interharmonic distortion rate setting, depending on Selection, can be a percentage and a voltage setting.

Start freq: Interharmonic sweep start frequency setting value.

End freq: Interharmonic sweep end frequency setting value.

Timer type: Select the timing reference, Time method or Cycle method.




Step time: Interharmonic scanning step time. If Cycle mode, it is the number of simple harmonic operation cycles (Step cycles).

Step freq: Interharmonic scanning step frequency.

Start angle: Interharmonic start phase angle.

9.7 Aircrafts Power Interruption Test

This instrument has built-in aircraft power interruption test regulation waveforms, and users can select the corresponding waveforms directly in the interface.



Parameter	Description
Category	Aircraft power interruption test category
	Users can select the test item waveforms in the Test1-
	Test7 category, or choose to customize them.
Frequrecy Hz	Frequency setting, default setting is 400Hz
	Can be set in customized mode.
Phase	Phase selection, three-phase can be any combination of choice
Trig source	Trigger source selection:
	Manual
	● Bus
	● Trig1
	● Trig2
Mode	interruption start mode:
	Time / Angle
	Time / Immediate
	Trig / Angle
	Trig / Immediate
Angle	Setting the angle at which power interruption starts.
	Only displayed in Time / Angle and Trig / Angle mode.
T1	Total drop time, settable in customized mode. Unit: ms
T2	Drop time from 90% to 10%. Unit: ms
T3	Recovery time from 10% to 90%, Unit: ms



Chapter 10 Technical Specifications

This chapter will introduce the main technical parameters of IT7900, such as rated voltage/current/power and so on. Besides, this part will introduce the working environment and storage temperature.

10.1 Supplemental characteristics

Recommended calibration frequency: once a year

Cooling style: fans

10.2 Main technical parameters

IT7902-350-10U

parameters				
	Wiring connection		3 phase 3wire + ground(PE)	
	Line voltage	RMS	(200~220) ±10%	V
			(380∼480) ±10%	
	Line current	RMS	< 8.1	A
AC Input	Apparent power		< 2.4	kVA
	Frequency		45~65	Hz
	Power factor	typ	0.98	
ut parameter	'S			
	Output voltage	VLN	0~350	V
		RMS (1phase)	10	A
	Output current	Crest Factor *1	6	
		Peak (1phase)	40	A
	Output power	Max. Power	2k	VA
		(1phase)		
	Voltage setting			
	Range	1phase	0~350	V
	Resolution		0.01	V
	Accuracy	16Hz~150Hz	< 0.1%+0.1% F.S.	
	Temperature		< 100ppm/°C F.S.	
	coefficient			
	DC Voltage Offset	typ	0.02	Vdc
	Current Limit setting	RMS	10	A
	Range	(1phase)	10	^
AC Output	Resolution	(Tpridoo)	0.01	A
	Accuracy	16Hz∼150Hz	< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/°C F.S.	
	Frequency			
	Range		16~150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform	50/60Hz	up to 50	orders
	synthesizer			
	Phase			
	Range		0~360	0
	Resolution		0.01	۰
	Voltage setting			
	Range	1phase	-499~499	Vdc
	Resolution		0.01	V



IILUII				Technical Spe
	Accuracy		<0.1%+0.1% F.S.	
	Temperatur		< 100ppm/°C F.S.	
	е		₹ 100ррпі/ € 1.3.	
	coefficient			
	Current setting			
	Range	1phase	-10∼10	Adc
DC Output	Resolution		0.01	A
	Accuracy		< 0.1%+0.2% F.S.	
	Temperatur		< 200ppm/°C F.S.	
	е			
	coefficient			
	Max. power			
	Total power	Max. Power	2k	W
		(1phase)		
	Line regulation		<0.05% F.S.	
	Load regulation	DC,16Hz~150Hz	<0.05% + 0.05% F.S.	
	THD *3	16Hz∼150Hz	<0.5%	
Voltage stability	Voltage ripple	RMS	< 0.4	V
	*4	typ	220	us
	Dynamic response			
	R Range	1phase	0~1000	mΩ
Programmabl e	L Range	1phase	0~1000	uH
impedance				
pedanee	P Range	1phase	0~2	kW
		·		
	QL Range	1phase	0∼2	kVar
	QC Range	1phase	0∼2	kVar
	R Range	1phase	1~1000	Ω
Anti-Islanding RLC	L Range	1phase	1~5000	mH
	C Range	1phase		mF
Valtage Class	C Kange	·	0.001~5	1111
Voltage Slew Rate,		≥2 V/µs with full-scale programmed	voltage step	
Typical				
Output Isolation		550Vac		
Measurement parai	neter			
	Resolution		0.01	V
Voltage RMS	Accuracy	DC,16Hz~150Hz	<0.1%+0.1% F.S.	
	Temperature		< 100ppm/°C F.S.	
	coefficient			
	Resolution		0.01	A
Current RMS	Accuracy	DC,16Hz∼	<0.1% + 0.2% F.S.	
		150Hz		
	Temperature		< 200ppm/°C F.S.	
	coefficient		0.01	A
Peak current	Resolution Accuracy	16Hz∼150Hz	<0.4% + 0.6% F.S.	^
	-	10⊓2: ~ 130⊓2		LAM
Output power	Resolution Accuracy	DC 16U	0.001	kW
	Accuracy	DC,16Hz∼ 150Hz	<0.4% +0.4% F.S.	
Harmonic	Max.		un to EO	orde
measurement	IVIAX.	50/60Hz	up to 50	orders
Regenerative	<u> </u>			
Max.		2k		VA
Regenerative				
power				
ITHD		< 5%		
Others				
Efficiency	typ		84%	
	OVP, OCP, OPP, OTP, FAN, Sense			
Protection				
		OVP, OCI 437.00mm (W) *88.40mm (H) *700		



Weight	15kg
Working	0°C-50°C
temperature	
Programming	2ms
response time	
Remote Sense	20V
Compensation	200
Voltage	
Communication	Built-in USB/CAN/LAN/Digital IO interface, optional GPIB / Analog&RS232
interface	

^{*1}Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.

IT7904-350-20U

ut parameters	T T				
	Wiring connection	3 phase 3	3 phase 3wire + ground(PE)		
	Line voltage	RMS	(200~220) ±10% *1 (380~480) ±10%	٧	
AC Input	Line current	RMS	< 9.4	Α	
	Apparent power		< 4.6	kVA	
	Frequency		45~65	Hz	
	Power factor	typ	0.98		
tput parameter	S				
	Output voltage	VLN	0∼350	V	
	output voltago	VLL (reverse)	0~700	V	
		RMS (1phase)	20	A	
		Crest Factor *2	6		
	Output current	Peak (1phase)	80	А	
		RMS (multichannel/reverse)	10	А	
		Peak (multichannel/reverse)	40	А	
		Per Phase/Per Channel	2k	VA	
	Output power	Max. Power (reverse phase)	4k	VA	
		Max. Power (1phase/multichannel)	4k	VA	
	Voltage setting				
	Pango	1phase/multichannel	0~350	V	
	Range	reverse	0~700	V	
	Resolution		0.01	V	
AC Output	Accuracy	16Hz∼150Hz	< 0.1%+0.1% F.S.		
	Temperature coefficient		< 100ppm/°C F.S.		
	DC Voltage Offset	typ	0.02	Vde	
	Current Limit setting				
	Range	RMS (1phase)	20	А	
		RMS (multichannel/reverse)	10	А	
	Resolution		0.01	Α	
	Accuracy	16Hz∼150Hz	< 0.1%+0.2% F.S.		
	Temperature coefficient		< 200ppm/°C F.S.		

^{*3} test condition is: Pure resistive loads, full power conditions.

^{*4} The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10uF



	Range		Technical S $16\sim150$	Hz
			16∼150 0.01	
	Resolution			Hz
	Accuracy		0.01%	
	waveform synthesizer	50/60Hz	up to 50	orde
	Phase			
	Range		0~360	0
	Resolution		0.01	0
	Voltage setting		5.5.	
	Tonago coming	1phase/multichannel	-499∼499	Vdo
	Range	· · · · · · · · · · · · · · · · · · ·	-998~998	Vdd
	Resolution	reverse	0.01	V
				V
	Accuracy		<0.1%+0.1% F.S.	
	Temperature coefficient		< 100ppm/°C F.S.	
	Current setting			
	_	multichannel/reverse	-10∼10	Add
DC Output	Range	1phase	-20~20	Add
	Resolution	1	0.01	A
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature			
	coefficient		< 200ppm/°C F.S.	
	Max. power			
	Channel power	Per Channel	2k	W
	Max. power (reverse	Max. Power (reverse phase)	4k	W
	phase) Total power	Max. Power (1phase/multichannel)	4k	W
	Line regulation	(Tphase/matterialmer /	<0.05% F.S.	
	Load regulation	DC,16Hz~150Hz	<0.05% + 0.05% F.S.	
	THD *4	16Hz~150Hz	<0.03% + 0.03% 1 .3. <0.5%	
Voltage stability		RMS	< 0.4	V
	Voltage ripple	RMS	< 0.4	V
	*5 Dynamic response	typ	220	us
		multichannel	0~1000	mΩ
	R Range	1phase	0∼500	mΩ
Programmable		reverse	0∼2000	mΩ
impedance		multichannel	0∼1000	u⊦
	L Range	1phase	0~500	ul
		reverse	0~2000	ul
	+	1phase	0~4	kV
	P Range	reverse	0~4	kW
	-	1phase	0~4	kVa
	QL Range	reverse		kVa
			0~4	kVa
	QC Range	1phase	0~4	
nti-Islanding RLC		reverse	0~4	kVa
	R Range	1phase	0.5~500	Ω
	_	reverse	2~2000	Ω
	L Range	1phase	0.5~2500	m⊦
		reverse	2~10000	mF
	C Range	1phase	0.002~10	mF
	- Nange	reverse	0.001~2.5	mF
Voltage Slew Rate, Typical		≥2 V/µs with full-scale programme	d voltage step	



Output Isolation		550Vac	recinical 3p	<u> </u>
Measurement parai	meter			
	Resolution		0.01	T v
Voltage RMS	Accuracy	DC,16Hz∼150Hz	<0.1%+0.1% F.S.	•
	Temperature coefficient	50,10112 100112	< 100ppm/°C F.S.	
	Resolution		0.01	А
Current RMS	Accuracy	DC,16Hz~150Hz	<0.1% + 0.2% F.S.	
	Temperature coefficient		< 200ppm/°C F.S.	
Peak current	Resolution		0.01	А
i can cuitefft	Accuracy	16Hz∼150Hz	<0.4% + 0.6% F.S.	
Output power	Resolution		0.001	kW
Output power	Accuracy	DC,16Hz~150Hz	<0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	orders
Regenerative				
Max. Regenerative power		4k		VA
ITHD		< 5%		
Others				
Efficiency	typ		88%	
Protection		OVP, OCP	P, OPP, OTP, FAN, ECP, Sense	
Dimension		437.00mm (W) *88.40mm (H) *700.00mm (D) (Instrument frame Dimesion) 483.00mm (W) *106.90mm (H) *785.30mm (D) (Include handle, foot and protect cover)		
Weight			16.5kg	
Working			0°C-50°C	
Programming response time			2ms	
Remote Sense Compensation Voltage			20V	
Communication interface		Built-in USB/CAN/LAN/Digit	al IO interface, optional GPIB / Analog	&RS232

 $^{^{*1}}$ Max. Output derated to 60% when input with ($200{\sim}220$) $\pm10\%$ Vac 1 ph or 3 ph

IT7905-350-30U

Input parameters					
	Wiring connection	3 phase 3wire + ground(PE) *1			
AC input	Line voltage	RMS	(200~220) ±10% *2 (380~480) ±10%	V	
	Line current	RMS	< 20	Α	
	Apparent power		< 5.8	kVA	
	Frequency		45~65	Hz	
	Power factor	typ	0.98		
Output parameters					
	Output voltage	VLN	0~350	V	
		RMS (1phase)	30	Α	

^{*2} Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.

^{*4} test condition is: Pure resistive loads, full power conditions.

^{*5} The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10uF



(IIII)	<u> </u>		Technical Specifica	10113
		Crest Factor*3	6	
	Output current	Peak (1phase)	180	Α
	Output power	Max. Power (1phase/3phase/multichannel)	5k	VA
	Voltage setting			
	Range	1phase	0∼350	V
	Resolution		0.01	V
	Accuracy	16Hz∼150Hz	0.1%+0.1% F.S	
	Temperature coefficient		< 100ppm/°C F.S	
AC output	DC Voltage Offset	typ	0.02	Vdc
	Current Limit setting			
	Range	RMS (1phase)	30	А
	Resolution		0.01	Α
	Accuracy	16Hz∼150Hz	0.1%+0.2% F.S	
	Temperature		< 200ppm/°C F.S	
	coefficient		,	
	Frequency		40, 450	T U
	Range	 	16~150	Hz
	Resolution		0.01 0.01%	Hz
	Accuracy waveform	50/60Hz	up to 50	orders
	synthesizer	30/0011Z	up to 50	orders
	Phase	<u> </u>		
	Range		0∼360	0
	Resolution		0.01	٥
	Voltage setting			
	Range	1phase/multichannel	-495 <i>~</i> 495	Vdc
	Resolution		0.01	V
	Accuracy		<0.1%+0.1% F.S	
	Temperature		< 100ppm/°C F.S	
	coefficient		Τουμριίί/ C Γ.δ	
DC Output	Current Setting			
	Range	1phase	-30~30	Adc
	Resolution	ļ	0.01	A
	Accuracy	 	< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/°C F.S	
	Max. power			
		Max. Power	5k	W
	Total power	(1phase)		
Programmable	R Range	ļ	0~1	Ω
impedance	L Range		0~1000	uH
	P Range	ļ	0~5	kW
Island RLC	QL Range	 	0~5	kVar
	QC Range R Range	-	0~5 1~1000	kVar Ω
	L Range		1~1000	mH
	C Range		0.001~5	mF
	Line regulation	+	<0.05% F.S.	1111
Voltage stability	Load regulation	DC,16Hz~150Hz	<0.05% + 0.05% F.S.	
	THD	16Hz∼150Hz		
	Voltage ripple	RMS	< 0.4	V
	Dynamic	typ	200	us
/altana Olava Dari	response*4			
oltage Slew Rate, Typical	:	≥2 V/µs with full-scale programmed volt	age step	
Output Isolation		550Vac		
leasurement para			0.04	
	Resolution		0.01	V
Voltage RMS	Accuracy	DC,16∼150Hz	<0.1%+0.1% F.S.	



			<u>'</u>	
	Temperature coefficient		< 100ppm/°C F.S.	
	Resolution		0.01	A
Current RMS	Accuracy	DC,16∼150Hz	<0.1% + 0.2% F.S.	,
	Temperature coefficient		< 200ppm/°C F.S.	
Peak current	Resolution		0.01	Α
. oak oan om	Accuracy	DC,16∼150Hz	<0.4% + 0.6% F.S.	
Output power	Resolution		0.001	kW
Carpat polito.	Accuracy	DC,16∼150Hz	<0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	orders
Regenerative				
Max. Regenerative power		5k		VA
Output Current THD		< 5%		
Others				
Efficiency	typ		91%	
Protection		OVP, OCP,	OPP, OTP, FAN,ECP,Sense	
dimension		439.50mm (W) *132.80mm (F 483.00mm (W) *151.30mm (F cover and foot)		
Weight			26.4kg	
Working temperature			0°C-50°C	
Programming response time		2ms		
Remote Sense Compensation Voltage			20V	
Communication interface			Digital IO interface, optional GPIE nalog&RS232	3 /

^{*1} supports single phase 220V input, power cord connecting refer to manual.

IT7906-350-30

Input parameters					
	Wiring connection	3 phas	3 phase 3wire + ground(PE)		
	Line voltage	RMS	(200~220) ±10% <mark>*1</mark> (380~480) ±10%	V	
AC Input	Line current	RMS	< 13.8	Α	
	Apparent power		< 7	kVA	
	Frequency		45~65	Hz	
	Power factor	typ	0.98		
Output parameters					
		VLN	0~350	V	
	Output voltage	VLL (3phase)	0~606	V	
		VLL (reverse)	0∼700	V	

^{*2} Max. Output derated to 60% when input with ($200\sim220$) ±10% Vac 1 ph or 3 ph

^{*3} Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.

^{*4}The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10uF



IECH I			reem	nical Specif
		RMS (1phase)	30	А
		Crest Factor *2	6	
	Output current	Peak (1phase)	120	А
	Cusput current	RMS	10	А
		(3phase/multichannel/reverse) Peak		
		(3phase/multichannel/reverse)	40	A
		Per Phase/Per Channel	2k	VA
	Output power	Max. Power (reverse phase)	4k	VA
		Max. Power	6k	VA
		(1phase/3phase/multichannel)	OK .	V/-
	Voltage setting			
	Range	1phase/3phase/multichannel	0∼350	V
	rtailgo	reverse	0∼700	V
	Resolution		0.01	V
AC Output	Accuracy	16Hz∼150Hz	< 0.1%+0.1% F.S.	
	Temperature		< 100ppm/°C F.S.	
	coefficient DC Voltage Offset	tvo	0.02	Vdo
	Current Limit setting	typ	0.02	ı vac
		RMS	30	А
	Range	(1phase) RMS		
		(3phase/multichannel/reverse)	10	A
	Resolution		0.01	А
	Accuracy	16Hz∼150Hz	< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/°C F.S.	
	coefficient Frequency		< 200ррпі/ СТ.З.	
	Range		16~150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform	50/60Hz	up to 50	orde
	synthesizer	30/00/12	ир 10 30	orde
	Phase			
	Range		0∼360	٥
	Resolution		0.01	0
	Resolution Voltage setting		0.01	۰
	Voltage setting	1phase/multichannel	0.01 -499~499	
		1phase/multichannel reverse		Vde
	Voltage setting	,	-499~499	Vdo
	Voltage setting Range	,	-499~499 -998~998	Vdo
	Range Resolution Accuracy Temperature	,	-499~499 -998~998 0.01 <0.1%+0.1% F.S.	Vdo
	Range Resolution Accuracy Temperature coefficient	,	-499~499 -998~998 0.01	Vdo
	Range Resolution Accuracy Temperature	reverse	-499~499 -998~998 0.01 <0.1%+0.1% F.S. < 100ppm/°C F.S.	Vda Vda V
DC Output	Range Resolution Accuracy Temperature coefficient	reverse multichannel/reverse	-499~499 -998~998 0.01 <0.1%+0.1% F.S. < 100ppm/°C F.S.	Vda Vda V
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range	reverse	-499~499 -998~998 0.01 <0.1%+0.1% F.S. < 100ppm/°C F.S. -10~10 -30~30	Vdd Vdd Vd Vd Add Add
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution	reverse multichannel/reverse	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01	Vdd Vdd Vd Vd Add Add
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy	reverse multichannel/reverse	-499~499 -998~998 0.01 <0.1%+0.1% F.S. < 100ppm/°C F.S. -10~10 -30~30	Vdd Vdd V
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature	reverse multichannel/reverse	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01	Vdd Vdd Vd Vd Add Add
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature	reverse multichannel/reverse	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S.	Vdd Vdd V
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient	multichannel/reverse 1phase	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vdd Vdd Vdd Vdd Add Add A
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient Max. power Channel power	multichannel/reverse 1phase Per Channel	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vdd Vdd V
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient	reverse multichannel/reverse 1phase Per Channel Max. Power (reverse phase)	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vdd Vdd V
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient Max. power Max. power (reverse	reverse multichannel/reverse 1phase Per Channel Max. Power (reverse phase) Max. Power	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vda
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient Max. power Channel power Max. power (reverse phase) Total power	reverse multichannel/reverse 1phase Per Channel Max. Power (reverse phase)	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vda
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient Max. power Channel power Max. power (reverse phase) Total power Line regulation	reverse multichannel/reverse 1phase Per Channel Max. Power (reverse phase) Max. Power (1phase/multichannel)	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vdd Vdd Vdd Vdd Vdd Add Add Add Add WW W
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient Max. power Channel power Max. power (reverse phase) Total power	reverse multichannel/reverse 1phase Per Channel Max. Power (reverse phase) Max. Power	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vdd Vdd Vdd Vdd Vdd Add Add Add Add WW W
DC Output	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient Max. power Channel power Max. power Channel power Max. power Line regulation *3	reverse multichannel/reverse 1phase Per Channel Max. Power (reverse phase) Max. Power (1phase/multichannel)	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vdc Vdc V Adc Adc A W W
	Range Resolution Accuracy Temperature coefficient Current setting Range Resolution Accuracy Temperature coefficient Max. power Channel power Max. power (reverse phase) Total power Line regulation *3 Load regulation	reverse multichannel/reverse 1phase Per Channel Max. Power (reverse phase) Max. Power (1phase/multichannel) DC,16Hz~150Hz	-499~499 -998~998 0.01 <0.1%+0.1% F.S. <100ppm/°C F.S. -10~10 -30~30 0.01 <0.1%+0.2% F.S. <200ppm/°C F.S.	Vdc Vdc V



III III III III			lech	nical Specific
		3phase/multichannel	0∼1000	mΩ
	R Range	1phase	0∼333.333	mΩ
Programmable mpedance		reverse	0~2000	mΩ
ripedance		3phase/multichannel	0~1000	uH
	L Range	1phase	0~333.333	uH
		reverse	0~2000	uH
		3phase	0∼2	kW
	P Range	1phase	0~6	kW
		reverse	0~4	kW
		3phase	0∼2	kVar
	QL Range	1phase	0∼6	kVar
_		reverse	0~4	kVar
		3phase	0∼2	kVar
	QC Range	1phase	0~6	kVar
Anti-Islanding RLC	-	reverse	0~4	kVar
		3phase	1~1000	Ω
	R Range	1phase	0.333~333.333	Ω
		reverse	2~2000	Ω
		3phase	1~5000	mH
	L Range	1phase	0.333~1666.667	mH
_		reverse	2~10000	mH
		3phase	0.001~5	mF
	C Range	1phase	0.003~15	mF
		reverse	0.001~2.5	mF
oltage Slew Rate,		≥2 V/µs with full-scale programmed	d voltage step	
Typical Output Isolation		550Vac		
Measurement param	eter	000 vac		
	Resolution		0.01	V
Voltage RMS	Accuracy	DC,16Hz∼150Hz	<0.1%+0.1% F.S.	
	Temperature		< 100ppm/°C F.S.	
	coefficient Resolution		0.01	A
Current RMS	Accuracy	DC,16Hz~150Hz	<0.1% + 0.2% F.S.	^
Current Kivis	Temperature	DC,16H2 ~ 150H2		
	coefficient		< 200ppm/°C F.S.	
Peak current	Resolution		0.1	А
	Accuracy	16Hz∼150Hz	<0.4% + 0.6% F.S.	
Output power	Resolution		0.001	kW
Catput pono.	Accuracy	DC,16Hz~150Hz	<0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	orders
Regenerative				
Max. Regenerative power		6k		VA
ITHD		<		
Others		5%		
	typ		88%	
Efficiency		OVP OCP	OPP, OTP, FAN, ECP, Sense	
Protection) *700.00mm (D) (Instrument fram	no Dimosion
Dimension		437.00mm (W) 88.40mm (H 483.00mm (W) *106.90mm (H	H) *785.30mm (D) (include handl	e and cover)
Weight			18kg	
Working temperature			0°C-50°C	
Programming response time			2ms	
Remote Sense Compensatio			20V	
Voltage Communication		Built-in USB/CAN/LAN/Digita	I IO interface, optional GPIB / Analog	&RS232
interface		l g		



- *1 Max. Output derated to 60% when input with ($200\sim220$) ±10% Vac 1 ph or 3 ph
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *4 test condition is: Pure resistive loads, full power conditions.
- *5 The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10uF

IT7906-350-90

nput parameters	S			
	Wiring connection	3 phase 3wire + ground(PE)		
	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V
AC Input	Line current	RMS	< 24	А
	Apparent power		< 8.1	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
output paramete	ers			
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
		RMS (1phase)	90	А
		Crest Factor *2	6	
		Peak (1phase)	270	А
	Output current	RMS (3phase/multichannel/reverse)	30	А
		Peak (3phase/multichannel/reverse)	90	А
	Output power	Per Phase/Per Channel	2k	VA
		Max. Power (reverse phase)	4k	VA
		Max. Power (1phase/3phase/multichannel)	6k	VA
	Voltage setting	1		-
	-	1phase/3phase/multichannel	0 ~ 350	V
400	Range	reverse	0 ~ 700	V
AC Output	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
	Temperature coefficient		< 100ppm/C° F.S.	
	DC Voltage Offset	typ	0.02	Vde
	Current Limit setting			
		RMS (1phase)	90	А
	Range	RMS (3phase/multichannel/reverse)	30	А
	Resolution		0.01	А
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/C° F.S.	
	Frequency			1
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	



	waveform	50/60Hz	up to 50	orders
	synthesizer			
	Phase		0.000	
	Range		0 ~ 360	0
	Resolution		0.01	
	Voltage setting	1	400 400	\
	Range	1phase/multichannel	-499 ~ 499	Vdc
	Desclution	reverse	-998 ~ 998	Vdc V
	Resolution		0.01 < 0.1%+0.1% F.S.	V
	Accuracy		< U.1%+U.1% F.S.	
	Temperature coefficient		< 100ppm/C° F.S.	
	Current setting			
	Current setting	multichannel/reverse	-30 ~ 30	Adc
	Range	1phase	-90 ~ 90	Adc
DC Output	Resolution	трпазе	0.01	A
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature			
	coefficient		< 200ppm/C° F.S.	
	Max. power			
	Channel power	Per Channel	2k	W
	Max. power	Max. Power		
	(reverse phase)	(reverse phase)	4k	W
	Tatal a success	Max. Power	CI	147
	Total power	(1phase/multichannel)	6k	W
	Line regulation		< 0.05% F.S.	
	*3	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
	Load regulation			
Voltage stability	THD *4	16Hz ~ 150Hz	< 0.5%	
	Voltage ripple	RMS	< 0.4	V
	*5	typ	200	us
	Dynamic response			
		3phase/multichannel	0 ~ 1000	mΩ
	R Range	1phase	0 ~ 333.333	mΩ
Programmable		reverse	0 ~ 2000	mΩ
impedance		3phase/multichannel	0 ~ 1000	uH
	L Range	1phase	0 ~ 333.333	uH
		reverse	0 ~ 2000	uH
	D Damas	3phase	0 ~ 2	kW
	P Range	1phase	0 ~ 6 0 ~ 4	kW
		reverse		kW
	Ol Banga	3phase	0 ~ 2 0 ~ 6	kVar
	QL Range	1phase	0 ~ 6 0 ~ 4	kVar kVar
		reverse	0 ~ 2	
	OC Panga	3phase	0 ~ 2 0 ~ 6	kVar
Islanding RLC	QC Range	1phase	0 ~ 6 0 ~ 4	kVar kVar
		reverse	1 ~ 1000	_
	P Pange	3phase	0.333 ~ 333.333	Ω
	R Range	1phase	0.333 ~ 333.333 2 ~ 2000	Ω
		reverse	2 ~ 2000 1 ~ 5000	+
	I Panga	3phase		mH
	L Range	1phase reverse	0.333 ~ 1666.667 2 ~ 10000	mH mH
	1			



1		1	<u> </u>	pecifications
		1phase	0.003 ~ 15	mF
		reverse	0.001 ~ 2.5	mF
Voltage Slew Rate, Typical	*	2 V/μs with full-scale programmed	l voltage step	
Output Isolation		550Vac		
Measurement para	ameter			
	Resolution		0.01	V
\/altana DMC	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage RMS	Temperature coefficient		< 100ppm/C° F.S.	
	Resolution		0.01	А
	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current RMS	Temperature coefficient		< 200ppm/C° F.S.	
	Resolution		0.1	А
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
	Resolution		0.001	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	orders
Regenerative				
Max.				
Regenerative		6k		VA
power				
THD		< 5%		
Others				
Efficiency	typ		76%	
Protection			P, OTP, FAN, ECP, Sense	
Dimension		439.50mm (W)*132.80mm (H)*73483.00mm (W)*151.30mm (H)*160ot)	, , ,	cover and
Weight			42kg	
Working temperature			0C°-50C°	
Programming response time			2ms	
Remote Sense Compensation Voltage			20V	
Communication interface		Built-in USB/CAN/LAN/Digital IC) interface, optional GPIB / Analo	g&RS232

- ± 1 (200 ~ 220) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.



IT7909-350-90

put paramete				
	Wiring connection	3 phase 3v	vire + ground(PE)	
	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V
AC Input	Line current	RMS	< 33	А
	Apparent power		< 11.1	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
utput parame	ters			
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
		RMS (1phase)	90	А
		Crest Factor *2	6	
		Peak (1phase)	270	А
	Output current	RMS	20	^
		(3phase/multichannel/reverse)	30	А
		Peak	00	Λ.
		(3phase/multichannel/reverse)	90	A
		Per Phase/Per Channel	3k	VA
	Output power	Max. Power	6k	VA
		(reverse phase)	OK .	VA
		Max. Power	9k	VA
		(1phase/3phase/multichannel)	9K	V
	Voltage setting			
	Range	1phase/3phase/multichannel	0 ~ 350	V
	Kange	reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
AC Output	Temperature		< 100ppm/C° F.S.	
, to Gatpat	coefficient		100ppm/C 1.3.	
	DC Voltage	typ	0.02	Vd
	Offset		0.02	
	Current Limit settin	-		
		RMS	90	А
	Range	(1phase)		
	95	RMS	30	А
		(3phase/multichannel/reverse)		
	Resolution		0.01	А
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/C° F.S.	
	coefficient			
	Frequency		40 450	1
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	_
	waveform	50/60Hz	up to 50	orde
	synthesizer		1	
	Phase		0 000	0
	Range		0 ~ 360	
	Resolution		0.01	



			Technical S	_
	Range	1phase/multichannel	-499 ~ 499	Vdd
		reverse	-998 ~ 998	Vdd
	Resolution		0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
	Temperature coefficient		< 100ppm/C° F.S.	
	Current setting			
	Pango	multichannel/reverse	-30 ~ 30	Add
	Range	1phase	-90 ~ 90	Add
	Resolution		0.01	А
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/C° F.S.	
	Max. power			I
	Channel power	Per Channel	3k	W
	Max. power (reverse phase)	Max. Power (reverse phase)	6k	W
	Total power	Max. Power (1phase/multichannel)	9k	W
	Line regulation	(1911030/1110101101111101)	< 0.05% F.S.	\dashv
	*3	DC 1011 45011		
	Load regulation	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
Voltage	THD *4	16Hz ~ 150Hz	< 0.5%	
stability	Voltage ripple	RMS	< 0.4	V
	*5 Dynamic response	typ	200	us
	R Range	3phase/multichannel	0 ~ 1000	mΩ
		1phase	0 ~ 333.333	mΩ
Programmable		reverse	0 ~ 2000	mΩ
impedance		3phase/multichannel	0 ~ 1000	u⊦
·	L Range	1phase	0 ~ 333.333	ul
		reverse	0 ~ 2000	u⊦
		3phase	0 ~ 3	kV
	P Range	1phase	0 ~ 9	kV
		reverse	0 ~ 6	kV
		3phase	0 ~ 3	kVa
	QL Range	1phase	0 ~ 9	kVa
		reverse	0 ~ 6	kVa
		3phase	0 ~ 3	kVa
	QC Range	1phase	0 ~ 9	kVa
	42.101.90	reverse	0 ~ 6	kVa
Islanding RLC		3phase	1 ~ 1000	Ω
	R Range	1phase	0.333 ~ 333.333	Ω
	_	reverse	2 ~ 2000	Ω
		3phase	1 ~ 5000	mŀ
	L Range	1phase	0.333 ~ 1666.667	mŀ
	90	reverse	2 ~ 10000	mŀ
		3phase	0.001 ~ 5	mf
	C Range	1phase	0.001 3	ml
		reverse	0.003 13	mf
Voltage Slew Rate, Typical	>	2 V/μs with full-scale programmed ν		



Output Isolation		550Vac		pecifications
Measurement pa	rameter			
	Resolution		0.01	V
Voltage RMS	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage Rivis	Temperature		< 100ppm/C° F.S.	
	coefficient		100ppn//C F.3.	
	Resolution		0.01	А
Current RMS	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current Kivio	Temperature coefficient		< 200ppm/C° F.S.	
Peak current	Resolution		0.1	А
Peak Current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
Output power	Resolution		0.001	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic	Max.	50/60Hz	up to 50	orders
measurement	IVIUX.	307 00112	up to 30	Oracis
Regenerative				
Max.				
Regenerative		9k		VA
power THD		< 5%		
Others		< 5%		
Efficiency	tun	T T T T T T T T T T T T T T T T T T T	83%	
Protection	typ	OVP OCP OPP	, OTP, FAN, ECP, Sense	
Dimension		439.50mm (W)*132.80mm (H)*777.5		
Difficusion		483.00mm (W)*151.30mm (H)*841.6	* * * *	nd foot)
Weight			42kg	
Working		0.	C°-50C°	
temperature		U	C -00C	
Programming			2ms	
response time			20	
Remote Sense Compensation Voltage			20V	
Communication interface		Built-in USB/CAN/LAN/Digital IO i	interface, optional GPIB / Analog	&RS232

- ± 1 (200 ~ 220) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 uF
- $\star 6 \; Test \; conditions : Input \; 380 \; VLL/50 \; Hz, \; output \; three-phase, \; \; per \; phase \; 350 \; Vrms/50 \; Hz/Full \; power.$

IT7912-350-90

Input parameters				
AC Input	Wiring connection	3 phase 3w		
	Line voltage	RMS	(200~220) ±10% *1 (380~480) ±10%	V
	Line current	RMS	< 29	А
	Apparent power		< 14.8	kVA



	T -	<u> </u>		pecificatio
	Frequency		45 ~ 65	Hz
Ot :	Power factor	typ	0.98	
Output parame	eters	\(\alpha\)	0 050	\ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \
	Outrout valta va	VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
		RMS (1phase)	90	А
		Crest Factor *2	6	
		Peak (1phase)	270	А
	Output current	RMS (3phase/multichannel/reverse)	30	А
		Peak (3phase/multichannel/reverse)	90	А
		Per Phase/Per Channel	4k	VA
		Max. Power	01	
	Output power	(reverse phase)	8k	VA
		Max. Power (1phase/3phase/multichannel)	12k	VA
	Voltage setting	(251000) Options (Training)		
		1phase/3phase/multichannel	0 ~ 350	V
	Range	reverse	0 ~ 700	V
	Resolution	Teverse	0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	v
AC Output		10112 1 130112	<u> </u>	_
AC Output	Temperature coefficient		<100ppm/C° F.S.	
	DC Voltage Offset	typ	0.02	Vdc
	Current Limit setting	ļ		
	Range	RMS	90	А
		(1phase)	30	
		RMS	30	А
		(3phase/multichannel/reverse)		/\
	Resolution		0.01	А
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/C° F.S.	
	Frequency			
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	112
	waveform			orde
	synthesizer	50/60Hz	up to 50	s
	Phase	1		s
	Range		0 ~ 360	0
	Resolution		0.01	0
	Voltage setting		0.01	
	voitage setting	1phase/multichannel	-499 ~ 499	Vdc
	Range	·	-499 ~ 499 -998 ~ 998	Vac
	Docalution	reverse		
	Resolution		0.01	V
DC Output	Accuracy		< 0.1%+0.1% F.S.	
	Temperature coefficient		< 100ppm/C° F.S.	
	Current setting			
	Range	multichannel/reverse	-30 ~ 30	Adc



ITEON			1	pecifications
	Resolution		0.01	А
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/C° F.S.	
	coefficient		× 200ββΠ// C 1.3.	
	Max. power			
	Channel power	Per Channel	4k	W
	Max. power (reverse phase)	Max. Power (reverse phase)	8k	W
	Total power	Max. Power (1phase/multichannel)	12k	W
	Line regulation	(Iphase/multichannel)	< 0.05% F.S.	
	*3		\ 0.00% F.S.	
Voltage	Load regulation	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
stability	THD *4	16Hz ~ 150Hz	< 0.5%	
oca o mey	Voltage ripple	RMS	< 0.4	V
	*5 Dynamic response	typ	200	us
		3phase/multichannel	0 ~ 1000	mΩ
	R Range	1phase	0 ~ 333.333	mΩ
Programmable		reverse	0 ~ 2000	mΩ
impedance		3phase/multichannel	0 ~ 1000	uН
•	L Range	1phase	0 ~ 333.333	uН
		reverse	0 ~ 2000	uН
		3phase	0 ~ 4	kW
	P Range	1phase	0~12	kW
		reverse	0 ~ 8	kW
		3phase	0 ~ 4	kVar
	QL Range	1phase	0~12	kVar
		reverse	0~8	kVar
		3phase	0 ~ 4	kVar
	QC Range	1phase	0~12	kVar
	Q	reverse	0 ~ 8	kVar
Islanding RLC		3phase	1 ~ 1000	Ω
	R Range	1phase	0.333 ~ 333.333	Ω
	it italigo	reverse	2 ~ 2000	Ω
		3phase	1 ~ 5000	mH
	L Range	1phase	0.333 ~ 1666.667	mH
	90	reverse	2 ~ 10000	mH
		3phase	0.001 ~ 5	mF
	C Range	1phase	0.003 ~ 15	mF
	o Kange	reverse	0.001 ~ 2.5	mF
Voltage Slew Rate,	>	>2 V/µs with full-scale programmed v	l	
Typical				
Output Isolation		550Vac		
Measurement p	arameter			
	Resolution		0.01	V
V-140 D140	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage RMS	Temperature		× 100 · · · · /00 F 0	
	coefficient		< 100ppm/C° F.S.	
	Resolution		0.01	А
Course at DA40	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current RMS	Temperature		< 200 nm /C° F C	
	coefficient		< 200ppm/C° F.S.	



	Resolution		0.1	А
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
Outrot a success	Resolution		0.001	kW
Output power -	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	order s
Regenerative				
Max. Regenerative power		12k		VA
THD		< 5%		
Others				
Efficiency	typ		83%	
Protection		OVP, OCP, OPP,	, OTP, FAN, ECP, Sense	
Dimension		439.50mm (W)*132.80mm (H)*777. 483.00mm (W)*151.30mm (H)*841.	, , ,	and foot)
Weight			42kg	
Working temperature		0.0	C°-50C°	
Programming response time			2ms	
Remote Sense Compensation Voltage			20V	
Communicatio n interface		Built-in USB/CAN/LAN/Digital IO i	interface, optional GPIB / Analc	g&RS232

- ± 1 ($200 \sim 220$) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 u F
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.
- All the above parameters are subject to change without prior notice from ITECH.

IT7915-350-90

Input parameters				
	Wiring connection	3 phase 3wire + ground(PE)		
	Line voltage	RMS	(200~220) ±10% *1 (380~480) ±10%	V
AC Input	Line current	RMS	< 34	Α
	Apparent power		< 17.4	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
Output parameters				
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
A.C. Quitout		RMS (1phase)	90	Α
AC Output		Crest Factor *2	6	
	Output current	Peak (1phase)	270	А
		RMS (3phase/multichannel/reverse)	30	А



	<u>- • • • • • • • • • • • • • • • • • • •</u>	1	Technical Sp	ecification
		Peak (3phase/multichannel/reverse)	90	А
		Per Phase/Per Channel	 5k	VA
		Max. Power		
	Output power	(reverse phase)	10k	VA
		Max. Power	151.	\/A
		(1phase/3phase/multichannel)	15k	VA
	Voltage setting			
	Range	1phase/3phase/multichannel	0 ~ 350	V
		reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
	Temperature		< 100ppm/C° F.S.	
	coefficient			- \
	DC Voltage Offset	typ	0.02	Vdc
	Current Limit setting	RMS		
		(1phase)	90	Α
	Range	RMS		+
		(3phase/multichannel/reverse)	30	Α
	Resolution	(-)	0.01	Α
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	1
	Temperature			
	coefficient		< 200ppm/C° F.S.	
	Frequency	,		
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform	50/60Hz	up to 50	order
	synthesizer	307 00112	<u>ар то зо</u>	S
	Phase			
	Range		0 ~ 360	0
	Resolution		0.01	0
	Voltage setting	4.1	400 400	1 1/1
	Range	1phase/multichannel	-499 ~ 499	Vdc
	Resolution	reverse	-998 ~ 998 0.01	Vdc V
	Accuracy		< 0.1%+0.1% F.S.	- V
	Temperature			_
	coefficient		< 100ppm/C° F.S.	
	Current setting	1		
		multichannel/reverse	-30 ~ 30	Adc
2	Range	1phase	-90 ~ 90	Adc
C Output	Resolution		0.01	А
	Accuracy		< 0.1%+0.2% F.S.	1
	Temperature		< 200nnm/C° FC	
	coefficient		< 200ppm/C° F.S.	
	Max. power			
	Channel power	Per Channel	5k	W
	Max. power	Max. Power	10k	W
	(reverse phase)	(reverse phase)	TOK	
		Max. Power	451	14/
	Total power	(1phase/multichannel)	15k	W



TIEGII II II			Technical Sp	ecification
	*3 Load regulation	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
	THD *4	16Hz ~ 150Hz	< 0.5%	
	Voltage ripple	RMS	< 0.4	V
	*5 Dynamic response	typ	200	us
		3phase/multichannel	0 ~ 1000	mΩ
	R Range	1phase	0 ~ 333.333	mΩ
Programmable		reverse	0 ~ 2000	mΩ
impedance		3phase/multichannel	0 ~ 1000	uН
	L Range	1phase	0 ~ 333.333	uН
		reverse	0 ~ 2000	uH
		3phase	0 ~ 5	kW
	P Range	1phase	0~15	kW
		reverse	0~10	kW
		3phase	0~5	kVar
	QL Range	1phase	0~15	kVar
		reverse	0~10	kVar
		3phase	0 ~ 5	kVar
	QC Range	1phase	0 ~ 15	kVar
Jalandina DLC		reverse	0~10	kVar
Islanding RLC		3phase	1 ~ 1000	Ω
	R Range	1phase	0.333 ~ 333.333	Ω
		reverse	2 ~ 2000	Ω
		3phase	1 ~ 5000	mH
	L Range	1phase	0.333 ~ 1666.667	mH
		reverse	2 ~ 10000	mH
		3phase	0.001 ~ 5	mF
	C Range	1phase	0.003 ~ 15	mF
		reverse	0.001 ~ 2.5	mF
Voltage Slew Rate, Typical	≥2 V/ ₁	μs with full-scale programmed vo	Itage step	
Output Isolation		550Vac		
Measurement paran	neter			
	Resolution		0.01	V
Maliana DMC	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage RMS	Temperature coefficient		< 100ppm/C° F.S.	
	Resolution		0.01	А
	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current RMS	Temperature coefficient	,	< 200ppm/C° F.S.	
	Resolution		0.1	А
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	, ,
	Resolution	200112	0.001	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
harmonic measurement	Max.	50/60Hz	up to 50	order s
Regenerative			<u> </u>	
Max. Regenerative		 15k		VA
power				VA
THD		< 5%		
Others				
Efficiency	typ *6		91%	



	·
Protection	OVP, OCP, OPP, OTP, FAN, ECP, Sense
Dimension	439.50mm (W)*132.80mm (H)*777.50mm (D) (frame Dimension)
	483.00mm (W)*151.30mm (H)*841.60mm(D)(include handle, cover and
	foot)
Weight	42kg
Working	0C°-50C°
temperature	00 -300
Programming	2ms
response time	21115
Sense	
compensation	20V
Voltage	
Communication	Built-in USB/CAN/LAN/Digital IO interface, optional GPIB /
interface	Analog&RS232

- ± 1 ($200 \sim 220$) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- *5 The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.
- All the above parameters are subject to change without prior notice from ITECH.

IT7930-350-180

nput paramete	rs			
	Wiring connection	3 phase 3wire + ground(PE)		
	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V
AC Input	Line current	RMS	< 67	А
	Apparent power		< 34.8	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
Output parame	ters			
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
	Output current	RMS (1phase)	180	А
		Crest Factor *2	6	
		Peak (1phase)	540	А
		RMS (3phase/multichannel/reverse)	60	А
400		Peak (3phase/multichannel/reverse)	180	А
AC Output		Per Phase/Per Channel	10k	VA
	Output power	Max. Power (reverse phase)	20k	VA
		Max. Power (1phase/3phase/multichannel)	30k	VA
	Voltage setting			•
	Dange	1phase/3phase/multichannel	0 ~ 350	V
	Range	reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	



IIEGH		1 ,	iecnnicais	Specificatio
	Temperature coefficient		<100ppm/C° F.S.	
	DC Voltage Offset	typ	0.02	Vdc
	Current Limit setting			,
	Dongo	RMS (1phase)	180	А
	Range	RMS (3phase/multichannel/reverse)	60	А
	Resolution		0.01	А
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/C° F.S.	
	Frequency			
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform synthesizer	50/60Hz	up to 50	order
	Phase			
	Range		0 ~ 360	0
	Resolution		0.01	0
	Voltage setting			
	Danas	1phase/multichannel	-499 ~ 499	Vdc
	Range	reverse	-998 ~ 998	Vdc
	Resolution		0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
	Temperature coefficient		< 100ppm/C° F.S.	
	Current setting			
	Dange	multichannel/reverse	-60 ~ 60	Adc
DC Output	Range	1phase	-180 ~ 180	Adc
DC Output	Resolution		0.01	А
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/C° F.S.	
	Max. power			
	Channel power	Per Channel	10k	W
	Max. power (reverse phase)	Max. Power (reverse phase)	20k	W
	Total power	Max. Power (1phase/multichannel)	30k	W
	Line regulation		< 0.05% F.S.	
Voltago	*3 Load regulation	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
Voltage stability	THD *4	16Hz ~ 150Hz	< 0.5%	
Stability	Voltage ripple	RMS	< 0.5	V
	*5 Dynamic response	typ	200	us
		3phase/multichannel	0 ~ 500	mΩ
	R Range	1phase	0 ~ 166.667	mΩ
Programmable		reverse	0 ~ 1000	mΩ
impedance		3phase/multichannel	0 ~ 500	uН
	L Range	1phase	0 ~ 166.667	uН
		reverse	0 ~ 1000	uН
Islanding RLC	P Range	3phase	0 ~ 10	kW



(112011			lechnical	Specifications
		1phase	0 ~ 30	kW
_		reverse	0 ~ 20	kW
		3phase	0 ~ 10	kVar
	QL Range	1phase	0 ~ 30	kVar
		reverse	0 ~ 20	kVar
		3phase	0 ~ 10	kVar
	QC Range	1phase	0 ~ 30	kVar
		reverse	0~20	kVar
		3phase	0.5 ~ 500	Ω
	R Range	1phase	0.167 ~ 166.667	Ω
		reverse	1 ~ 1000	Ω
		3phase	0.5 ~ 2500	mH
	L Range	1phase	0.167 ~ 833.333	mH
	_	reverse	1 ~ 5000	mH
_		3phase	0.002 ~ 10	mF
	C Range	1phase	0.006 ~ 30	mF
	 	reverse	0.001 ~ 5	mF
Voltage Slew		.0.0.00	0.002	
Rate,	>	2 V/μs with full-scale programmed	voltage step	
Typical		,	J 1	
Output		FF0\/		
Isolation		550Vac		
Measurement par	rameter			
	Resolution		0.01	V
\/=\t==== D\/C	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage RMS	Temperature		1100 100 F.C	
	coefficient		< 100ppm/C° F.S.	
	Resolution		0.1	А
Comment DNAC	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current RMS	Temperature		1000 100 F.C	
	coefficient		< 200ppm/C° F.S.	
D 1	Resolution		0.1	А
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
0	Resolution		0.001	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic		50,0001		
measurement	Max.	50/60Hz	up to 50	orders
Regenerative				
Max.				
Regenerative		30k		VA
power				
THD		< 5%		
Others				
Efficiency	typ *6		91%	
Protection		OVP, OCP, OP	P, OTP, FAN, ECP, Sense	
Dimension		439.50mm (W)*266.00mm (H) 483.00mm (W)*347.2mm (H)* foot)	*777.50mm (D) (frame Dime	
Weight			42kg	
Working				
temperature			0C°-50C°	
Programming			2	
		i	2ms	



Remote Sense Compensation Voltage	20V
Communication interface	Built-in USB/CAN/LAN/Digital IO interface, optional GPIB / Analog&RS232

- ± 1 ($200 \sim 220$) $~\pm 10\%,$ the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.

IT7945-350-270

Input parameters				
	Wiring connection	3 phase 3wire	+ ground(PE)	
	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V
AC Input	Line current	RMS	< 100	Α
	Apparent power		< 52.2	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
Output parameters				
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
		RMS (1phase)	270	Α
		Crest Factor *2	6	
	Output current	Peak (1phase)	810	Α
		RMS	90	А
		(3phase/multichannel/reverse) Peak (3phase/multichannel/reverse)	270	Α
	Output power	Per Phase/Per Channel	15k	VA
		Max. Power (reverse phase)	30k	VA
AC Output		Max. Power (1phase/3phase/multichannel)	45k	VA
	Voltage setting			
	Range	1phase/3phase/multichannel	0 ~ 350	V
	Kange	reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
	Temperature coefficient		< 100ppm/C° F.S.	
	DC Voltage Offset	typ	0.02	Vdc
	Current Limit setting			
	Panga	RMS (1phase)	270	А
	Range	RMS (3phase/multichannel/reverse)	90	А
	Resolution	·	0.01	Α



ITECH			Technical Sp	ecificatio
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/C° F.S.	
	coefficient		×200μμπ/C 1.5.	
	Frequency			
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform	E0/0011-	+- 50	
	synthesizer	50/60Hz	up to 50	order
	Phase			
	Range		0 ~ 360	0
	Resolution		0.01	0
	Voltage setting		<u>.</u>	
DC Output		1phase/multichannel	-499 ~ 499	Vdc
	Range	reverse	-998 ~ 998	Vdc
	Resolution		0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
	Temperature			
	coefficient		< 100ppm/C° F.S.	
	Current setting			
	_	multichannel/reverse	-90 ~ 90	Adc
	Range	1phase	-270 ~ 270	Adc
	Resolution	I pridec	0.01	A
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature		· 0.1/0 · 0.2/0 · .5.	
	coefficient		< 200ppm/C° F.S.	
	Max. power			
	Channel power	Per Channel	15k	W
	Max. power	Max. Power	1510	***
	(reverse phase)	(reverse phase)	30k	W
	(reverse pridse)	Max. Power		
	Total power	(1phase/multichannel)	45k	W
Voltage stability	Line regulation	(iphase/matterialmer)	< 0.05% F.S.	
			< 0.03% 1.3.	
	*3	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
	Load regulation THD *4	16Hz ~ 150Hz	< 0.5%	
			< 0.5%	V
	Voltage ripple	RMS	< 0.6	V
	*5	typ	200	us
	Dynamic response		0 . 222 222	
Programmable impedance	D. Donore	3phase/multichannel	0 ~ 333.333	mΩ
	R Range	1phase	0 ~ 111.111	mΩ
		reverse	0 ~ 666.667	mΩ
	I Danier	3phase/multichannel	0 ~ 333.333	uH
	L Range	1phase	0 ~ 111.111	uH
Islanding RLC		reverse	0 ~ 666.667	uH
		3phase	0 ~ 15	kW
	P Range	1phase	0 ~ 45	kW
		reverse	0 ~ 30	kW
		3phase	0 ~ 15	kVaı
	QL Range	1phase	0 ~ 45	kVaı
		reverse	0 ~ 30	kVaı
		3phase	0 ~ 15	kVar
	QC Range	1phase	0 ~ 45	kVar
			0 ~ 30	



(ITEOTI			Technical Sp	becincation
		3phase	0.333 ~ 333.333	Ω
	R Range	1phase	0.111 ~ 111.111	Ω
		reverse	0.667 ~ 666.667	Ω
		3phase	0.333 ~ 1666.667	mH
	L Range	1phase	0.111 ~ 555.556	mH
		reverse	0.667 ~ 3333.333	mH
		3phase	0.003 ~ 15	mF
	C Range	1phase	0.009 ~ 45	mF
		reverse	0.002 ~ 7.5	mF
Voltage Slew Rate, Typical	≥2 \	V/μs with full-scale programmed vo	ltage step	
Output Isolation		550Vac		
Measurement parame	eter			
	Resolution		0.01	V
Valta aa DMC	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage RMS	Temperature coefficient		< 100ppm/C° F.S.	
	Resolution		0.1	Α
	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current RMS	Temperature coefficient	,	< 200ppm/C° F.S.	
	Resolution		0.1	Α
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
_	Resolution		0.001	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	orders
Regenerative				
Max. Regenerative		45k		VA
THD		< 5%		
Others				
Efficiency	typ *6		91%	
Protection		OVP, OCP, OPP, (OTP, FAN, ECP, Sense	
Dimension		550.0000mm (W) *907.6 (H) *	*841mm(D)(909mm inclu	de cover)
Weight		4	42kg	
Working		0.0	°-50C°	
temperature		00		
Programming response time		2ms		
Remote Sense Compensation Voltage		20V		
Communication interface		Built-in USB/CAN/LAN/Digi Analo	ital IO interface, optional GP g&RS232	IB /

- ± 1 ($200 \sim 220$) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- *5 The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power. All the above parameters are subject to change without prior notice from ITECH.



IT7960-350-360

Input parameters				
	Wiring	3 phasa 3wir	o + ground(DE)	
	connection	3 phase 3wire + ground(PE)		
AC Loss d	Line voltage	RMS	(200 ~ 220) ± 10% *1 (380 ~ 480) ±10%	V
AC Input	Line current	RMS	< 133	Α
	Apparent power		< 69.6	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
Output parameters				
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
		RMS (1phase)	360	Α
		Crest Factor *2	6	
	-	Peak (1phase)	1080	Α
	Output current	RMS	100	
		(3phase/multichannel/reverse)	120	Α
	<u> </u>	Peak		
		(3phase/multichannel/reverse)	360	Α
		Per Phase/Per Channel	20k	VA
	Output power	Max. Power		
		(reverse phase)	40k	VA
		Max. Power		
		(1phase/3phase/multichannel)	60k	VA
	Voltage setting			
		1phase/3phase/multichannel	0 ~ 350	V
	Range	reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
	Temperature	20112 2001.12		
AC Output	coefficient		<100ppm/C° F.S.	
	DC Voltage			
	Offset	typ	0.02	Vdc
	Current Limit settin	ng .		
		RMS		
	_	(1phase)	360	Α
	Range	RMS	100	
		(3phase/multichannel/reverse)	120	Α
	Resolution		0.1	Α
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature		.000 /00 50	
	coefficient		< 200ppm/C° F.S.	
	Frequency			
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform	F0/6011		
	synthesizer	50/60Hz	up to 50	order
	Phase		•	
	Range		0 ~ 360	0
	Resolution		0.01	0
DC Output	Voltage setting		1	



IIICII	 		Technical Sp	ecincati
	Range	1phase/multichannel	-499 ~ 499	Vdc
	Range	reverse	-998 ~ 998	Vdc
	Resolution		0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
	Temperature		< 100ppm/C° F.S.	
	coefficient		> 100hhiii/C L'2.	
	Current setting			
	Range	multichannel/reverse	-120 ~ 120	Add
	Kange	1phase	-360 ~ 360	Add
	Resolution		0.1	Α
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/C° F.S.	
	coefficient		< 200ppin/C F.3.	
	Max. power			
	Channel power	Per Channel	20k	W
	Max. power	Max. Power	40k	W
	(reverse phase)	(reverse phase)	40K	٧٧
	Total power	Max. Power	60k	W
	iotai powei	(1phase/multichannel)		٧٧
	Line regulation		< 0.05% F.S.	
	*3	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
	Load regulation			
Voltage stability	THD *4	16Hz ~ 150Hz	< 0.5%	
	Voltage ripple	RMS	< 0.7	V
	*5			
	Dynamic	typ	1000	us
	response			
		3phase/multichannel	0 ~ 250	mΩ
	R Range	1phase	0 ~ 83.333	mΩ
Programmable		reverse	0 ~ 500	mΩ.
impedance		3phase/multichannel	0 ~ 250	uH
	L Range	1phase	0 ~ 83.333	uH
_		reverse	0 ~ 500	uН
		3phase	0 ~ 20	kW
	P Range	1phase	0 ~ 60	kW
		reverse	0 ~ 40	kW
		3phase	0 ~ 20	kVa
	QL Range	1phase	0 ~ 60	kVa
		reverse	0 ~ 40	kVa
		3phase	0 ~ 20	kVa
	QC Range	1phase	0 ~ 60	kVa
Islanding RLC		reverse	0 ~ 40	kVa
isianding NEC		3phase	0.25 ~ 250	Ω
	R Range	1phase	0.083 ~ 83.333	Ω
		reverse	0.5 ~ 500	Ω
		3phase	0.25 ~ 1250	mH
	L Range	1phase	0.083 ~ 416.667	m⊦
		reverse	0.5 ~ 2500	mH
		3phase	0.004 ~ 20	mF
	C Range	1phase	0.012 ~ 60	mF
		reverse	0.002 ~ 10	mF
Voltage Slew Rate,			•	
Typical	≥2	V/μs with full-scale programmed v	oitage step	
Output Isolation	1	550Vac		



Measurement param	neter				
	Resolution			0.01	V
Voltage RMS	Accuracy	DC,16Hz ~ 1	150Hz	< 0.1%+0.1% F.S.	
Voilage Rivis	Temperature			< 100ppm/C° F.S.	
	coefficient			< 100ppiii/C	
	Resolution			0.1	А
Current RMS	Accuracy	DC,16Hz ~ 1	150Hz	< 0.1% + 0.2% F.S.	
Current Nivio	Temperature			< 200ppm/C° F.S.	
	coefficient			< 200ppin/C F.3.	
Peak current	Resolution			1	А
reak current	Accuracy	16Hz ~ 15	i0Hz	< 0.4% + 0.6% F.S.	
Output power	Resolution			0.001	kW
Output power	Accuracy	DC,16Hz ~ 1	150Hz	< 0.4% +0.4% F.S.	
Harmonic	Max.	50/60H	17	up to 50	orders
measurement	IVIAX.	30/001	12	up to 30	Orders
Regenerative					1
Max. Regenerative		f	60k		VA
power					• • • • • • • • • • • • • • • • • • • •
THD		<	< 5%		
Others					
Efficiency	typ *6		919		
Protection				P, FAN, ECP, Sense	
Dimension		600.0000mm (W)	*1475 (H) *841	.mm (D) (909mm inclu	de cover)
Weight			42k	g	
Working			0C°-5	UC ₀	
temperature			00 -0		
Programming			2m	ς	
response time			2111		
Remote Sense					
Compensation			20\	<i>I</i>	
Voltage					
Communication		Built-in USB/CAN/LAN/Digital IO interface, optional GPIB / Analog&RS232			
interface				, op.:	950.10202

- ± 1 (200 ~ 220) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- *5 The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.
- All the above parameters are subject to change without prior notice from ITECH.

IT7975-350-450

Input paramete	Input parameters				
Wiring connection		3 phase	e 3wire + ground(PE)		
	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V	
AC Input	Line current	RMS	< 167	А	
	Apparent power		< 87	kVA	
	Frequency		45 ~ 65	Hz	
	Power factor	typ	0.98		
Output parameters					



			Technical :	peemeati
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
		RMS (1phase)	450	А
		Crest Factor *2	6	
		Peak (1phase)	1350	А
	Output current	RMS	150	А
		(3phase/multichannel/reverse)	130	A
		Peak	450	А
		(3phase/multichannel/reverse)	430	
		Per Phase/Per Channel	25k	VA
		Max. Power	50k	VA
	Output power	(reverse phase)	JUK	٧٨
		Max. Power	75k	VA
		(1phase/3phase/multichannel)	75K	VA
	Voltage setting			
	Range	1phase/3phase/multichannel	0 ~ 350	V
	Kange	reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
AC Output	Temperature		< 100ppm/C° F.S.	
7.0 Output	coefficient		< 100ppm/C F.3.	
	DC Voltage	typ	0.02	Vd
	Offset	typ	0.02	Vui
	Current Limit settin	g		
		RMS	450	А
	Range	(1phase)	400	
		RMS	150	А
		(3phase/multichannel/reverse)		
	Resolution		0.1	А
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/C° F.S.	
	coefficient		200pp 0 1.0.	
	Frequency			
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform	50/60Hz	up to 50	orde
	synthesizer	33, 33.12		0.00
	Phase			
	Range		0 ~ 360	۰
	Resolution		0.01	0
	Voltage setting			
	Range	1phase/multichannel	-499 ~ 499	Vd
		reverse	-998 ~ 998	Vd
	Resolution		0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
DC Output	Temperature		< 100ppm/C° F.S.	
- 25.4566	coefficient			
	Current setting			
	Range	multichannel/reverse	-150 ~ 150	Ad
		1phase	-450 ~ 450	Ad
	Resolution		0.1	Α
	Accuracy		< 0.1%+0.2% F.S.	



(II LOII)			Technical S	pecifications
	Temperature coefficient		< 200ppm/C° F.S.	
	Max. power			
	Channel power	Per Channel	25k	W
	Max. power	Max. Power	50k	W
	(reverse phase)	(reverse phase)	JUK	VV
	Total power	Max. Power (1phase/multichannel)	75k	W
	Line regulation		< 0.05% F.S.	
	*3 Load regulation	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
Voltage	THD *4	16Hz ~ 150Hz	< 0.5%	
stability	Voltage ripple	RMS	< 0.8	V
	*5 Dynamic	typ	1000	us
	response	3phase/multichannel	0 ~ 200	mΩ
	P Dango	1phase	0 ~ 200	mΩ
Programmable	R Range	reverse	0 ~ 400	mΩ
impedance				_
impedance	I Dance	3phase/multichannel 1phase	0 ~ 200 0 ~ 66.667	uH uH
	L Range	'	0 ~ 60.007	
		reverse		uH
	D.D	3phase	0 ~ 25	kW
	P Range	1phase	0~75	kW
		reverse	0 ~ 50	kW
		3phase	0 ~ 25	kVar
	QL Range	1phase	0 ~ 75	kVar
		reverse	0 ~ 50	kVar
		3phase	0 ~ 25	kVar
	QC Range	1phase	0 ~ 75	kVar
Islanding RLC		reverse	0 ~ 50	kVar
		3phase	0.2 ~ 200	Ω
	R Range	1phase	0.067 ~ 66.667	Ω
		reverse	0.4 ~ 400	Ω
		3phase	0.2 ~ 1000	mH
	L Range	1phase	0.067 ~ 333.333	mH
		reverse	0.4 ~ 2000	mH
		3phase	0.005 ~ 25	mF
	C Range	1phase	0.015 ~ 75	mF
		reverse	0.003 ~ 12.5	mF
Voltage Slew Rate, Typical		≥2 V/µs with full-scale programmed v	voltage step	
Output		550Vac		
Isolation		550Vac		
Measurement p	arameter			
	Resolution		0.01	V
Voltage RMS	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
voitage Kivis	Temperature coefficient		< 100ppm/C° F.S.	
	Resolution		0.1	А
0 0140	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current RMS	Temperature coefficient		< 200ppm/C° F.S.	
Peak current	Resolution		1	А
	1	i		



	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
Output namer	Resolution		0.001	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic	Max.	50/60Hz	up to 50	orders
measurement	IVIAA.	30/00112	up to 50	Olders
Regenerative				
Max.				
Regenerative		75k		VA
power				
THD		< 5%		
Others				
Efficiency	typ *6	91%		
Protection		OVP, OCP, OPP, OTP, FAN, ECP, Sense		
Dimension		600.0000mm (W)*1475(H)*841mm(D)(909mm include cover)		
Weight		42kg		
Working		00	°-50C°	
temperature		00	500	
Programming		2ms		
response time		21110		
Remote Sense				
Compensation		20V		
Voltage				
Communication		Built-in USB/CAN/LAN/Digital IO interface, optional GPIB / Analog&RS232		
interface		Built-iii 035/04/1/LAIV/Digital 10 liiteriace, optional 9F16 / Alialog&R3232		

- ± 1 ($200 \sim 220$) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- *5 The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.
- All the above parameters are subject to change without prior notice from ITECH.

IT7990-350-540

Input paramete	rs				
_	Wiring connection	3 phase 3	3 phase 3wire + ground(PE)		
	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V	
AC Input	Line current	RMS	< 200	А	
	Apparent power		< 104.4	kVA	
	Frequency		45 ~ 65	Hz	
	Power factor	typ	0.98		
Output paramet	ters				
		VLN	0 ~ 350	V	
	Output voltage	VLL (3phase)	0 ~ 606	V	
		VLL (reverse)	0 ~ 700	V	
		RMS (1phase)	540	А	
AC Output		Crest Factor *2	6		
AC Output		Peak (1phase)	1620	А	
	Output current	RMS (3phase/multichannel/reverse)	180	А	
		Peak (3phase/multichannel/reverse)	540	A kVA Hz	



ITECH			Technical S	pecificatio
		Per Phase/Per Channel	30k	VA
		Max. Power	60k	VA
	Output power	(reverse phase)	OUK	VA
		Max. Power	90k	VA
		(1phase/3phase/multichannel)		VA
	Voltage setting			
	Range	1phase/3phase/multichannel	0 ~ 350	V
		reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
	Temperature		< 100ppm/C° F.S.	
	coefficient			
	DC Voltage Offset	typ	0.02	Vdd
	Current Limit setting			
		RMS	540	А
	Range	(1phase)		
		RMS	180	А
	B 1.0	(3phase/multichannel/reverse)	0.4	
	Resolution	4011 45011	0.1	А
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/C° F.S.	
	coefficient			
	Frequency		40, 450	<u> </u>
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform synthesizer	50/60Hz	up to 50	ordei
	Phase			
	Range		0 ~ 360	0
	Resolution		0.01	0
	Voltage setting		0.01	
	voltage setting	1phase/multichannel	-499 ~ 499	Vdo
	Range	reverse	-998 ~ 998	Vdo
	Resolution	1070130	0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
	Temperature			
	coefficient		< 100ppm/C° F.S.	
	Current setting	L		I
	_	multichannel/reverse	-180 ~ 180	Add
DC 0	Range	1phase	-540 ~ 540	Add
DC Output	Resolution	·	0.1	А
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/C° F.S.	
	coefficient		< 200ppm/C F.S.	
	Max. power			
	Channel power	Per Channel	30k	W
	Max. power	Max. Power	60k	W
	(reverse phase)	(reverse phase)	UUK	VV
	Total power	Max. Power	90k	W
	-	(1phase/multichannel)		٧٧
	Line regulation		< 0.05% F.S.	
Voltage	*3	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
stability	Load regulation			
	THD *4	16Hz ~ 150Hz	< 0.5%	





Dimension	600	0.0000mm (W) *1475 (H) *841mm (D) (909mm include cover)
Weight		42kg
Working temperature		0C°-50C°
Programming response time		2ms
Remote Sense Compensation Voltage		20V
Communication interface	Buil	lt-in USB/CAN/LAN/Digital IO interface, optional GPIB / Analog&RS232

- ± 1 (200 ~ 220) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- $\star 2$ Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 u F
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.
- All the above parameters are subject to change without prior notice from ITECH.

IT79105-350-630

Input paramete	rs			
	Wiring connection	3 phase 3v	vire + ground(PE)	
	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V
AC Input	Line current	RMS	< 233	А
	Apparent power		< 121.8	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
Output paramet	ters			
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
	Output current	RMS (1phase)	630	А
		Crest Factor *2	6	
		Peak (1phase)	1890	А
		RMS (3phase/multichannel/reverse)	210	А
		Peak (3phase/multichannel/reverse)	630	А
	Output power	Per Phase/Per Channel	35k	VA
AC Output		Max. Power (reverse phase)	70k	VA
		Max. Power (1phase/3phase/multichannel)	105k	VA
	Voltage setting		•	W.
	Dange	1phase/3phase/multichannel	0 ~ 350	V
	Range	reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
	Temperature coefficient		< 100ppm/C° F.S.	
	DC Voltage Offset	typ	0.02	Vdc



	Commont Limit and in a		Technical Sp		
	Current Limit setting			1	
		RMS	630	А	
	Range	(1phase)			
	95	RMS (3phase/multichannel/reverse)	210	А	
	Resolution		0.1	А	
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.		
	Temperature	20112 200112			
	coefficient		< 200ppm/C° F.S.		
	Frequency				
	Range		16 ~ 150	Hz	
	Resolution		0.01	Hz	
	Accuracy		0.01%		
				order	
	waveform synthesizer	50/60Hz	up to 50	S	
	Phase				
	Range		0 ~ 360	0	
	Resolution		0.01	0	
	Voltage setting	,		ı	
	Range	1phase/multichannel	-499 ~ 499	Vdc	
	Runge	reverse	-998 ~ 998	Vdc	
	Resolution		0.01	V	
	Accuracy		< 0.1%+0.1% F.S.		
	Temperature		< 100ppm/C° F.S.		
	coefficient		· 100ppiii/C F.3.		
	Current setting				
	Range	multichannel/reverse	-210 ~ 210	Adc	
DC Output		1phase	-630 ~ 630	Adc	
DC Output	Resolution		0.1	А	
	Accuracy		< 0.1%+0.2% F.S.		
	Temperature		< 200ppm/C° F.S.		
	coefficient		· 200μμπ/C Γ.3.		
	Max. power				
	Channel power	Per Channel	35k	W	
	Max. power	Max. Power	701,	\ \ /	
	(reverse phase)	(reverse phase)	70k	W	
		Max. Power	1051.	147	
	Total power	(1phase/multichannel)	105k	W	
	Line regulation		< 0.05% F.S.		
	*3	DC 16Hz c. 150Hz	< 0.0E% + 0.0E% E.C		
Voltage	Load regulation	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.		
Voltage stability	THD *4	16Hz ~ 150Hz	< 0.5%		
Stability	Voltage ripple	RMS	< 1	V	
	*5	typ	1000	us	
	Dynamic response				
		3phase/multichannel	0 ~ 142.857	mΩ	
	R Range	1phase	0 ~ 47.619	mΩ	
Programmable		reverse	0 ~ 285.714	mΩ	
impedance		3phase/multichannel	0 ~ 142.857	uH	
	L Range	1phase	0 ~ 47.619	uH	
		reverse	0 ~ 285.714	uH	
		3phase	0 ~ 35	kW	
Islanding RLC	P Range	1phase	0 ~ 105	kW	
	1	reverse	0 ~ 70	kW	



(ITEOTI			iecnnicai Sį	pecifications
		3phase	0 ~ 35	kVar
	QL Range	1phase	0 ~ 105	kVar
		reverse	0 ~ 70	kVar
		3phase	0 ~ 35	kVar
	QC Range	1phase	0 ~ 105	kVar
		reverse	0~70	kVar
		3phase	0.143 ~ 142.857	Ω
	R Range	1phase	0.048 ~ 47.619	Ω
		reverse	0.286 ~ 285.714	Ω
<u> </u>		3phase	0.143 ~ 714.286	mH
	L Range	1phase	0.048 ~ 238.095	mH
	90	reverse	0.286 ~ 1428.571	mH
-		3phase	0.007 ~ 35	mF
	C Range	1phase	0.021 ~ 105	mF
	O Range	reverse	0.004 ~ 17.5	mF
Voltage Slew Rate, Typical	≥;	2 V/μs with full-scale programmed ν		
Output Isolation		550Vac		
Measurement par	ameter			
	Resolution		0.01	V
Value DMC	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage RMS –	Temperature coefficient		< 100ppm/C° F.S.	
	Resolution		0.1	А
	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current RMS -	Temperature coefficient		< 200ppm/C° F.S.	
	Resolution		1	Α
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
	Resolution		0.1	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	1111
Harmonic measurement	Max.	50/60Hz	up to 50	order s
Regenerative				3
Max. Regenerative		105k		VA
power				
THD		< 5%		
Others				
Efficiency	typ *6		91%	
Protection			P, OTP, FAN, ECP, Sense	
Dimension		600.0000mm (W) *1475 (H)) *841mm (D) (909mm inclu	ide cover)
Weight			42kg	
Working			0C°-50C°	
temperature			00 -000	
Programming			2ms	
response time			۵۱۱۱۵	
Remote Sense				
Compensation			20V	
Voltage				
Communicatio		Built-in USB/CAN/LAN/Digita	Il IO interface, optional	GPIB /
n interface		Analog&RS232		



- ± 1 ($200 \sim 220$) $~\pm 10\%,$ the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.
- All the above parameters are subject to change without prior notice from ITECH.

IT79120-350-720

nput paramete						
	Wiring connection	3 phase	3wire + ground(PE)			
	Line voltage	RMS	(200~220) ±10% *1 (380~480) ±10%	V		
AC Input	Line current	RMS	< 266	А		
	Apparent power		< 139.1	kVA		
	Frequency		45 ~ 65	Hz		
	Power factor	typ	0.98			
Output parame	ters	**	-			
		VLN	0 ~ 350	V		
	Output voltage	VLL (3phase)	0 ~ 606	V		
		VLL (reverse)	0 ~ 700	V		
		RMS (1phase)	720	А		
		Crest Factor *2	6			
		Peak (1phase)	2160	А		
	Output current	RMS	240			
		(3phase/multichannel/reverse)		Α		
		Peak (3phase/multichannel/reverse)	720	А		
		Per Phase/Per Channel	40k	VA		
		Max. Power				
	Output power	(reverse phase)	80k	VA		
		Max. Power (1phase/3phase/multichannel)	120k	VA		
AC Output	Voltage setting					
·		1phase/3phase/multichannel	0 ~ 350	V		
	Range	reverse	0 ~ 700	V		
	Resolution		0.01	V		
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.			
	Temperature coefficient		< 100ppm/C° F.S.			
	DC Voltage Offset	typ	0.02	Vdc		
	Current Limit sett	ing	-			
		RMS (1phase)	720	А		
	Range	RMS (3phase/multichannel/reverse)	240	А		
	Resolution	(opilase/maidenamien/everse)	0.1	А		
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.			



IILUII		ı	lechnical	Specificatio
	Temperature coefficient		< 200ppm/C° F.S.	
	Frequency			<u> </u>
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform	50/0011		
	synthesizer	50/60Hz	up to 50	order
	Phase	·		•
	Range		0 ~ 360	0
	Resolution		0.01	0
	Voltage setting	·		
	Dange	1phase/multichannel	-499 ~ 499	Vdc
	Range	reverse	-998 ~ 998	Vdc
	Resolution		0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
	Temperature		< 100mm / C° F C	
	coefficient		< 100ppm/C° F.S.	
	Current setting	·		
	Dange	multichannel/reverse	-240 ~ 240	Adc
DC Output	Range	1phase	-720 ~ 720	Adc
DC Output	Resolution		0.1	А
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature		.000 (00 50	
	coefficient		< 200ppm/C° F.S.	
	Max. power			W.
	Channel power	Per Channel	40k	W
	Max. power	Max. Power	001	147
	(reverse phase)	(reverse phase)	80k	W
	Total power	Max. Power	120k	W
		(1phase/multichannel)	.0.05% 5.0	
	Line regulation		< 0.05% F.S.	
	*3 Load regulation	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
Voltage	THD *4	16Hz ~ 150Hz	< 0.5%	
stability		10H2 ~ 15UH2 RMS	< 1.1	V
Stability	Voltage ripple *5	CIVIN	× 1.1	V
	Dynamic	typ	1000	us
	response	3phase/multichannel	0 ~ 125	mΩ
	R Range	1phase	0 ~ 41.667	mΩ
Programmable	Kitange	reverse	0 ~ 250	mΩ
impedance		3phase/multichannel	0 ~ 125	uH
impedance	L Range	1phase	0 ~ 41.667	uH
	Likange	reverse	0 ~ 250	uH
		3phase	0 ~ 40	kW
	P Range	1phase	0 ~ 120	kW
	i Kange	reverse	0~120	kW
		3phase	0 ~ 40	kVar
	OL Panca	'	0 ~ 40 0 ~ 120	kVar
Islanding RLC	QL Range	1phase		
		reverse	0 ~ 80	kVar
	00.0	3phase	0 ~ 40	kVar
	QC Range	1phase	0 ~ 120	kVar
		reverse	0 ~ 80	kVar
	R Range	3phase	0.125 ~ 125	Ω



IILOII			Technical	Specifications
		1phase	0.042 ~ 41.667	Ω
		reverse	0.25 ~ 250	Ω
		3phase	0.125 ~ 625	mH
	L Range	1phase	0.042 ~ 208.333	mH
		reverse	0.25 ~ 1250	mH
		3phase	0.008 ~ 40	mF
	C Range	1phase	0.024 ~ 120	mF
		reverse	0.004 ~ 20	mF
Voltage Slew Rate, Typical		≥2 V/µs with full-scale programm	ed voltage step	
Output Isolation		550Vac		
Measurement pa	arameter			
	Resolution		0.01	V
\/oltaga DMC	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage RMS	Temperature coefficient		< 100ppm/C° F.S.	
	Resolution		0.1	А
C DMC	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current RMS	Temperature coefficient		< 200ppm/C° F.S.	
	Resolution		1	А
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
	Resolution		0.1	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	orders
Regenerative				<u> </u>
Max. Regenerative power		120k		VA
THD		< 5%		
Others				
Efficiency	typ *6		91%	
Protection			P, OTP, FAN, ECP, Sense	
Dimension		600.0000mm (W) *1919 (H)) *841mm (D) (909mm includ	e cover)
Weight			42kg	
Working			0C°-50C°	
temperature				
Programming			2ms	
response time				
Remote Sense Compensation Voltage			20V	
Communication interface		Built-in USB/CAN/LAN/Digital IC	D interface, optional GPIB / Analog	g&RS232

- ± 1 ($200 \sim 220$) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 uF



 $\star 6$ Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power. All the above parameters are subject to change without prior notice from ITECH.

IT79135-350-810

Input paramete	ers			
mpat paramote	Wiring			
	connection	3 phase 3wire + ground(PE)		
	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V
AC Input	Line current	RMS	< 299	А
	Apparent power		< 156.5	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
Output parame	ters			
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
		VLL (reverse)	0 ~ 700	V
		RMS (1phase)	810	А
		Crest Factor *2	6	
		Peak (1phase)	2430	А
	Output current	RMS		
		(3phase/multichannel/reverse)	270	A
		Peak		
		(3phase/multichannel/reverse)	810	A
		Per Phase/Per Channel	45k	VA
	Output power	Max. Power	201	
		(reverse phase)	90k	VA
		Max. Power		
		(1phase/3phase/multichannel)	135k	VA
	Voltage setting	, , , , , , , , , , , , , , , , , , , ,		ı
	Range	1phase/3phase/multichannel	0 ~ 350	V
		reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
AC Output	Temperature coefficient		< 100ppm/C° F.S.	
	DC Voltage Offset	typ	0.02	Vdd
	Current Limit settin	g		I
		RMS	010	
	Da	(1phase)	810	A
	Range	RMS	270	А
		(3phase/multichannel/reverse)		
	Resolution		0.1	А
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/C° F.S.	
	Frequency			
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform	50/60Hz	up to 50	orde
	synthesizer	30/ 001 12	αρ το J0	orde
	Phase			
	Range		0 ~ 360	0



ITEOH				pecification
	Resolution		0.01	0
	Voltage setting		T	
	Range	1phase/multichannel	-499 ~ 499	Vdc
	_	reverse	-998 ~ 998	Vdc
	Resolution		0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
	Temperature		< 100ppm/C° F.S.	
	coefficient		100000000000000000000000000000000000000	
	Current setting			1
	Range	multichannel/reverse	-270 ~ 270	Adc
DC Output	_	1phase	-810 ~ 810	Adc
20 Catpat	Resolution		0.1	А
	Accuracy		< 0.1%+0.2% F.S.	
	Temperature		< 200ppm/C° F.S.	
	coefficient		1200ββπη σ 1.0.	
	Max. power			
	Channel power	Per Channel	45k	W
	Max. power	Max. Power	90k	W
	(reverse phase)	(reverse phase)	JOK	V V
	Total power	Max. Power	135k	W
	-	(1phase/multichannel)		٧٧
	Line regulation		< 0.05% F.S.	
	*3	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
	Load regulation	DC,10112 100112	(0.03/// 0.03/// 1.3.	
Voltage	THD *4	16Hz ~ 150Hz	< 0.5%	
stability	Voltage ripple	RMS	< 1.2	V
	*5			
	Dynamic	typ	1000	us
	response			
		3phase/multichannel	0 ~ 111.111	mΩ
	R Range	1phase	0 ~ 37.037	mΩ
Programmable		reverse	0 ~ 222.222	mΩ
impedance		3phase/multichannel	0 ~ 111.111	uH
	L Range	1phase	0 ~ 37.037	uH
		reverse	0 ~ 222.222	uH
		3phase	0 ~ 45	kW
	P Range	1phase	0 ~ 135	kW
		reverse	0 ~ 90	kW
		3phase	0 ~ 45	kVar
	QL Range	1phase	0 ~ 135	kVar
		reverse	0 ~ 90	kVar
		3phase	0 ~ 45	kVar
	QC Range	1phase	0 ~ 135	kVar
Islanding RLC		reverse	0 ~ 90	kVar
isiananig NEO		3phase	0.111 ~ 111.111	Ω
	R Range	1phase	0.037 ~ 37.037	Ω
		reverse	0.222 ~ 222.222	Ω
		3phase	0.111 ~ 555.556	mH
	L Range	1phase	0.037 ~ 185.185	mH
		reverse	0.222 ~ 1111.111	mH
		3phase	0.009 ~ 45	mF
	C Range	1phase	0.027 ~ 135	mF
		reverse	0.005 ~ 22.5	mF



Voltage Slew		>0\// - 2\ -(-	<u> </u>	Decinications
Rate, Typical		≥2 V/µs with full-scale programme	a voitage step	
Output Isolation		550Vac		
Measurement pa	arameter			
	Resolution		0.01	V
Maltara DNAC	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage RMS	Temperature		< 100ppm/C° F.S.	
	coefficient		100ppii//C 1.3.	
	Resolution		0.1	А
Current RMS	Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
Current Kivis	Temperature coefficient		< 200ppm/C° F.S.	
Dealessurent	Resolution		1	А
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
0	Resolution		0.1	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	orders
Regenerative				
Max.				
Regenerative		135k		VA
power				
THD		< 5%		
Others		<u></u>		
Efficiency	typ *6		91%	
Protection			P, OTP, FAN, ECP, Sense	
Dimension		600.0000mm (W) *1919 (H)		cover)
Weight			42kg	
Working			0C°-50C°	
temperature				
Programming			2ms	
response time				
Remote Sense Compensation			20V	
Voltage			2UV	
Communication				
interface		Built-in USB/CAN/LAN/Digital IC) interface, optional GPIB / Analog	&RS232
IIICHACE				

- ± 1 ($200 \sim 220$) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- *5 The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.
- All the above parameters are subject to change without prior notice from ITECH.

IT79150-350-900

Input paramete	rs			
	Wiring connection	3 phase 3	wire + ground(PE)	
AC Input	Line voltage	RMS	(200 ~ 220) ±10% *1 (380 ~ 480) ±10%	V
	Line current	RMS	< 333	А



(ITEOH			Technical	specification	
	Apparent power		< 173.9	kVA	
	Frequency		45 ~ 65	Hz	
	Power factor	typ	0.98		
Output parame	eters	,			
		VLN	0 ~ 350	V	
	Output voltage	VLL (3phase)	0 ~ 606	V	
		VLL (reverse)	0 ~ 700	V	
		RMS (1phase)	900	А	
		Crest Factor *2	6		
		Peak (1phase)	2700	А	
	Output current	RMS	300	А	
		(3phase/multichannel/reverse)			
		Peak (3phase/multichannel/reverse)	900	А	
		Per Phase/Per Channel	50k	VA	
		Max. Power	JUK	VA	
	Output power	(reverse phase)	100k	VA	
	Output power	Max. Power			
		(1phase/3phase/multichannel)	150k	VA	
	Voltage setting	(
		1phase/3phase/multichannel	0 ~ 350	V	
	Range	reverse	0 ~ 700	V	
	Resolution		0.01	V	
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.		
AC Output	Temperature				
	coefficient		< 100ppm/C° F.S.		
	DC Voltage Offset	typ	0.02	Vdc	
	Current Limit setting				
	Range	RMS	900	А	
		(1phase)			
		RMS	300	А	
		(3phase/multichannel/reverse)	300	/ \	
	Resolution		0.1	А	
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.		
	Temperature		< 200ppm/C° F.S.		
	coefficient				
	Frequency	1	40. 450		
	Range		16 ~ 150	Hz	
	Resolution		0.01	Hz	
	Accuracy	<u> </u>	0.01%		
	waveform synthesizer	50/60Hz	up to 50	orders	
	Phase				
	Range		0 ~ 360	0	
	Resolution		0.01	0	
	Voltage setting		0.01		
	voitage setting	1phase/multichannel	-499 ~ 499	Vdc	
	Range	reverse	-998 ~ 998	Vdc	
	Resolution	10,0130	0.01	Vuc	
DC Output	Accuracy	†	< 0.1%+0.1% F.S.	•	
20 Oatput	Temperature				
	coefficient		<100ppm/C° F.S.		
	Current setting	1			
	Range	multichannel/reverse	-300 ~ 300	Adc	



Resolution Accuracy	1phase	-900 ~ 900 0.1	Adc A
		0.1	ı A
Accuracy		. 0 40/ 0 00/ = 0	+
		< 0.1%+0.2% F.S.	
Temperature		< 200ppm/C° F.S.	
coefficient			
Max. power			
Channel power	Per Channel	50k	W
		100k	W
phase)			
Total power		150k	W
	(1phase/multichannel)		
		< 0.05% F.S.	
	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
			_
	RMS	< 1.3	V
	tvp	1000	us
Dynamic response			
L	· ·		mΩ
R Range	1phase		mΩ
	reverse		mΩ
L Range	3phase/multichannel		uH
	1phase	0 ~ 33.333	uН
	reverse	0 ~ 200	uН
	3phase	0 ~ 50	kW
P Range	1phase	0 ~ 150	kW
	reverse	0 ~ 100	kW
	3phase	0 ~ 50	kVai
QL Range	1phase	0 ~ 150	kVai
	reverse	0 ~ 100	kVaı
	3phase	0 ~ 50	kVai
OC Range	·		kVaı
	·		kVai
			Ω
R Range			Ω
it italigo			Ω
			mH
I Range	·		mH
L Range	·		mH
			mF
C Pango	·		mF
Citalige	· ·		
	10,0126	0.000 20	mF
>21	//us with full-scale programmed w	oltage sten	
>	7/μ3 With run-scale programmed vo	onage step	
			+
	550Vac		
rameter			
		0.01	V
	DC 16Hz ~ 150Hz		+ -
	20,20112 100112		+
-		< 100ppm/C° F.S.	
Resolution	-	0.1	А
Accuracy	DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	
	Max. power (reverse phase) Total power Line regulation *3 Load regulation THD *4 Voltage ripple *5 Dynamic response R Range L Range P Range QL Range QC Range R Range C Range	Max. power (reverse phase)Max. Power (reverse phase)Total powerMax. Power (1phase/multichannel)Line regulation*3Load regulationDC,16Hz ~ 150HzTHD *416Hz ~ 150HzVoltage rippleRMS*5typDynamic response3phase/multichannelR Range1phase reverseL Range1phase reverseP Range1phase reverseQL Range1phase reverseQC Range1phase reverseQC Range1phase reverseR Range1phase reverseL Range1phase reverseL Range1phase reverseL Range1phase 	Max. power (reverse phase) Max. Power (reverse phase) 100k Total power phase) Max. Power (1phase/multichannel) 150k Line regulation +3 Load regulation C 0.05% + 0.00% +



	Temperature coefficient		< 200ppm/C° F.S.	
Peak current	Resolution		1	Α
reak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	
Output power	Resolution		0.1	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic measurement	Max.	50/60Hz	up to 50	orders
Regenerative				
Max. Regenerative power		150k		VA
THD		< 5%		
Others				
Efficiency	typ *6		91%	
Protection		OVP, OCP, C	OPP, OTP, FAN, ECP, Sense	
Dimension		600.0000mm (W) *1919 (H) *841mm (D) (909mm includ	de cover)
Weight			42kg	
Working temperature			0C°-50C°	
Programming response time			2ms	
Remote Sense Compensation Voltage			20V	
Communication interface		Built-in USB/CAN/LAN/Digital	IO interface, optional GPIB / Analo	g&RS232

- ± 1 ($200 \sim 220$) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.

All the above parameters are subject to change without prior notice from ITECH.

IT79165-350-990

Input paramete	rs			
	Wiring connection	3 phase 3	Bwire + ground(PE)	
	Line voltage	RMS	(200~220) ±10% *1 (380~480) ±10%	V
AC Input	Line current	RMS	< 366	А
	Apparent power		< 191.3	kVA
	Frequency		45 ~ 65	Hz
	Power factor	typ	0.98	
Output paramet	ers			
		VLN	0 ~ 350	V
	Output voltage	VLL (3phase)	0 ~ 606	V
AC Outout		VLL (reverse)	0 ~ 700	V
AC Output		RMS (1phase)	990	А
	Output current	Crest Factor *2	6	
		Peak (1phase)	2970	А



IIECH			Technical S	pecineatio
		RMS	330	А
		(3phase/multichannel/reverse)		
		Peak (3phase/multichannel/reverse)	990	А
		Per Phase/Per Channel	55k	VA
	Output power	Max. Power (reverse phase)	110k	VA
		Max. Power (1phase/3phase/multichannel)	165k	VA
	Voltage setting	(Iphase/ophase/matterialmer)		
		1phase/3phase/multichannel	0 ~ 350	V
	Range	reverse	0 ~ 700	V
	Resolution		0.01	V
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
	Temperature			
	coefficient		< 100ppm/C° F.S.	
	DC Voltage Offset	typ	0.02	Vdc
	Current Limit setting			
		RMS (1phase)	990	А
	Range	RMS	330	А
		(3phase/multichannel/reverse)		
	Resolution		0.1	А
	Accuracy	16Hz ~ 150Hz	< 0.1%+0.2% F.S.	
	Temperature coefficient		< 200ppm/C° F.S.	
	Frequency			
	Range		16 ~ 150	Hz
	Resolution		0.01	Hz
	Accuracy		0.01%	
	waveform synthesizer	50/60Hz	up to 50	orde
	Phase			
	Range		0 ~ 360	0
	Resolution		0.01	0
	Voltage setting			•
	Dongo	1phase/multichannel	-499 ~ 499	Vdd
	Range	reverse	-998 ~ 998	Vdc
	Resolution		0.01	V
	Accuracy		< 0.1%+0.1% F.S.	
	Temperature		< 100ppm/C° F.S.	
	coefficient		< 100ppm/C° F.S.	
	- I			
	coefficient	multichannel/reverse	-330 ~ 330	
DC Output	coefficient Current setting Range	multichannel/reverse 1phase	-330 ~ 330 -990 ~ 990	Adc
DC Output	coefficient Current setting Range Resolution		-330 ~ 330 -990 ~ 990 0.1	
DC Output	coefficient Current setting Range Resolution Accuracy		-330 ~ 330 -990 ~ 990	Adc
DC Output	coefficient Current setting Range Resolution Accuracy Temperature		-330 ~ 330 -990 ~ 990 0.1	Add
DC Output	coefficient Current setting Range Resolution Accuracy Temperature coefficient		-330 ~ 330 -990 ~ 990 0.1 < 0.1%+0.2% F.S.	Add
DC Output	Range Resolution Accuracy Temperature coefficient Max. power	1phase	-330 ~ 330 -990 ~ 990 0.1 < 0.1%+0.2% F.S. < 200ppm/C° F.S.	Adc
DC Output	coefficient Current setting Range Resolution Accuracy Temperature coefficient		-330 ~ 330 -990 ~ 990 0.1 < 0.1%+0.2% F.S. < 200ppm/C° F.S.	Adc A
DC Output	coefficient Current setting Range Resolution Accuracy Temperature coefficient Max. power Channel power	1phase Per Channel	-330 ~ 330 -990 ~ 990 0.1 < 0.1%+0.2% F.S. < 200ppm/C° F.S.	



IIIEUII III				Specification
	Line regulation		< 0.05% F.S.	
	*3	DC,16Hz ~ 150Hz	< 0.05% + 0.05% F.S.	
Voltage	Load regulation THD *4	16Hz ~ 150Hz	4 O FW	
stability	Voltage ripple	16HZ ~ 15UHZ RMS	< 0.5% < 1.4	V
	*5	RIVIS	< 1.4	V
	Dynamic response	typ	1000	us
		3phase/multichannel	0 ~ 90.909	mΩ
	R Range	1phase	0 ~ 30.303	mΩ
Programmable		reverse	0 ~ 181.818	mΩ
impedance		3phase/multichannel	0 ~ 90.909	uН
	L Range	1phase	0 ~ 30.303	uH
		reverse	0 ~ 181.818	uН
		3phase	0 ~ 55	kW
	P Range	1phase	0 ~ 165	kW
		reverse	0 ~ 110	kW
		3phase	0 ~ 55	kVar
	QL Range	1phase	0 ~ 165	kVar
		reverse	0 ~ 110	kVar
		3phase	0 ~ 55	kVar
	QC Range	1phase	0 ~ 165	kVar
Latera d'a la Diago		reverse	0 ~ 110	kVar
Islanding RLC		3phase	0.091 ~ 90.909	Ω
	R Range	1phase	0.03 ~ 30.303	Ω
		reverse	0.182 ~ 181.818	Ω
		3phase	0.091 ~ 454.545	mH
	L Range	1phase	0.03 ~ 151.515	mH
		reverse	0.182 ~ 909.091	mH
		3phase	0.011 ~ 55	mF
	C Range	1phase	0.033 ~ 165	mF
		reverse	0.006 ~ 27.5	mF
Voltage Slew Rate, Typical	≥2	V/μs with full-scale programmed ν	voltage step	
Output Isolation		550Vac		
Measurement pa	arameter			
	Resolution		0.01	V
Voltage RMS	Accuracy	DC,16Hz ~ 150Hz	< 0.1%+0.1% F.S.	
Voltage MIVIO	Temperature coefficient		< 100ppm/C° F.S.	
	Resolution		0.1	A
		DC,16Hz ~ 150Hz	< 0.1% + 0.2% F.S.	A
Current RMS	Accuracy	DC,10HZ ~ ZHUZ		
	Temperature coefficient		< 200ppm/C° F.S.	
	Resolution		1	А
Peak current	Accuracy	16Hz ~ 150Hz	< 0.4% + 0.6% F.S.	/ \
	Resolution	10112 100112	0.1	kW
Output power	Accuracy	DC,16Hz ~ 150Hz	< 0.4% +0.4% F.S.	
Harmonic	Max.	50/60Hz	up to 50	orders
measurement	ividă.	JU/ UUTIZ	άρ το σο	orders
Regenerative				
Max.		165k		VA



THD	< 5%						
Others							
Efficiency	typ *6	91%					
Protection		OVP, OCP, OPP, OTP, FAN, ECP, Sense					
Dimension		600.0000mm (W) *1919 (H) *841mm (D) (909mm include cover)					
Weight		42kg					
Working		0C°-50C°					
temperature		UC*-50C*					
Programming		2ms					
response time		21113					
Remote Sense							
Compensation		20V					
Voltage							
Communication		Built-in USB/CAN/LAN/Digital IO interface, optional GPIB / Analog&RS232					
interface		built-iii 03b/CAIW/LAIW/Digital 10 iiiteHace, optional GPB/ Analog@R3232					

- ± 1 ($200 \sim 220$) $\pm 10\%$, the power of 12kw and above is 60% of the rated.
- *2 Under the output frequency of 50Hz/60Hz, the maximum CF is 6 without exceeding the peak current; under the condition of full current and full power, the maximum CF is 3.
- *3 30kW and above models need to be tested in sense remote measurement mode.
- *4 test condition is: Pure resistive loads, full power conditions.
- $\star 5$ The dynamic response time test condition is: DC high-speed mode, and the capacitance of the DUT to be less than 10 uF
- *6 Test conditions: Input 380VLL/50Hz, output three-phase, per phase 350Vrms/50Hz/Full power.

All the above parameters are subject to change without prior notice from ITECH.



Chapter11 Remote Control

This series power supply comes standard with four communication interfaces: USB, LAN and CAN, and supports two optional communication interfaces: GPIB, RS-232. You can choose one of them to communicate with your computer.



When you use the remote interface to send SCPI instructions, if you use the programming commands that involve modifying the instrument settings, such as modifying the output voltage value, after completing the communication connection between the instrument and the host computer, and after the communication settings are completed, you must execute the SYST:REM command firstly.

11.1 USB Interface

Use cables with both USB ends to connect with IT7900 and PC. All functions are programmable over the USB.

The USB488 interface capabilities are described below:


- The interface is 488.2 USB488 interface.
- The interface accepts REN_CONTROL, GO_TO_LOCAL, and LOCAL_LOCKOUT requests.
- The interface accepts MsgID = TRIGGER USBTMC command message and forwards TRIGGER requests to the function layer.

The USB488 device functions are described below:

- The device understands all mandatory SCPI commands.
- The device is SR1 capable.
- The device is RL1 capable.

The device is DT1 capable.

The operation steps to change the USB interface type in System Menu are as follows.





(System) on the front panel

- 2. Select Communication -> USB and press [Enter].
- 3. Select **USB Type** to Device, and press [Enter].
- 4. Select the USB device class to TMC or VCP.

11.2 LAN Interface

When the user connect PC through LAN interface, the following is required to use the LAN interface. The LAN interface complies with the LXI standard.



Connect Interface

Use the following steps to quickly connect your instrument to your LAN and con-figure it. Two typical LAN interface systems are described below: private LAN and site LAN.

Connect to the private LAN

A private LAN is a network in which LAN-enabled instruments and computers are directly connected. They are typically small, with no centrally-managed resources. When connected to a computer, a standard network cable can be used to connect directly to the computer via the LAN interface.

Connect to the site LAN

A site LAN is a local area network in which LAN-enabled instruments and computers are connected to the network through routers, hubs, and/or switches. They are typically large, centrally-managed networks with services such as DHCP and DNS servers. When connected to a computer, a network cable can be used to connect to the router, and the computer is also con-nected to the router.

∭_{Note}

- When using one crossover cable to connect PC directly, the gateway address of the instrument should be consistent with that of the PC, and the IP address should be at the same network segment with the PC's IP address.
- When the instrument and computer are connected to the router, an independent IP address must be assigned for the instrument.

Configure LAN Interface Information

The configurable parameters of the IT7900 series power supply are described as follows.

LAN Config:

- Mode: IP Address setting method, user can selectautomatically configure the address of the instrument(DHCP) or manually.
- IP: This value is the Internet Protocol (IP) address of the instrument. An IP address is required for all IP and TCP/IP communications with the instrument. An IP Address consists of 4 decimal numbers separated by periods. Each decimal number ranges from 0 through 255 with no leading zeros (for example, 169.254.2.20).
- Mask: This value is used to enable the instrument to determine if a client IP address is on the same local subnet. The same numbering notation applies as for the IP Address. When a client IP address is on a different subnet, all packets must be sent to the Default Gateway.
- Gateway: This value is the IP Address of the default gateway that allows the
 instrument to communicate with systems that are not on the local subnet, as
 determined by the subnet mask setting. The same numbering notation applies as for the IP Address. A value of 0.0.0.0 indicates that no default gate-



way is defined.

 Socket Port: This value indicates the port number corresponding to the service.

How to Configure

Take manual configuration as an example. The steps are as follows:

- 1. Press the composite keys [Shift]+ (System) on the front panel to enter the system menu.
- 2. Select Communication and press [Enter].
- 3. Press the Left/Right key to select **LAN** and set the mode to **Manual**.
- 4. Set the **IP**, **Mask** and the other parameters in turns, and press **[Enter]**.

Using Web Server

The instrument has a built-in Web server for monitoring and controlling the instrument via a Web browser in PC. To use the Web server, connect the instrument and PC over LAN interface and enter the instrument's IP address into the address bar at the top of your PC's Web browser, you can access the front panel control functions including the LAN configuration parameters.

The format of the address entered in the address bar of the browser is **http:// 192.168.0.100**. The specific IP address is subject to the actual instrument settings.

The opened page is displayed as follows:



You can select different pages by clicking the buttons shown in the navigation bar on the left side of the window. The detailed descriptions are as follows.

- Home: Web home interface, displays the model and appearance of the instrument;
- Information: Displays the serial number of the instrument and more system information as well as LAN configuration parameters;
- Web Control: Enables the Web control to begin controlling the instrument.



This page allows you to monitor and control the instrument;

- LAN Configuration: Reconfigure the LAN parameters;
- Manual: Go to the ITECH official website and view or download the relevant documents.
- Upload: Performs a system upgrade.

Click **CONNECT** to connect the PC with the instrument, then click **Select File** to select the system upgrade installation package (for example, IT7900P-U-V000.001.029all.itech), and then click **UPLOAD** performs the upgrade operation. After the upgrade is complete, the instrument needs to be restarted.

11.3 CAN Interface

The CAN interface is located on the rear panel of the instrument and is connected to the computer using a CAN communication cable.

Definition of CAN Pins

The definition of CAN pins are as follows.


Pins	Description
Н	CAN_H
L	CAN_L

CAN Configuration

The user needs to configure the CAN interface parameters in the system menu before using the remote control. The CAN interface parameters are as follows.

Name	Description
Address	Range: 0 – 127
Baud rate	Select the baud rate from the following options: 5k/10k/20k/40k/50k/80k/100k/125k/200k/250k/400k
	/500k/600k/800k/1000k

The operation steps are as follows.



- 2. Select Communication and press [Enter].
- Select CAN and press [Enter].
- 4. Set the baud rate and address, press [Enter].


CAN Troubleshooting

If you meet some problems when communicating with PC by CAN interface,




please check the following items:

- PC and the instrument must have the same baud rate.
- Ensure you have used the correct communication cable (CAN_H, CAN_L). Please pay attention that some cable may not have a correct internal wiring even it is with an appropriate plug.
- The interface cable is correctly connected (CAN_H to CAN_H, CAN_L to CAN_L).
- If the communication signal is poor or unstable, it is recommended to con- nect a 120 Ω terminating resistance.
 - The connection diagram of a single device is as below.



The connection diagram of multiple devices is as below.





When multiple devices are connected, it is recommended to connect the pin 8 (GND) of the P-IO terminal on the rear panel of these devi- ces in parallel, and the communication quality will be improved in the entire CAN network.

11.4 GPIB Interface (Optional)

The GPIB (IEEE-488) interface is assembled in the IT-E176 communication board. Use a GPIB cable to connect GPIB interfaces of the instrument and PC. Please ensure that the screws have been screwed down in order to have a full connection.

GPIB Configuration

Each device on the GPIB (IEEE-488) interface must have a unique whole numaddress between 1 and 30. Your computer's GPIB interface card address must not conflict with any instrument on the interface bus. This setting is nonvolatile; it will not be changed by *RST.

When you purchase the interface accessory and successfully insert it into the corresponding position on the rear panel of the instrument, the menu item for changing the GPIB address appears in the System menu. The specific steps are as follows:




- 1. Ensure that the instrument's power switch is off, that is, the instrument is in Power Off state.
- 2. Insert the separately purchased GPIB interface card into the card slot on the rear panel of the instrument.
- 3. Connect the instrument with the computer via the GPIB cable. After the connection is successful, turn on the power switch of the instrument.
- 4. Press the composite keys [Shift]+ (System) on the front panel to en- ter the system menu.
- 5. Select **Communication** and press **[Enter]**.
- 6. Select GPIB and press [Enter].
- 7. Press the numeric keys to set the GPIB address and press [Enter].

11.5 RS-232 Interface (Optional)

The RS-232 interface shares the same communication card (IT-E177) with the analog function.

Definition of RS-232 Pins

The definition of RS-232 pins are as follows.



When using the RS-232 interface for communication, connect the pin 1, pin 2, and pin 3 of the IT-E177 to the PC. The pin description is as follows:

Pins	Description
1	TXD, transmit data
2	RXD, receive data
3	DGND, ground

RS-232 Configuration

When you purchase the interface accessory and successfully insert it into the corresponding position on the rear panel of the instrument, the RS-232 menu



item will appear in the System menu. The specific steps are as follows:

- 1. Ensure that the instrument's power switch is off, that is, the instrument is in Power Off state.
- 2. Insert the separately purchased RS–232 interface card into the card slot on the rear panel of the instrument.
- 3. Connect the instrument to the computer via an RS–232 cable. After the connection is successful, turn on the power switch of the instrument.
- 4. Press the composite keys [Shift]+ (System) on the front panel to en- ter the system menu.
- 5. Select **Communication** and press **[Enter]**.
- 6. Select RS232 and press [Enter].
- 7. Set the relevant communication parameters in turn, and press [Enter].

RS-232 Troubleshooting

If you meet some problems when communicating with PC by RS-232 interface, please check the following items:

- Check that whether the baud rate of the computer and instrument are the same:
- Make sure the correct cable and adapter are connected. Note that internal wiring may not be correct even if the cable has a suitable plug;
- The cable must be connected to the correct serial ports (COM1, COM2, etc) of PC.

11.6 Commonly Used Commands Overview

The IT7900 series power supply can be connected with the remote control device through the communication interface to realize the remote operation instrument by sending SCPI commands. This series of power supplies provides a detailed commands reference IT7900 Programming Guide.

Listed below are common commands for users to quickly implement common operations. For more command information, refer to the corresponding commands reference.

```
#IDN? //Query identification of instrument

#SYSTEM: FUNCtion ONE //Set the power mode to 1-phase mode

#UNCtion AC //Set the output mode to AC mode

#UNCtion AC //Set the voltage RMS to 220V

#FREQuency 60.0 //Set the frequency to 60Hz

##CURRent: PROTection: RMS 90 //Set the current RMS protect value to
```



90A

```
CURRent: PROTection: PEAK 270
                                     //Set the current Peak protect value to
 270A
 OUTPut. ON
                                  //Turn on/off the output
                                   //Measure the output Vrms
 MEASure: VOLTage?
 MEASure: CURRent?
                                  // Measure the output Irms
                                   // Measure the output power
 MEASure: POWer?
                            //Query the error information of the instrument
 SYSTem: ERRor?
                                    // To clear the error queue
 SYSTem:CLEar
 OUTPut: PROTection: CLEar
                                     //To clear the protection status
```

11.7 Demo Software Introduction

IT7900 series AC/DC power supply provides the following computer control software, convenient for users to achieve different test control. The standard version of the software can be downloaded from the official website, and the optional version of the software needs to be purchased separately.

• IT9000-PV7900 Demo Software(Standard)

IT7900 series power supply supporting remote control software, users can directly obtain the software from ITECH agent and install to the PC, to achieve remote control equipment by visualization method, IT9000-PV7900 software can achieve all the instrument panel operation. And the interface is simple and convenient to operate. For detailed introduction, please refer to IT9000-PV7900 Software Manual.

ASS4000 Avionic & Shipboard Software(Optional)

IT7900 series power supply supporting ASS4000 series software to achieve complex aviation, Marine and IEC regulatory certification testing, through the software interface engineers can choose IEC61000-4-11, 4-13, 4-14, 4-28, MIL704 and ABD regulations.



Appendix

Specifications of Red and Black Test Lines

ITECH provides you with optional red and black test lines, the user can choose the company's test line for testing. For specifications of ITECH test lines and maximum current values, refer to the table below.

Model	Specifi cations	Length	Description
IT-E30110-AB	10A	1m	Alligator clips-Banana plugs A pair of red and black test line
IT-E30110-BB	10A	1m	Banana plugs - Banana plugs A pair of red and black test line
IT-E30110-BY	10A	1m	Banana plugs - Y-type terminals A pair of red and black test line
IT-E30312-YY	30A	1.2m	Y-type terminals - A pair of red and black test line
IT-E30320-YY	30A	2m	Y-type terminals - A pair of red and black test line
IT-E30615-OO	60A	1.5m	Ring terminals - A pair of red and black test line
IT-E31220-OO	120A	2m	Ring terminals - A pair of red and black test line
IT-E32410-OO	240A	1m	Ring terminals - A pair of red and black test line
IT-E32420-OO	240A	2m	Ring terminals - A pair of red and black test line
IT-E33620-OO	360A	2m	Ring terminals - A pair of red and black test line

For maximum current of AWG copper wire, refer to table blow.

AWG	10	12	14	16	18	20	22	24	26	28
The Maximum current value(A)	40	25	20	13	10	7	5	3.5	2.5	1.7

Note: AWG (American Wire Gage), it means X wire (marked on the wire). The table above lists current capacity of single wire at working temperature of 30°C. For reference only.



Connect with us

Thank you for purchasing ITECH products. Any questions, pls. feel free to let us know.

